

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROMA02
Lab Code: ACE **SDG NO.:** Q3725
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL01 **Date Analyzed:** 12/01/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115465.D **Time Analyzed:** 10:38

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.479	5.387	5.587	496.310	500.000	-0.7
Aroclor-1016-2	5.500	5.408	5.608	487.540	500.000	-2.5
Aroclor-1016-3	5.562	5.470	5.670	518.240	500.000	3.6
Aroclor-1016-4	5.657	5.566	5.766	496.140	500.000	-0.8
Aroclor-1016-5	5.948	5.858	6.058	488.160	500.000	-2.4
Aroclor-1260-1	7.061	6.971	7.171	460.180	500.000	-8.0
Aroclor-1260-2	7.315	7.225	7.425	499.040	500.000	-0.2
Aroclor-1260-3	7.672	7.582	7.782	483.760	500.000	-3.2
Aroclor-1260-4	7.894	7.803	8.003	471.850	500.000	-5.6
Aroclor-1260-5	8.200	8.110	8.310	466.800	500.000	-6.6
Decachlorobiphenyl	9.915	9.830	10.030	57.190	50.000	14.4
Tetrachloro-m-xylene	4.340	4.248	4.448	56.280	50.000	12.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROMA02
Lab Code: ACE **SDG NO.:** Q3725
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL01 **Date Analyzed:** 12/01/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115465.D **Time Analyzed:** 10:38

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.717	4.628	4.828	454.430	500.000	-9.1
Aroclor-1016-2	4.737	4.647	4.847	473.050	500.000	-5.4
Aroclor-1016-3	4.911	4.821	5.021	458.020	500.000	-8.4
Aroclor-1016-4	4.953	4.864	5.064	462.250	500.000	-7.6
Aroclor-1016-5	5.165	5.076	5.276	472.970	500.000	-5.4
Aroclor-1260-1	6.193	6.106	6.306	426.140	500.000	-14.8
Aroclor-1260-2	6.381	6.294	6.494	402.960	500.000	-19.4
Aroclor-1260-3	6.533	6.446	6.646	417.810	500.000	-16.4
Aroclor-1260-4	7.004	6.917	7.117	422.800	500.000	-15.4
Aroclor-1260-5	7.244	7.158	7.358	424.970	500.000	-15.0
Decachlorobiphenyl	8.623	8.541	8.741	52.650	50.000	5.3
Tetrachloro-m-xylene	3.642	3.548	3.748	53.050	50.000	6.1