

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROMA02
Lab Code: ACE **SDG NO.:** Q3725
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL03 **Date Analyzed:** 12/02/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115493.D **Time Analyzed:** 01:00

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.475	5.387	5.587	532.930	500.000	6.6
Aroclor-1016-2	5.498	5.408	5.608	469.670	500.000	-6.1
Aroclor-1016-3	5.559	5.470	5.670	495.780	500.000	-0.8
Aroclor-1016-4	5.654	5.566	5.766	496.330	500.000	-0.7
Aroclor-1016-5	5.945	5.858	6.058	493.280	500.000	-1.3
Aroclor-1260-1	7.059	6.971	7.171	473.900	500.000	-5.2
Aroclor-1260-2	7.313	7.225	7.425	519.720	500.000	3.9
Aroclor-1260-3	7.671	7.582	7.782	506.120	500.000	1.2
Aroclor-1260-4	7.893	7.803	8.003	543.380	500.000	8.7
Aroclor-1260-5	8.200	8.110	8.310	487.790	500.000	-2.4
Decachlorobiphenyl	9.914	9.830	10.030	58.990	50.000	18.0
Tetrachloro-m-xylene	4.338	4.248	4.448	58.770	50.000	17.5

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROMA02
Lab Code: ACE **SDG NO.:** Q3725
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL03 **Date Analyzed:** 12/02/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115493.D **Time Analyzed:** 01:00

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.715	4.628	4.828	476.460	500.000	-4.7
Aroclor-1016-2	4.734	4.647	4.847	472.020	500.000	-5.6
Aroclor-1016-3	4.908	4.821	5.021	471.220	500.000	-5.8
Aroclor-1016-4	4.948	4.864	5.064	301.800	500.000	-39.6
Aroclor-1016-5	5.162	5.076	5.276	432.630	500.000	-13.5
Aroclor-1260-1	6.190	6.106	6.306	435.440	500.000	-12.9
Aroclor-1260-2	6.379	6.294	6.494	422.980	500.000	-15.4
Aroclor-1260-3	6.531	6.446	6.646	419.420	500.000	-16.1
Aroclor-1260-4	7.003	6.917	7.117	413.070	500.000	-17.4
Aroclor-1260-5	7.244	7.158	7.358	425.130	500.000	-15.0
Decachlorobiphenyl	8.623	8.541	8.741	53.910	50.000	7.8
Tetrachloro-m-xylene	3.640	3.548	3.748	53.870	50.000	7.7