

**Laboratory Control Sample/Laboratory Control Sample Duplicate Summary**

**SW-846**

|                 |                           |                           |                   |
|-----------------|---------------------------|---------------------------|-------------------|
| <b>SDG No.:</b> | <u>Q3725</u>              | <b>Analytical Method:</b> | <u>8082A</u>      |
| <b>Client:</b>  | <u>Roman E&amp;G Corp</u> | <b>Datafile :</b>         | <u>PO115485.D</u> |

| Lab Sample ID            | Parameter | Spike | Result | Units | Rec | RPD | Qual | RPD  | Limits  |           | RPD |
|--------------------------|-----------|-------|--------|-------|-----|-----|------|------|---------|-----------|-----|
|                          |           |       |        |       |     |     |      | Qual | Low     | High      |     |
| PB170761BS<br>(Column 1) | AR1016    | 166.5 | 143    | ug/kg | 86  |     |      |      | 40 (71) | 140 (120) |     |
|                          | AR1260    | 166.5 | 136    | ug/kg | 82  |     |      |      | 40 (65) | 140 (130) |     |
| PB170761BS<br>(Column 2) | AR1016    | 166.5 | 133    | ug/kg | 80  |     |      |      | 40 (71) | 140 (120) |     |
|                          | AR1260    | 166.5 | 120    | ug/kg | 72  |     |      |      | 40 (65) | 140 (130) |     |