

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROMA02
Lab Code: ACE **SDG NO.:** Q3725
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/14/2025 11/14/2025

Client Sample No.: CCAL01 **Date Analyzed:** 12/01/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091403.D **Time Analyzed:** 10:26

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.695	6.599	6.799	44.700	50.000	-10.6
4,4'-DDE	6.185	6.089	6.289	46.690	50.000	-6.6
4,4'-DDT	7.010	6.914	7.114	44.380	50.000	-11.2
Aldrin	5.258	5.163	5.363	47.230	50.000	-5.5
alpha-BHC	3.990	3.893	4.093	48.950	50.000	-2.1
alpha-Chlordane	6.015	5.919	6.119	44.600	50.000	-10.8
beta-BHC	4.508	4.411	4.611	46.600	50.000	-6.8
Decachlorobiphenyl	9.057	8.963	9.163	42.270	50.000	-15.5
Dieldrin	6.335	6.239	6.439	46.600	50.000	-6.8
Endosulfan I	6.062	5.966	6.166	46.390	50.000	-7.2
Endosulfan II	6.776	6.679	6.879	45.210	50.000	-9.6
Endosulfan sulfate	7.139	7.043	7.243	44.510	50.000	-11.0
Endrin	6.563	6.466	6.666	46.620	50.000	-6.8
gamma-BHC (Lindane)	4.321	4.224	4.424	48.460	50.000	-3.1
gamma-Chlordane	5.934	5.838	6.038	47.500	50.000	-5.0
Heptachlor	4.917	4.821	5.021	45.800	50.000	-8.4
Heptachlor epoxide	5.679	5.582	5.782	46.270	50.000	-7.5
Methoxychlor	7.484	7.387	7.587	42.430	50.000	-15.1
Tetrachloro-m-xylene	3.541	3.444	3.644	47.240	50.000	-5.5

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROMA02
Lab Code: ACE **SDG NO.:** Q3725
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/14/2025 11/14/2025

Client Sample No.: CCAL01 **Date Analyzed:** 12/01/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091403.D **Time Analyzed:** 10:26

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.913	5.817	6.017	47.690	50.000	-4.6
4,4'-DDE	5.359	5.263	5.463	50.420	50.000	0.8
4,4'-DDT	6.167	6.071	6.271	49.540	50.000	-0.9
Aldrin	4.355	4.258	4.458	51.440	50.000	2.9
alpha-BHC	3.383	3.286	3.486	51.310	50.000	2.6
alpha-Chlordane	5.175	5.079	5.279	51.330	50.000	2.7
beta-BHC	4.015	3.918	4.118	50.290	50.000	0.6
Decachlorobiphenyl	8.054	7.958	8.158	44.000	50.000	-12.0
Dieldrin	5.497	5.401	5.601	50.310	50.000	0.6
Endosulfan I	5.232	5.135	5.335	51.130	50.000	2.3
Endosulfan II	6.066	5.969	6.169	49.570	50.000	-0.9
Endosulfan sulfate	6.467	6.371	6.571	49.110	50.000	-1.8
Endrin	5.774	5.678	5.878	49.170	50.000	-1.7
gamma-BHC (Lindane)	3.718	3.621	3.821	51.080	50.000	2.2
gamma-Chlordane	5.111	5.014	5.214	51.720	50.000	3.4
Heptachlor	4.070	3.973	4.173	49.140	50.000	-1.7
Heptachlor epoxide	4.858	4.762	4.962	50.580	50.000	1.2
Methoxychlor	6.738	6.641	6.841	44.320	50.000	-11.4
Tetrachloro-m-xylene	2.872	2.774	2.974	50.640	50.000	1.3