

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROMA02
Lab Code: ACE **SDG NO.:** Q3725
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/14/2025 11/14/2025

Client Sample No.: CCAL02 **Date Analyzed:** 12/01/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091414.D **Time Analyzed:** 15:11

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.704	6.599	6.799	49.440	50.000	-1.1
4,4'-DDE	6.194	6.089	6.289	51.150	50.000	2.3
4,4'-DDT	7.019	6.914	7.114	49.550	50.000	-0.9
Aldrin	5.268	5.163	5.363	52.100	50.000	4.2
alpha-BHC	3.998	3.893	4.093	54.790	50.000	9.6
alpha-Chlordane	6.024	5.919	6.119	49.410	50.000	-1.2
beta-BHC	4.517	4.411	4.611	51.610	50.000	3.2
Decachlorobiphenyl	9.069	8.963	9.163	46.220	50.000	-7.6
Dieldrin	6.344	6.239	6.439	51.070	50.000	2.1
Endosulfan I	6.072	5.966	6.166	50.770	50.000	1.5
Endosulfan II	6.785	6.679	6.879	49.380	50.000	-1.2
Endosulfan sulfate	7.148	7.043	7.243	48.480	50.000	-3.0
Endrin	6.572	6.466	6.666	53.040	50.000	6.1
gamma-BHC (Lindane)	4.329	4.224	4.424	54.050	50.000	8.1
gamma-Chlordane	5.944	5.838	6.038	51.870	50.000	3.7
Heptachlor	4.926	4.821	5.021	52.220	50.000	4.4
Heptachlor epoxide	5.688	5.582	5.782	51.400	50.000	2.8
Methoxychlor	7.493	7.387	7.587	47.090	50.000	-5.8
Tetrachloro-m-xylene	3.549	3.444	3.644	51.530	50.000	3.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROMA02
Lab Code: ACE **SDG NO.:** Q3725
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/14/2025 11/14/2025

Client Sample No.: CCAL02 **Date Analyzed:** 12/01/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091414.D **Time Analyzed:** 15:11

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.916	5.817	6.017	47.640	50.000	-4.7
4,4'-DDE	5.360	5.263	5.463	50.130	50.000	0.3
4,4'-DDT	6.170	6.071	6.271	50.770	50.000	1.5
Aldrin	4.356	4.258	4.458	52.400	50.000	4.8
alpha-BHC	3.383	3.286	3.486	53.220	50.000	6.4
alpha-Chlordane	5.177	5.079	5.279	51.270	50.000	2.5
beta-BHC	4.015	3.918	4.118	51.640	50.000	3.3
Decachlorobiphenyl	8.058	7.958	8.158	41.200	50.000	-17.6
Dieldrin	5.500	5.401	5.601	50.050	50.000	0.1
Endosulfan I	5.234	5.135	5.335	50.560	50.000	1.1
Endosulfan II	6.068	5.969	6.169	48.730	50.000	-2.5
Endosulfan sulfate	6.470	6.371	6.571	46.480	50.000	-7.0
Endrin	5.776	5.678	5.878	51.670	50.000	3.3
gamma-BHC (Lindane)	3.719	3.621	3.821	52.770	50.000	5.5
gamma-Chlordane	5.113	5.014	5.214	51.780	50.000	3.6
Heptachlor	4.071	3.973	4.173	50.960	50.000	1.9
Heptachlor epoxide	4.860	4.762	4.962	51.170	50.000	2.3
Methoxychlor	6.741	6.641	6.841	44.990	50.000	-10.0
Tetrachloro-m-xylene	2.872	2.774	2.974	52.350	50.000	4.7