

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF110525\
 Data File : BF144175.D
 Acq On : 05 Nov 2025 16:19
 Operator : RC/JU
 Sample : SSTDICC060
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC060

Quant Time: Nov 05 17:24:17 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF110525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 05 17:20:41 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.875	152	63801	20.000	ng	0.00
21) Naphthalene-d8	8.157	136	239743	20.000	ng	0.00
39) Acenaphthene-d10	9.916	164	126385	20.000	ng	0.00
64) Phenanthrene-d10	11.410	188	213136	20.000	ng	0.00
76) Chrysene-d12	14.057	240	145867	20.000	ng	0.00
86) Perylene-d12	15.539	264	205758	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.492	112	509184	124.890	ng	0.00
7) Phenol-d6	6.510	99	559443	121.101	ng	0.00
23) Nitrobenzene-d5	7.439	82	519687	125.195	ng	0.00
42) 2,4,6-Tribromophenol	10.710	330	198226	124.986	ng	0.00
45) 2-Fluorobiphenyl	9.233	172	1047440	122.404	ng	0.00
79) Terphenyl-d14	12.998	244	1057426	115.147	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.604	88	109742	65.106	ng	99
3) Pyridine	3.375	79	307664	65.424	ng	97
4) n-Nitrosodimethylamine	3.334	42	166098	67.619	ng	97
6) Aniline	6.539	93	409356	62.194	ng	99
8) 2-Chlorophenol	6.663	128	247503	61.902	ng	99
9) Benzaldehyde	6.428	77	150534	57.706	ng	98
10) Phenol	6.522	94	342646	60.707	ng	90
11) bis(2-Chloroethyl)ether	6.610	93	256657	62.405	ng	96
12) 1,3-Dichlorobenzene	6.816	146	303642	61.543	ng	98
13) 1,4-Dichlorobenzene	6.892	146	306973	62.104	ng	99
14) 1,2-Dichlorobenzene	7.045	146	293432	61.936	ng	99
15) Benzyl Alcohol	7.016	79	229107	62.632	ng	97
16) 2,2'-oxybis(1-Chloropr...	7.145	45	467387	61.712	ng	96
17) 2-Methylphenol	7.128	107	191705	60.269	ng	96
18) Hexachloroethane	7.386	117	103964	63.308	ng	99
19) n-Nitroso-di-n-propyla...	7.292	70	179314	58.963	ng	96
20) 3+4-Methylphenols	7.281	107	224891	59.977	ng	91
22) Acetophenone	7.286	105	302907	58.897	ng	94
24) Nitrobenzene	7.463	77	285924	63.439	ng	98
25) Isophorone	7.698	82	485836	61.876	ng	99
26) 2-Nitrophenol	7.775	139	144140	64.606	ng	96
27) 2,4-Dimethylphenol	7.810	122	212249	63.471	ng	95
28) bis(2-Chloroethoxy)met...	7.904	93	304411	62.085	ng	99
29) 2,4-Dichlorophenol	8.016	162	234718	60.899	ng	99
30) 1,2,4-Trichlorobenzene	8.098	180	240858	60.858	ng	98
31) Naphthalene	8.181	128	684579	60.028	ng	100
32) Benzoic acid	7.933	122	136682	55.761	ng	96
33) 4-Chloroaniline	8.228	127	255554	61.441	ng	98
34) Hexachlorobutadiene	8.298	225	182626	61.229	ng	98
35) Caprolactam	8.610	113	62840	59.512	ng	95
36) 4-Chloro-3-methylphenol	8.710	107	208910	60.481	ng	99
37) 2-Methylnaphthalene	8.875	142	447909	58.743	ng	98
38) 1-Methylnaphthalene	8.975	142	431529	58.654	ng	100
40) 1,2,4,5-Tetrachloroben...	9.039	216	271508	65.549	ng	99
41) Hexachlorocyclopentadiene	9.022	237	95909	64.419	ng	96
43) 2,4,6-Trichlorophenol	9.151	196	185412	65.769	ng	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF110525\
 Data File : BF144175.D
 Acq On : 05 Nov 2025 16:19
 Operator : RC/JU
 Sample : SSTDICC060
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC060

Quant Time: Nov 05 17:24:17 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF110525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 05 17:20:41 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.192	196	190324	62.639	ng	99
46) 1,1'-Biphenyl	9.339	154	549700	62.310	ng	99
47) 2-Chloronaphthalene	9.363	162	474514	63.056	ng	99
48) 2-Nitroaniline	9.457	65	141938	65.370	ng	98
49) Acenaphthylene	9.775	152	636324	62.050	ng	100
50) Dimethylphthalate	9.633	163	521205	62.250	ng	99
51) 2,6-Dinitrotoluene	9.698	165	110140	64.983	ng	99
52) Acenaphthene	9.951	154	423653	61.778	ng	99
53) 3-Nitroaniline	9.869	138	119967	64.749	ng	100
54) 2,4-Dinitrophenol	9.980	184	63597	62.138	ng	98
55) Dibenzofuran	10.122	168	592400	61.680	ng	98
56) 4-Nitrophenol	10.033	139	80163	59.812	ng	97
57) 2,4-Dinitrotoluene	10.104	165	142910	62.984	ng	97
58) Fluorene	10.469	166	435687	59.664	ng	99
59) 2,3,4,6-Tetrachlorophenol	10.239	232	140953	65.569	ng	97
60) Diethylphthalate	10.333	149	468931	60.755	ng	98
61) 4-Chlorophenyl-phenyle...	10.457	204	249857	60.068	ng	97
62) 4-Nitroaniline	10.486	138	109153	61.862	ng	97
63) Azobenzene	10.616	77	453890	63.187	ng	99
65) 4,6-Dinitro-2-methylph...	10.516	198	90823	62.727	ng	97
66) n-Nitrosodiphenylamine	10.574	169	421720	61.520	ng	99
67) 4-Bromophenyl-phenylether	10.951	248	170395	62.634	ng	99
68) Hexachlorobenzene	11.016	284	204258	63.744	ng	96
69) Atrazine	11.104	200	134002	61.467	ng	100
70) Pentachlorophenol	11.216	266	103466	64.842	ng	99
71) Phenanthrene	11.433	178	641461	60.448	ng	99
72) Anthracene	11.486	178	656819	60.871	ng	100
73) Carbazole	11.639	167	546322	59.542	ng	100
74) Di-n-butylphthalate	11.969	149	677444	61.075	ng	99
75) Fluoranthene	12.627	202	667021	60.849	ng	99
77) Benzidine	12.745	184	277725	64.368	ng	100
78) Pyrene	12.857	202	686663	58.384	ng	100
80) Butylbenzylphthalate	13.468	149	249274	61.152	ng	100
81) Benzo(a)anthracene	14.051	228	653341	64.625	ng	100
82) 3,3'-Dichlorobenzidine	14.010	252	232807	67.182	ng	96
83) Chrysene	14.086	228	580449	62.196	ng	98
84) Bis(2-ethylhexyl)phtha...	14.027	149	351501	62.991	ng	98
85) Di-n-octyl phthalate	14.645	149	631065	65.547	ng	100
87) Indeno(1,2,3-cd)pyrene	17.056	276	938738	63.557	ng	100
88) Benzo(b)fluoranthene	15.109	252	700099	59.849	ng	99
89) Benzo(k)fluoranthene	15.139	252	699851	63.919	ng	98
90) Benzo(a)pyrene	15.480	252	680720	63.078	ng	99
91) Dibenzo(a,h)anthracene	17.074	278	746544	62.936	ng	98
92) Benzo(g,h,i)perylene	17.515	276	757880	64.699	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Nov 05 17:24:17 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF110525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 05 17:20:41 2025
Response via : Initial Calibration

