

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF110525

Review By	rahul	Review On	11/5/2025 2:56:50 PM
Supervise By	Jagrut	Supervise On	11/6/2025 10:21:33 AM
SubDirectory	BF110525	HP Acquire Method	BNA_F
		HP Processing Method	BF110525

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13181,10ul/1000ul sample
ICV/I.BLK	SP6872
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF144168.D	05 Nov 2025 12:21		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BF144169.D	05 Nov 2025 12:50		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BF144170.D	05 Nov 2025 13:20		RC/JU	Ok,M
4	SSTDICC010	SSTDICC010	BF144171.D	05 Nov 2025 13:50		RC/JU	Ok
5	SSTDICC020	SSTDICC020	BF144172.D	05 Nov 2025 14:20		RC/JU	Ok
6	SSTDICCC040	SSTDICCC040	BF144173.D	05 Nov 2025 14:50	Compound #41 kept on QR	RC/JU	Ok
7	SSTDICC050	SSTDICC050	BF144174.D	05 Nov 2025 15:49		RC/JU	Ok
8	SSTDICC060	SSTDICC060	BF144175.D	05 Nov 2025 16:19		RC/JU	Ok
9	SSTDICC080	SSTDICC080	BF144176.D	05 Nov 2025 16:49		RC/JU	Ok
10	SSTDICV040	ICVBF110525	BF144177.D	05 Nov 2025 17:21		RC/JU	Ok,M
11	PB170372BL	PB170372BL	BF144178.D	05 Nov 2025 18:10		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF120125

Review By	Jagrut	Review On	12/2/2025 4:27:26 PM
Supervise By	mohammad	Supervise On	12/3/2025 1:12:18 AM
SubDirectory	BF120125	HP Acquire Method	BNA_F
		HP Processing Method	BF110525

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13184,10ul/1000ul sample
ICV/I.BLK	SP6872
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF144381.D	01 Dec 2025 12:12		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF144382.D	01 Dec 2025 12:41	CCC fail high for compound #9 & #77	RC/JU	Ok,M
3	PB170749BL	PB170749BL	BF144383.D	01 Dec 2025 13:13		RC/JU	Ok
4	PB170749BS	PB170749BS	BF144384.D	01 Dec 2025 13:42		RC/JU	Ok
5	PB170763BL	PB170763BL	BF144385.D	01 Dec 2025 14:11		RC/JU	Ok
6	PB170763BS	PB170763BS	BF144386.D	01 Dec 2025 14:40		RC/JU	Ok
7	Q3740-01	WC1	BF144387.D	01 Dec 2025 15:15		RC/JU	Ok,M
8	Q3719-11	SB-1125-16A	BF144388.D	01 Dec 2025 15:45		RC/JU	Ok,M
9	Q3725-04	STOCKPILE-SAMPLES	BF144389.D	01 Dec 2025 16:14		RC/JU	Ok,M
10	Q3732-01	HD-01-11-26-2025	BF144390.D	01 Dec 2025 16:44		RC/JU	Ok
11	Q3735-01	BU-3-112625	BF144391.D	01 Dec 2025 17:13		RC/JU	Ok
12	Q3735-01MS	BU-3-112625MS	BF144392.D	01 Dec 2025 17:43		RC/JU	Ok
13	Q3735-01MSD	BU-3-112625MSD	BF144393.D	01 Dec 2025 18:12		RC/JU	Ok
14	Q3719-13	SB-1125-17A	BF144394.D	01 Dec 2025 18:41	Need Straight Run	RC/JU	Not Ok
15	Q3719-14	SB-1125-17B	BF144395.D	01 Dec 2025 19:11		RC/JU	Ok,M
16	Q3733-01	SU-04-11-26-2025	BF144396.D	01 Dec 2025 19:40	Double IS added	RC/JU	Ok

M : Manual Integration