

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY112625

Review By	Semsettin Yesilyurt	Review On	12/1/2025 9:50:11 AM
Supervise By	John Carlone	Supervise On	12/1/2025 2:45:01 PM
SubDirectory	VY112625	HP Acquire Method	MSVOA_Y HP Processing Method 82y112625s.m

STD. NAME	STD REF.#
Tune/Reschk	VP136432
Initial Calibration Stds	VP136433,VP136434,VP136435,VP136436,VP136437,VP136438
CCC	VP136440
Internal Standard/PEM	VP136146
ICV/I.BLK	VP136439
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY023808.D	26 Nov 2025 07:11		SY/MD	Ok
2	VSTDICC005	VSTDICC005	VY023809.D	26 Nov 2025 08:04		SY/MD	Ok,M
3	VSTDICC010	VSTDICC010	VY023810.D	26 Nov 2025 08:27		SY/MD	Ok,M
4	VSTDICC020	VSTDICC020	VY023811.D	26 Nov 2025 08:49		SY/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VY023812.D	26 Nov 2025 09:17		SY/MD	Ok,M
6	VSTDICC100	VSTDICC100	VY023813.D	26 Nov 2025 09:39		SY/MD	Ok,M
7	VSTDICC150	VSTDICC150	VY023814.D	26 Nov 2025 10:02		SY/MD	Ok,M
8	VIBLK	VIBLK	VY023815.D	26 Nov 2025 10:27		SY/MD	Ok
9	VSTDICV050	ICVVY112625	VY023816.D	26 Nov 2025 10:50		SY/MD	Ok,M
10	VY1126SBL01	VY1126SBL01	VY023817.D	26 Nov 2025 11:13		SY/MD	Ok
11	VY1126SBS01	VY1126SBS01	VY023818.D	26 Nov 2025 11:47		SY/MD	Ok,M
12	VY1126SBSD01	VY1126SBSD01	VY023819.D	26 Nov 2025 12:09		SY/MD	Ok,M
13	Q3604-04	S-1	VY023820.D	26 Nov 2025 12:34	not reported,vial A	SY/MD	Not Ok
14	Q3604-05	S-2	VY023821.D	26 Nov 2025 12:57	not reported,vial A	SY/MD	Not Ok
15	VIBLK	VIBLK	VY023822.D	26 Nov 2025 13:21		SY/MD	Ok
16	Q3721-01	EO-2-112525	VY023823.D	26 Nov 2025 13:44	vial A	SY/MD	Ok
17	Q3723-01	CC-1-112525	VY023824.D	26 Nov 2025 14:08	vial A	SY/MD	Ok
18	Q3725-04	STOCKPILE-SAMPLES	VY023825.D	26 Nov 2025 14:31	vial A	SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY112625

Review By	Semsettin Yesilyurt	Review On	12/1/2025 9:50:11 AM		
Supervise By	John Carlone	Supervise On	12/1/2025 2:45:01 PM		
SubDirectory	VY112625	HP Acquire Method	MSVOA_Y	HP Processing Method	82y112625s.m

STD. NAME	STD REF.#
Tune/Reschk	VP136432
Initial Calibration Stds	VP136433,VP136434,VP136435,VP136436,VP136437,VP136438
CCC	VP136440
Internal Standard/PEM	VP136146
ICV/I.BLK	VP136439
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	VIBLK	VIBLK	VY023826.D	26 Nov 2025 14:54		SY/MD	Ok
20	VSTDCCC050	VSTDCCC050EC	VY023827.D	26 Nov 2025 15:17		SY/MD	Ok,M

M : Manual Integration