

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: ROMA02
 Lab Code: ACE SDG NO.: Q3725
 Lab File ID: VY023812.D Date Analyzed: 11/26/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:17
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	569944	7.71	897275	8.62	788206	11.41
UPPER LIMIT	1139890	8.207	1794550	9.116	1576410	11.914
LOWER LIMIT	284972	7.207	448638	8.116	394103	10.914
EPA SAMPLE NO.						
STOCKPILE-SAMPLES	711387	7.71	1273779	8.62	920727	11.41
VY1126SBL01	798267	7.71	1469518	8.62	1241361	11.41
VY1126SBS01	585172	7.71	926752	8.62	802455	11.41
VY1126SBSD01	572632	7.71	914870	8.62	774262	11.41

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.