

Initial and Continuing Calibration Verification

Client: ATG-GREENVILLE AEC

SDG No.: Q3728

Project: AEC-2025-0013

RunNo.: LB138070

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1						
Bromide	mg/L	9.7	10	97	90-110	11/03/2025
Chloride	mg/L	2.9	3	97	90-110	11/03/2025
Fluoride	mg/L	2	2	100	90-110	11/03/2025
Nitrite	mg/L	2.9	3	97	90-110	11/03/2025
Nitrate	mg/L	2.4	2.5	96	90-110	11/03/2025
Sulfate	mg/L	14.5	15	97	90-110	11/03/2025
Orthophosphate as P	mg/L	5.2	5	104	90-110	11/03/2025
Sample ID: CCV1						
Bromide	mg/L	10.4	10	104	90-110	11/26/2025
Chloride	mg/L	3.1	3	103	90-110	11/26/2025
Fluoride	mg/L	2.1	2	105	90-110	11/26/2025
Nitrite	mg/L	3.1	3	103	90-110	11/26/2025
Nitrate	mg/L	2.5	2.5	100	90-110	11/26/2025
Sulfate	mg/L	15.6	15	104	90-110	11/26/2025
Orthophosphate as P	mg/L	4.9	5	98	90-110	11/26/2025
Sample ID: CCV2						
Bromide	mg/L	10.4	10	104	90-110	11/26/2025
Chloride	mg/L	3.1	3	103	90-110	11/26/2025
Fluoride	mg/L	2.1	2	105	90-110	11/26/2025
Nitrite	mg/L	3.1	3	103	90-110	11/26/2025
Nitrate	mg/L	2.5	2.5	100	90-110	11/26/2025
Sulfate	mg/L	15.4	15	103	90-110	11/26/2025
Orthophosphate as P	mg/L	5.1	5	102	90-110	11/26/2025

Initial and Continuing Calibration Verification

Client: ATG-GREENVILLE AEC

SDG No.: Q3728

Project: AEC-2025-0013

RunNo.: LB138131

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 TKN	mg/L	5	5	100	90-110	12/05/2025
Sample ID: CCV1 TKN	mg/L	5.1	5	102	90-110	12/05/2025
Sample ID: CCV2 TKN	mg/L	5	5	100	90-110	12/05/2025
Sample ID: CCV3 TKN	mg/L	5.1	5	102	90-110	12/05/2025

Initial and Continuing Calibration Verification

Client: ATG-GREENVILLE AEC

SDG No.: Q3728

Project: AEC-2025-0013

RunNo.: LB138132

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Phosphorus, Total	mg/L	0.522	0.50	104	90-110	12/05/2025
Sample ID: CCV1 Phosphorus, Total	mg/L	0.499	0.50	100	90-110	12/05/2025
Sample ID: CCV2 Phosphorus, Total	mg/L	0.507	0.50	101	90-110	12/05/2025
Sample ID: CCV3 Phosphorus, Total	mg/L	0.498	0.50	100	90-110	12/05/2025