

Laboratory Control Sample Summary

Client: ATG-GREENVILLE AEC

SDG No.: Q3728

Project: AEC-2025-0013

Run No.: LB138060

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138060BS							
TSS	mg/L	550	568		103	1	90-110	11/26/2025

Laboratory Control Sample Summary

Client: ATG-GREENVILLE AEC

SDG No.: Q3728

Project: AEC-2025-0013

Run No.: LB138070

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138070BSW							
Bromide	mg/L	10	10.4		104	1	90-110	11/26/2025
Chloride	mg/L	3	3.10		103	1	90-110	11/26/2025
Fluoride	mg/L	2	2.10		105	1	90-110	11/26/2025
Nitrite	mg/L	3	3.10		103	1	90-110	11/26/2025
Nitrate	mg/L	2.5	2.50		100	1	90-110	11/26/2025
Sulfate	mg/L	15	15.4		103	1	90-110	11/26/2025
Orthophosphate as P	mg/L	5	5.00		100	1	90-110	11/26/2025

Laboratory Control Sample Summary

Client: ATG-GREENVILLE AEC

SDG No.: Q3728

Project: AEC-2025-0013

Run No.: LB138131

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB170831BS							
TKN	mg/L	5	5.00		100	1	90-110	12/05/2025

Laboratory Control Sample Summary

Client: ATG-GREENVILLE AEC

SDG No.: Q3728

Project: AEC-2025-0013

Run No.: LB138132

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB170844BS							
Phosphorus, Total	mg/L	0.50	0.52		104	1	90-110	12/05/2025