

Matrix Spike Summary

Client:	ATG-GREENVILLE AEC	SDG No.:	Q3729
Project:	Accuracy	Sample ID:	Q3778-01
Client ID:	SW-2MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Phosphorus, Total	mg/L	90-110	0.53		0.064		0.5	1	93		12/05/2025

Matrix Spike Summary

Client:	ATG-GREENVILLE AEC	SDG No.:	Q3729
Project:	Accuracy	Sample ID:	Q3778-01
Client ID:	SW-2MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Phosphorus, Total	mg/L	90-110	0.52		0.064		0.5	1	92		12/05/2025

Matrix Spike Summary

Client:	ATG-GREENVILLE AEC	SDG No.:	Q3729
Project:	Accuracy	Sample ID:	Q3785-04
Client ID:	WATER-TREATMEN DISCHARGEMS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
COD	mg/L	75-125	48.2		1.50	U	50.0	1	96		12/05/2025

Matrix Spike Summary

Client:	ATG-GREENVILLE AEC	SDG No.:	Q3729
Project:	Accuracy	Sample ID:	Q3785-04
Client ID:	WATER-TREATMEN DISCHARGEMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
COD	mg/L	75-125	47.1		1.50	U	50.0	1	94		12/05/2025