

Laboratory Control Sample Summary

Client: ATG-GREENVILLE AEC

SDG No.: Q3729

Project: Accuracy

Run No.: LB138060

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138060BS							
TSS	mg/L	550	568		103	1	90-110	11/26/2025

Laboratory Control Sample Summary

Client: ATG-GREENVILLE AEC

SDG No.: Q3729

Project: Accuracy

Run No.: LB138064

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138064BS							
Oil and Grease	mg/L	20.0	18.9		94	1	78-114	12/01/2025

Laboratory Control Sample Summary

Client: ATG-GREENVILLE AEC

SDG No.: Q3729

Project: Accuracy

Run No.: LB138064

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138064BSD							
Oil and Grease	mg/L	20.0	19.2		96	1	78-114	12/01/2025

Laboratory Control Sample Summary

Client: ATG-GREENVILLE AEC

SDG No.: Q3729

Project: Accuracy

Run No.: LB138124

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138124BS							
COD	mg/L	50	50.2		100	1	90-110	12/05/2025

Laboratory Control Sample Summary

Client: ATG-GREENVILLE AEC

SDG No.: Q3729

Project: Accuracy

Run No.: LB138132

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB170844BS							
Phosphorus, Total	mg/L	0.50	0.52		104	1	90-110	12/05/2025