



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: JPCL Engineering

SDG No.: Q3731

Project: NYCDPR BGCY Playground - W 145th St

RunNo.: LB138058

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Flash Point	ICV ° F	82.7	81	102	78-84	11/26/2025