



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: JPCL Engineering

SDG No.: Q3731

Contract: JPCL01

Lab Code: ACE

Initial Calibration Source:

Continuing Calibration Source:

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Arsenic	22.7	20.0	114	65 - 135	P	12/01/2025	13:50	LB138072
	Barium	102	100	102	65 - 135	P	12/01/2025	13:50	LB138072
	Cadmium	6.21	6.0	104	65 - 135	P	12/01/2025	13:50	LB138072
	Chromium	9.87	10.0	99	65 - 135	P	12/01/2025	13:50	LB138072
	Lead	13.2	12.0	110	65 - 135	P	12/01/2025	13:50	LB138072
	Selenium	22.4	20.0	112	65 - 135	P	12/01/2025	13:50	LB138072
	Silver	10.2	10.0	102	65 - 135	P	12/01/2025	13:50	LB138072
CRA	Mercury	0.15	0.2	73	70 - 130	CV	12/01/2025	15:09	lb138068