

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: JPCL Engineering

SDG No.: Q3731

Contract: JPCL01

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	4.50			-20	20	12/01/2025	13:55	LB138072
	Barium	4.98	6.0	83	-94	106	12/01/2025	13:55	LB138072
	Cadmium	4.78	1.0	478	-5	7	12/01/2025	13:55	LB138072
	Chromium	59.8	52.0	115	42	62	12/01/2025	13:55	LB138072
	Lead	2.07			-12	12	12/01/2025	13:55	LB138072
	Selenium	-14.4			-20	20	12/01/2025	13:55	LB138072
	Silver	-8.77			-10	10	12/01/2025	13:55	LB138072
ICSAB01	Arsenic	106	104	102	88.4	120	12/01/2025	14:09	LB138072
	Barium	515	537	96	437	637	12/01/2025	14:09	LB138072
	Cadmium	979	972	101	826	1120	12/01/2025	14:09	LB138072
	Chromium	528	542	97	460	624	12/01/2025	14:09	LB138072
	Lead	53.1	49.0	108	37	61	12/01/2025	14:09	LB138072
	Selenium	41.4	46.0	90	26	66	12/01/2025	14:09	LB138072
	Silver	174	201	87	170	232	12/01/2025	14:09	LB138072