

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>JPCL Engineering</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3731</u>
Contract:	<u>JPCL01</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q3731-01</u>	Client ID:	<u>W1DUP</u>
Percent Solids for Sample:	NA	Duplicate ID	Q3731-01DUP	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	0.84		0.82		2		CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>JPCL Engineering</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3731</u>
Contract:	<u>JPCL01</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q3731-01MS</u>	Client ID:	<u>W1MSD</u>
Percent Solids for Sample:	NA	Duplicate ID	Q3731-01MSD	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	4.09		4.00		2		CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>JPCL Engineering</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3731</u>
Contract:	<u>JPCL01</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q3737-02</u>	Client ID:	<u>MW4DUP</u>
Percent Solids for Sample:	NA	Duplicate ID	Q3737-02DUP	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	10.0	U	10.0	U			P
Barium	ug/L	20	115		112		3		P
Cadmium	ug/L	20	0.32	J	0.37	J	13		P
Chromium	ug/L	20	20.8		21.8		5		P
Lead	ug/L	20	6.12		4.67	J	27		P
Selenium	ug/L	20	10.0	U	10.0	U			P
Silver	ug/L	20	5.00	U	5.00	U			P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>JPCL Engineering</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3731</u>
Contract:	<u>JPCL01</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q3737-02MS</u>	Client ID:	<u>MW4MSD</u>
Percent Solids for Sample:	NA	Duplicate ID	Q3737-02MSD	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	435		433		0		P
Barium	ug/L	20	209		211		1		P
Cadmium	ug/L	20	103		102		1		P
Chromium	ug/L	20	216		208		4		P
Lead	ug/L	20	498		494		1		P
Selenium	ug/L	20	1130		1130		0		P
Silver	ug/L	20	38.9		38.6		1		P