

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: JPCL Engineering

SDG No.: Q3731

Contract: JPCL01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170768BS							
Arsenic	ug/L	400	402		100	80 - 120	P
Barium	ug/L	100	96.9		97	80 - 120	P
Cadmium	ug/L	100	96.9		97	80 - 120	P
Chromium	ug/L	200	196		98	80 - 120	P
Lead	ug/L	500	474		95	80 - 120	P
Selenium	ug/L	1000	1060		106	80 - 120	P
Silver	ug/L	37.5	38.3		102	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: JPCL Engineering

SDG No.: Q3731

Contract: JPCL01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170772BS Mercury	ug/L	4.0	4.04		101	80 - 120	CV