

Metals
- 5b -
POST DIGEST SPIKE SUMMARY

Client: JPCL Engineering

SDG No.: Q3731

Contract: JPCL01

Lab Code: ACE

Matrix: Water

Level: LOW

Client ID: D1A

Sample ID: Q3736-02

Spiked ID: Q3736-02A

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Barium	ug/L	75 - 125	4290		3230		1000	105		P