

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: <u>JPCL Engineering</u>	SDG No.: <u>Q3736</u>
Contract: <u>JPCL01</u>	Lab Code: <u>ACE</u>
ICS Source: <u>EPA</u>	Instrument ID: <u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	4.50			-20	20	12/01/2025	13:55	LB138072
	Barium	4.98	6.0	83	-94	106	12/01/2025	13:55	LB138072
	Cadmium	4.78	1.0	478	-5	7	12/01/2025	13:55	LB138072
	Chromium	59.8	52.0	115	42	62	12/01/2025	13:55	LB138072
	Lead	2.07			-12	12	12/01/2025	13:55	LB138072
	Selenium	-14.4			-20	20	12/01/2025	13:55	LB138072
	Silver	-8.77			-10	10	12/01/2025	13:55	LB138072
ICSAB01	Arsenic	106	104	102	88.4	120	12/01/2025	14:09	LB138072
	Barium	515	537	96	437	637	12/01/2025	14:09	LB138072
	Cadmium	979	972	101	826	1120	12/01/2025	14:09	LB138072
	Chromium	528	542	97	460	624	12/01/2025	14:09	LB138072
	Lead	53.1	49.0	108	37	61	12/01/2025	14:09	LB138072
	Selenium	41.4	46.0	90	26	66	12/01/2025	14:09	LB138072
	Silver	174	201	87	170	232	12/01/2025	14:09	LB138072
ICSA01	Arsenic	2.54			-20	20	12/02/2025	11:52	LB138085
	Barium	4.18	6.0	70	-94	106	12/02/2025	11:52	LB138085
	Cadmium	4.20	1.0	420	-5	7	12/02/2025	11:52	LB138085
	Chromium	46.6	52.0	90	42	62	12/02/2025	11:52	LB138085
	Lead	5.64			-12	12	12/02/2025	11:52	LB138085
	Selenium	-2.16			-20	20	12/02/2025	11:52	LB138085
	Silver	8.41			-10	10	12/02/2025	11:52	LB138085
ICSAB01	Arsenic	108	104	104	88.4	120	12/02/2025	11:57	LB138085
	Barium	500	537	93	437	637	12/02/2025	11:57	LB138085
	Cadmium	1000	972	103	826	1120	12/02/2025	11:57	LB138085
	Chromium	538	542	99	460	624	12/02/2025	11:57	LB138085
	Lead	52.6	49.0	107	37	61	12/02/2025	11:57	LB138085
	Selenium	41.4	46.0	90	26	66	12/02/2025	11:57	LB138085
	Silver	185	201	92	170	232	12/02/2025	11:57	LB138085