



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: JPCL Engineering

SDG No.: Q3736

Contract: JPCL01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB132	Mercury	0.076	+/-0.2	U	0.20	CV	12/01/2025	14:56	lb138068

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: JPCL Engineering

SDG No.: Q3736

Contract: JPCL01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB44	Mercury	0.076	+/-0.2	U	0.20	CV	12/01/2025	15:04	lb138068
CCB45	Mercury	0.076	+/-0.2	U	0.20	CV	12/01/2025	15:37	lb138068
CCB46	Mercury	0.076	+/-0.2	U	0.20	CV	12/01/2025	16:08	lb138068
CCB47	Mercury	0.076	+/-0.2	U	0.20	CV	12/01/2025	16:26	lb138068

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: JPCL Engineering

SDG No.: Q3736

Contract: JPCL01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	5.12	+/-10	U	20.0	P	12/01/2025	13:46	LB138072
	Barium	14.6	+/-50	U	100	P	12/01/2025	13:46	LB138072
	Cadmium	0.50	+/-3	U	6.00	P	12/01/2025	13:46	LB138072
	Chromium	2.12	+/-5	U	10.0	P	12/01/2025	13:46	LB138072
	Lead	2.30	+/-6	U	12.0	P	12/01/2025	13:46	LB138072
	Selenium	9.64	+/-10	U	20.0	P	12/01/2025	13:46	LB138072
	Silver	1.62	+/-5	U	10.0	P	12/01/2025	13:46	LB138072

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: JPCL Engineering

SDG No.: Q3736

Contract: JPCL01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	5.12	+/-10	U	20.0	P	12/01/2025	14:34	LB138072
	Barium	14.6	+/-50	U	100	P	12/01/2025	14:34	LB138072
	Cadmium	0.50	+/-3	U	6.00	P	12/01/2025	14:34	LB138072
	Chromium	2.12	+/-5	U	10.0	P	12/01/2025	14:34	LB138072
	Lead	2.30	+/-6	U	12.0	P	12/01/2025	14:34	LB138072
	Selenium	9.64	+/-10	U	20.0	P	12/01/2025	14:34	LB138072
	Silver	1.62	+/-5	U	10.0	P	12/01/2025	14:34	LB138072
CCB02	Arsenic	5.12	+/-10	U	20.0	P	12/01/2025	15:31	LB138072
	Barium	14.6	+/-50	U	100	P	12/01/2025	15:31	LB138072
	Cadmium	0.50	+/-3	U	6.00	P	12/01/2025	15:31	LB138072
	Chromium	2.12	+/-5	U	10.0	P	12/01/2025	15:31	LB138072
	Lead	2.30	+/-6	U	12.0	P	12/01/2025	15:31	LB138072
	Selenium	9.64	+/-10	U	20.0	P	12/01/2025	15:31	LB138072
	Silver	1.62	+/-5	U	10.0	P	12/01/2025	15:31	LB138072
CCB03	Arsenic	5.12	+/-10	U	20.0	P	12/01/2025	16:38	LB138072
	Barium	14.6	+/-50	U	100	P	12/01/2025	16:38	LB138072
	Cadmium	0.50	+/-3	U	6.00	P	12/01/2025	16:38	LB138072
	Chromium	2.12	+/-5	U	10.0	P	12/01/2025	16:38	LB138072
	Lead	2.30	+/-6	U	12.0	P	12/01/2025	16:38	LB138072
	Selenium	9.64	+/-10	U	20.0	P	12/01/2025	16:38	LB138072
	Silver	1.62	+/-5	U	10.0	P	12/01/2025	16:38	LB138072
CCB04	Arsenic	5.12	+/-10	U	20.0	P	12/01/2025	17:29	LB138072
	Barium	14.6	+/-50	U	100	P	12/01/2025	17:29	LB138072
	Cadmium	0.50	+/-3	U	6.00	P	12/01/2025	17:29	LB138072
	Chromium	2.12	+/-5	U	10.0	P	12/01/2025	17:29	LB138072
	Lead	2.30	+/-6	U	12.0	P	12/01/2025	17:29	LB138072
	Selenium	9.64	+/-10	U	20.0	P	12/01/2025	17:29	LB138072
	Silver	1.62	+/-5	U	10.0	P	12/01/2025	17:29	LB138072
CCB05	Arsenic	5.12	+/-10	U	20.0	P	12/01/2025	18:20	LB138072
	Barium	14.6	+/-50	U	100	P	12/01/2025	18:20	LB138072
	Cadmium	0.50	+/-3	U	6.00	P	12/01/2025	18:20	LB138072
	Chromium	2.12	+/-5	U	10.0	P	12/01/2025	18:20	LB138072
	Lead	2.30	+/-6	U	12.0	P	12/01/2025	18:20	LB138072
	Selenium	9.64	+/-10	U	20.0	P	12/01/2025	18:20	LB138072
	Silver	1.62	+/-5	U	10.0	P	12/01/2025	18:20	LB138072
CCB06	Arsenic	5.12	+/-10	U	20.0	P	12/01/2025	19:11	LB138072
	Barium	14.6	+/-50	U	100	P	12/01/2025	19:11	LB138072
	Cadmium	0.50	+/-3	U	6.00	P	12/01/2025	19:11	LB138072
	Chromium	2.12	+/-5	U	10.0	P	12/01/2025	19:11	LB138072



- 3a -

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Lead	2.30	+/-6	U	12.0	P	12/01/2025	19:11	LB138072
	Selenium	9.64	+/-10	U	20.0	P	12/01/2025	19:11	LB138072
	Silver	1.62	+/-5	U	10.0	P	12/01/2025	19:11	LB138072
CCB07	Arsenic	5.12	+/-10	U	20.0	P	12/01/2025	19:28	LB138072
	Barium	14.6	+/-50	U	100	P	12/01/2025	19:28	LB138072
	Cadmium	0.50	+/-3	U	6.00	P	12/01/2025	19:28	LB138072
	Chromium	2.12	+/-5	U	10.0	P	12/01/2025	19:28	LB138072
	Lead	2.30	+/-6	U	12.0	P	12/01/2025	19:28	LB138072
	Selenium	9.64	+/-10	U	20.0	P	12/01/2025	19:28	LB138072
	Silver	1.62	+/-5	U	10.0	P	12/01/2025	19:28	LB138072

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: JPCL Engineering

SDG No.: Q3736

Contract: JPCL01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	5.12	+/-10	U	20.0	P	12/02/2025	11:38	LB138085
	Barium	14.6	+/-50	U	100	P	12/02/2025	11:38	LB138085
	Cadmium	0.50	+/-3	U	6.00	P	12/02/2025	11:38	LB138085
	Chromium	2.12	+/-5	U	10.0	P	12/02/2025	11:38	LB138085
	Lead	2.30	+/-6	U	12.0	P	12/02/2025	11:38	LB138085
	Selenium	9.64	+/-10	U	20.0	P	12/02/2025	11:38	LB138085
	Silver	1.62	+/-5	U	10.0	P	12/02/2025	11:38	LB138085

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: JPCL Engineering

SDG No.: Q3736

Contract: JPCL01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	5.12	+/-10	U	20.0	P	12/02/2025	12:14	LB138085
	Barium	14.6	+/-50	U	100	P	12/02/2025	12:14	LB138085
	Cadmium	0.50	+/-3	U	6.00	P	12/02/2025	12:14	LB138085
	Chromium	2.12	+/-5	U	10.0	P	12/02/2025	12:14	LB138085
	Lead	2.30	+/-6	U	12.0	P	12/02/2025	12:14	LB138085
	Selenium	9.64	+/-10	U	20.0	P	12/02/2025	12:14	LB138085
	Silver	1.62	+/-5	U	10.0	P	12/02/2025	12:14	LB138085
CCB02	Arsenic	5.12	+/-10	U	20.0	P	12/02/2025	13:17	LB138085
	Barium	14.6	+/-50	U	100	P	12/02/2025	13:17	LB138085
	Cadmium	0.50	+/-3	U	6.00	P	12/02/2025	13:17	LB138085
	Chromium	2.12	+/-5	U	10.0	P	12/02/2025	13:17	LB138085
	Lead	2.30	+/-6	U	12.0	P	12/02/2025	13:17	LB138085
	Selenium	9.64	+/-10	U	20.0	P	12/02/2025	13:17	LB138085
	Silver	1.62	+/-5	U	10.0	P	12/02/2025	13:17	LB138085
CCB03	Arsenic	5.12	+/-10	U	20.0	P	12/02/2025	14:15	LB138085
	Barium	14.6	+/-50	U	100	P	12/02/2025	14:15	LB138085
	Cadmium	0.50	+/-3	U	6.00	P	12/02/2025	14:15	LB138085
	Chromium	2.12	+/-5	U	10.0	P	12/02/2025	14:15	LB138085
	Lead	2.30	+/-6	U	12.0	P	12/02/2025	14:15	LB138085
	Selenium	9.64	+/-10	U	20.0	P	12/02/2025	14:15	LB138085
	Silver	1.62	+/-5	U	10.0	P	12/02/2025	14:15	LB138085
CCB04	Arsenic	5.12	+/-10	U	20.0	P	12/02/2025	15:20	LB138085
	Barium	14.6	+/-50	U	100	P	12/02/2025	15:20	LB138085
	Cadmium	0.50	+/-3	U	6.00	P	12/02/2025	15:20	LB138085
	Chromium	2.12	+/-5	U	10.0	P	12/02/2025	15:20	LB138085
	Lead	2.30	+/-6	U	12.0	P	12/02/2025	15:20	LB138085
	Selenium	9.64	+/-10	U	20.0	P	12/02/2025	15:20	LB138085
	Silver	1.62	+/-5	U	10.0	P	12/02/2025	15:20	LB138085
CCB05	Arsenic	5.12	+/-10	U	20.0	P	12/02/2025	16:09	LB138085
	Barium	14.6	+/-50	U	100	P	12/02/2025	16:09	LB138085
	Cadmium	0.50	+/-3	U	6.00	P	12/02/2025	16:09	LB138085
	Chromium	2.12	+/-5	U	10.0	P	12/02/2025	16:09	LB138085
	Lead	2.30	+/-6	U	12.0	P	12/02/2025	16:09	LB138085
	Selenium	9.64	+/-10	U	20.0	P	12/02/2025	16:09	LB138085
	Silver	1.62	+/-5	U	10.0	P	12/02/2025	16:09	LB138085
CCB06	Arsenic	5.12	+/-10	U	20.0	P	12/02/2025	16:59	LB138085
	Barium	14.6	+/-50	U	100	P	12/02/2025	16:59	LB138085
	Cadmium	0.50	+/-3	U	6.00	P	12/02/2025	16:59	LB138085
	Chromium	2.12	+/-5	U	10.0	P	12/02/2025	16:59	LB138085

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: JPCL Engineering

SDG No.: Q3736

Contract: JPCL01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Lead	2.30	+/-6	U	12.0	P	12/02/2025	16:59	LB138085
	Selenium	9.64	+/-10	U	20.0	P	12/02/2025	16:59	LB138085
	Silver	1.62	+/-5	U	10.0	P	12/02/2025	16:59	LB138085