

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: JPCL Engineering

SDG No.: Q3736

Contract: JPCL01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170766BS							
Arsenic	ug/L	4000	4120		103	80 - 120	P
Barium	ug/L	1000	1080		108	80 - 120	P
Cadmium	ug/L	1000	1010		101	80 - 120	P
Chromium	ug/L	2000	2060		103	80 - 120	P
Lead	ug/L	5000	5010		100	80 - 120	P
Selenium	ug/L	10000	10400		104	80 - 120	P
Silver	ug/L	380	401		106	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: JPCL Engineering

SDG No.: Q3736

Contract: JPCL01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170770BS Mercury	ug/L	4.0	3.77		94	80 - 120	CV