

## Prep Standard - Chemical Standard Summary

**Order ID :** Q3737

**Test :** Alkalinity,Anions Group1

**Prepbatch ID :**

**Sequence ID/Qc Batch ID:** LB138084, LB138090,

**Standard ID :**

WP115344, WP115865, WP115866, WP115867, WP115868, WP115869, WP115870, WP115871, WP115872, WP115873, WP115874, WP115878, WP115879, WP115892, WP115894,

**Chemical ID :**

M6186, W2647, W3112, W3150, W3163, W3178, W3180, W3197, W3217,



| <u>Recipe ID</u>   | <u>NAME</u>                    | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u>            | <u>PipetteID</u> | <u>Supervised By</u>         |
|--|--------------------------------|--------------------------|------------------|------------------------|--------------------|---------------------------|------------------|------------------------------|
| 4035   | IC ELUENT CONCENTRATE FOR IC-1 | <a href="#">WP115344</a> | 10/23/2025       | 04/23/2026             | Iwona Zarych       | WETCHEM_SCALE_5 (WC SC-5) | None             | Jignesh Parikh<br>10/28/2025 |
| <b><u>FROM</u></b> 2.10000gram of W2647 + 84.75000gram of W3163 + 913.15000ml of W3112 = Final Quantity: 1000.000 ml |                                |                          |                  |                        |                    |                           |                  |                              |

| <u>Recipe ID</u>   | <u>NAME</u>                            | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u> | <u>Supervised By</u>         |
|--|--|--------------------------|------------------|------------------------|--------------------|----------------|------------------|------------------------------|
| 2487   | Anions 300/9056 calibration standard 1 | <a href="#">WP115865</a> | 12/01/2025       | 12/02/2025             | Iwona Zarych       | None           | None             | Jignesh Parikh<br>12/01/2025 |
| <b><u>FROM</u></b> 10.00000ml of W3112 = Final Quantity: 10.000 ml |  |                          |                  |                        |                    |                |                  |                              |



| <u>Recipe ID</u>   | <u>NAME</u>                            | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u>          | <u>Supervised By</u>         |
|--|--|--------------------------|------------------|------------------------|--------------------|----------------|---------------------------|------------------------------|
| 24   | Anions 300/9056 calibration standard 2 | <a href="#">WP115866</a> | 12/01/2025       | 12/02/2025             | Iwona Zarych       | None           | WETCHEM_PIPETTE_3<br>(WC) | Jignesh Parikh<br>12/01/2025 |
| <b><u>FROM</u></b> 0.20000ml of W3180 + 9.80000ml of W3112 = Final Quantity: 10.000 ml |  |                          |                  |                        |                    |                |                           |                              |

| <u>Recipe ID</u>   | <u>NAME</u>                            | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u>          | <u>Supervised By</u>         |
|--|--|--------------------------|------------------|------------------------|--------------------|----------------|---------------------------|------------------------------|
| 25   | Anions 300/9056 calibration standard 3 | <a href="#">WP115867</a> | 12/01/2025       | 12/02/2025             | Iwona Zarych       | None           | WETCHEM_PIPETTE_3<br>(WC) | Jignesh Parikh<br>12/01/2025 |
| <b><u>FROM</u></b> 0.40000ml of W3180 + 9.60000ml of W3112 = Final Quantity: 10.000 ml |  |                          |                  |                        |                    |                |                           |                              |

## Wet Chemistry STANDARD PREPARATION LOG

| <u>Recipe ID</u>  | <u>NAME</u>                            | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u>              | <u>Supervised By</u>             |
|---|--|--------------------------|------------------|------------------------|--------------------|----------------|-------------------------------|----------------------------------|
| 26  | Anions 300/9056 calibration standard 4 | <a href="#">WP115868</a> | 12/01/2025       | 12/02/2025             | Iwona Zarych       | None           | WETCHEM_F<br>IPETTE_3<br>(WC) | Jignesh Parikh<br><br>12/01/2025 |
| <b>FROM</b> 0.50000ml of W3180 + 9.50000ml of W3112 = Final Quantity: 10.000 ml |  |                          |                  |                        |                    |                |                               |                                  |

| <u>Recipe ID</u>   | <u>NAME</u>                                | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u>              | <u>Supervised By</u>             |
|--|--|--------------------------|------------------|------------------------|--------------------|----------------|-------------------------------|----------------------------------|
| 3680   | Anions 300/9056 calibration standard 5-CCV | <a href="#">WP115869</a> | 12/01/2025       | 12/02/2025             | Iwona Zarych       | None           | WETCHEM_F<br>IPETTE_3<br>(WC) | Jignesh Parikh<br><br>12/01/2025 |
| <b>FROM</b> 45.00000ml of W3112 + 5.00000ml of W3180 = Final Quantity: 50.000 ml |  |                          |                  |                        |                    |                |                               |                                  |



| <u>Recipe ID</u>   | <u>NAME</u>                            | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u>          | <u>Supervised By</u>         |
|--|--|--------------------------|------------------|------------------------|--------------------|----------------|---------------------------|------------------------------|
| 3679   | Anions 300/9056 calibration standard 6 | <a href="#">WP115870</a> | 12/01/2025       | 12/02/2025             | Iwona Zarych       | None           | WETCHEM_PIPETTE_3<br>(WC) | Jignesh Parikh<br>12/01/2025 |
| <b><u>FROM</u></b> 2.00000ml of W3180 + 8.00000ml of W3112 = Final Quantity: 10.000 ml |  |                          |                  |                        |                    |                |                           |                              |

| <u>Recipe ID</u>   | <u>NAME</u>                            | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u>          | <u>Supervised By</u>         |
|--|--|--------------------------|------------------|------------------------|--------------------|----------------|---------------------------|------------------------------|
| 3681   | Anions 300/9056 calibration standard 7 | <a href="#">WP115871</a> | 12/01/2025       | 12/02/2025             | Iwona Zarych       | None           | WETCHEM_PIPETTE_3<br>(WC) | Jignesh Parikh<br>12/01/2025 |
| <b><u>FROM</u></b> 2.50000ml of W3180 + 7.50000ml of W3112 = Final Quantity: 10.000 ml |  |                          |                  |                        |                    |                |                           |                              |



| <u>Recipe ID</u>  | <u>NAME</u>                 | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u>          | <u>Supervised By</u>         |
|---|-----------------------------|--------------------------|------------------|------------------------|--------------------|----------------|---------------------------|------------------------------|
| 3233  | Anions 300/9056 ICV-LCS std | <a href="#">WP115872</a> | 12/01/2025       | 12/02/2025             | Iwona Zarych       | None           | WETCHEM_FIPETTE_3<br>(WC) | Jignesh Parikh<br>12/01/2025 |
| <b><u>FROM</u></b> 45.00000ml of W3112 + 5.00000ml of W3197 = Final Quantity: 50.000 ml |                             |                          |                  |                        |                    |                |                           |                              |

| <u>Recipe ID</u>  | <u>NAME</u>        | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u>          | <u>Supervised By</u>         |
|---|--------------------|--------------------------|------------------|------------------------|--------------------|----------------|---------------------------|------------------------------|
| 4036  | IC ELUENT FOR IC-1 | <a href="#">WP115873</a> | 12/01/2025       | 01/01/2026             | Iwona Zarych       | None           | WETCHEM_PIPETTE_3<br>(WC) | Jignesh Parikh<br>12/01/2025 |
| <b><u>FROM</u></b> 1980.00000ml of W3112 + 20.00000ml of WP115344 = Final Quantity: 2000.000 ml |                    |                          |                  |                        |                    |                |                           |                              |

## Wet Chemistry STANDARD PREPARATION LOG

| <u>Recipe ID</u> | <u>NAME</u>       | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u>  | <u>Supervised By</u> |
|------------------|-------------------|--------------------------|------------------|------------------------|--------------------|----------------|-------------------|----------------------|
| 4037             | IC H2SO4 FOR IC-1 | <a href="#">WP115874</a> | 12/01/2025       | 01/01/2026             | Iwona Zarych       | None           | WETCHEM_FIPETTE_3 | Jignesh Parikh       |

(WC)

**FROM** 5.60000ml of M6186 + 994.40000ml of W3112 = Final Quantity: 1000.000 ml

| <u>Recipe ID</u> | <u>NAME</u>                                | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u>  | <u>Supervised By</u> |
|------------------|--|--------------------------|------------------|------------------------|--------------------|----------------|-------------------|----------------------|
| 3680             | Anions 300/9056 calibration standard 5-CCV | <a href="#">WP115878</a> | 12/02/2025       | 12/03/2025             | Iwona Zarych       | None           | WETCHEM_FIPETTE_3 | Jignesh Parikh       |

(WC)

**FROM** 45.00000ml of W3112 + 5.00000ml of W3180 = Final Quantity: 50.000 ml



| <u>Recipe ID</u>  | <u>NAME</u>                 | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u>          | <u>Supervised By</u>         |
|---|-----------------------------|--------------------------|------------------|------------------------|--------------------|----------------|---------------------------|------------------------------|
| 3233  | Anions 300/9056 ICV-LCS std | <a href="#">WP115879</a> | 12/02/2025       | 12/03/2025             | Iwona Zarych       | None           | WETCHEM_PIPETTE_3<br>(WC) | Jignesh Parikh<br>12/04/2025 |
| <b><u>FROM</u></b> 45.00000ml of W3112 + 5.00000ml of W3197 = Final Quantity: 50.000 ml |                             |                          |                  |                        |                    |                |                           |                              |

| <u>Recipe ID</u>  | <u>NAME</u>                             | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u>            | <u>PipetteID</u> | <u>Supervised By</u>         |
|---|---|--------------------------|------------------|------------------------|--------------------|---------------------------|------------------|------------------------------|
| 3407  | Acidity-Alkalinity Stock Std(-+2500PPM) | <a href="#">WP115892</a> | 12/03/2025       | 12/10/2025             | Iwona Zarych       | WETCHEM_SCALE_5 (WC SC-5) | None             | Jignesh Parikh<br>12/04/2025 |
| <b><u>FROM</u></b> 0.62500gram of W3163 + 249.40000ml of W3112 = Final Quantity: 250.000 ml |   |                          |                  |                        |                    |                           |                  |                              |



## Wet Chemistry STANDARD PREPARATION LOG

| <u>Recipe ID</u> | <u>NAME</u>            | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u>   | <u>Supervised By</u>             |
|------------------|------------------------|--------------------------|------------------|------------------------|--------------------|----------------|--------------------|----------------------------------|
| 293              | alkalinity LCSW 50 ppm | <a href="#">WP115894</a> | 12/03/2025       | 12/10/2025             | Iwona Zarych       | None           | Glass<br>Pipette-A | Jignesh Parikh<br><br>12/04/2025 |

**FROM** 196.00000ml of W3112 + 4.00000ml of WP115892 = Final Quantity: 200.000 ml

## CHEMICAL RECEIPT LOG BOOK

| Supplier         | ItemCode / ItemName                                     | Lot #      | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|------------------|---|------------|-----------------|-------------------------|-----------------------------|----------------|
| Seidler Chemical | BA-9673-33 / Sulfuric Acid, Instra-Analyzed (cs/6c2.5L) | 23D2462010 | 07/12/2026      | 08/13/2025 / Sagar      | 08/06/2025 / Sagar          | M6186          |

| Supplier                    | ItemCode / ItemName                           | Lot #      | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-----------------------------|---|------------|-----------------|-------------------------|-----------------------------|----------------|
| PCI Scientific Supply, Inc. | J3506-5 / SODIUM BICARBONATE, PWD, ACS, 2.5KG | 0000240594 | 06/03/2026      | 02/24/2020 / AMANDEEP   | 01/20/2020 / apatel         | W2647          |

| Supplier         | ItemCode / ItemName | Lot #               | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|------------------|---------------------|---------------------|-----------------|-------------------------|-----------------------------|----------------|
| Seidler Chemical | DIW / DI Water      | Daily Lab-Certified | 07/03/2029      | 07/03/2024 / lwona      | 07/03/2024 / lwona          | W3112          |

| Supplier                    | ItemCode / ItemName                  | Lot #  | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-----------------------------|--------------------------------------|--------|-----------------|-------------------------|-----------------------------|----------------|
| PCI Scientific Supply, Inc. | AL74050-8 / SULFURIC ACID, 0.02N, 4L | 235420 | 03/31/2029      | 11/04/2024 / lwona      | 11/04/2024 / lwona          | W3150          |

| Supplier                    | ItemCode / ItemName                         | Lot #      | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-----------------------------|---|------------|-----------------|-------------------------|-----------------------------|----------------|
| PCI Scientific Supply, Inc. | EM-SX0395-3 / SODIUM CARBONATE ANHYDR 2.5KG | 24E3156178 | 09/30/2027      | 12/10/2024 / lwona      | 12/10/2024 / lwona          | W3163          |

| Supplier                    | ItemCode / ItemName              | Lot #   | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-----------------------------|----------------------------------|---------|-----------------|-------------------------|-----------------------------|----------------|
| PCI Scientific Supply, Inc. | AL14055-3 / PH 4 BUFFER SOLUTION | 2411A93 | 10/30/2026      | 04/01/2025 / JIGNESH    | 01/27/2025 / jignesh        | W3178          |

### CHEMICAL RECEIPT LOG BOOK

| Supplier           | ItemCode / ItemName                                  | Lot #        | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|--------------------|--|--------------|-----------------|-------------------------|-----------------------------|----------------|
| Inorganic Ventures | 300-CAL-A-500ML / 300.0 Calibration Standard, 500 ml | V2-MEB742616 | 02/19/2026      | 02/19/2025 / lwona      | 01/27/2025 / lwona          | W3180          |

| Supplier           | ItemCode / ItemName                                  | Lot #  | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|--------------------|--|--------|-----------------|-------------------------|-----------------------------|----------------|
| Inorganic Ventures | 300-CAL-A-500ML / 300.0 Calibration Standard, 500 ml | 040525 | 04/05/2027      | 04/08/2025 / lwona      | 04/08/2025 / lwona          | W3197          |

| Supplier                    | ItemCode / ItemName                     | Lot #   | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-----------------------------|---|---------|-----------------|-------------------------|-----------------------------|----------------|
| PCI Scientific Supply, Inc. | AL14455-3 / buffer solution pH 7 yellow | 2504D34 | 03/31/2027      | 07/02/2025 / jignesh    | 06/26/2025 / lwona          | W3217          |

Sodium Bicarbonate, Powder  
BAKER ANALYZED® A.C.S. Reagent

(sodium hydrogen carbonate)



Material No.: 3506-05  
Batch No.: 0000240594  
Manufactured Date: 2019/06/05  
Retest Date: 2026/06/03  
Revision No: 1

## Certificate of Analysis

Meets ACS Reagent Chemical Requirements,

| Test  | Specification  | Result  |
|---|----------------|---------|
| Assay (NaHCO <sub>3</sub> ) (dried basis)     | 99.7 – 100.3 % | 100.1   |
| Insoluble Matter                              | <= 0.015 %     | < 0.002 |
| Chloride (Cl)                                 | <= 0.003 %     | 0.003   |
| Phosphate (PO <sub>4</sub> )                  | <= 0.001 %     | 0.001   |
| Sulfur Compounds (as SO <sub>4</sub> )        | <= 0.003 %     | 0.003   |
| Calcium (Ca)                                  | <= 0.02 %      | 0.02    |
| Trace Impurities – Iron (Fe)                  | <= 0.001 %     | 0.001   |
| Magnesium (Mg)                                | <= 0.005 %     | 0.005   |
| Potassium (K)                                 | <= 0.005 %     | 0.005   |
| Ammonium (NH <sub>4</sub> )                   | <= 5 ppm       | 5       |
| Trace Impurities – ACS – Heavy Metals (as Pb) | <= 5 ppm       | 5       |

For Laboratory, Research or Manufacturing Use  
Meets Reagent Specifications for testing USP/NF monographs

Country of Origin: US  
Packaging Site: Paris Mfg Ctr & DC

  
Jamie Ethier  
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700  
Avantor Performance Materials, LLC  
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700

Sulfuric Acid  
BAKER INSTRA-ANALYZED® Reagent  
For Trace Metal Analysis  
Low Selenium

avantor™



M6186

Recieve Date :- 08/06/25

Material No.: 9673-33  
Batch No.: 23D2462010  
Manufactured Date: 2023-03-22  
Retest Date: 2028-03-20  
Revision No.: 0

## Certificate of Analysis

| Test   | Specification | Result      |
|--|---------------|-------------|
| ACS - Assay (H <sub>2</sub> SO <sub>4</sub> )                | 95.0 - 98.0 % | 96.1 %      |
| Appearance   | Passes Test   | Passes Test |
| ACS - Color (APHA)   | ≤ 10          | 5           |
| ACS - Residue after Ignition                                 | ≤ 3 ppm       | < 1 ppm     |
| ACS - Substances Reducing Permanganate (as SO <sub>2</sub> ) | ≤ 2 ppm       | < 2 ppm     |
| Ammonium (NH <sub>4</sub> )                                  | ≤ 1 ppm       | 1 ppm       |
| Chloride (Cl)  | ≤ 0.1 ppm     | < 0.1 ppm   |
| Nitrate (NO <sub>3</sub> )                                   | ≤ 0.2 ppm     | < 0.1 ppm   |
| Phosphate (PO <sub>4</sub> )                                 | ≤ 0.5 ppm     | < 0.1 ppm   |
| Trace Impurities - Aluminum (Al)                             | ≤ 30.0 ppb    | < 5.0 ppb   |
| Arsenic and Antimony (as As)                                 | ≤ 4.0 ppb     | < 2.0 ppb   |
| Trace Impurities - Boron (B)                                 | ≤ 10.0 ppb    | 8.5 ppb     |
| Trace Impurities - Cadmium (Cd)                              | ≤ 2.0 ppb     | < 0.3 ppb   |
| Trace Impurities - Chromium (Cr)                             | ≤ 6.0 ppb     | < 0.4 ppb   |
| Trace Impurities - Cobalt (Co)                               | ≤ 0.5 ppb     | < 0.3 ppb   |
| Trace Impurities - Copper (Cu)                               | ≤ 1.0 ppb     | < 0.1 ppb   |
| Trace Impurities - Gold (Au)                                 | ≤ 10.0 ppb    | 0.5 ppb     |
| Heavy Metals (as Pb)   | ≤ 500.0 ppb   | < 100.0 ppb |
| Trace Impurities - Iron (Fe)                                 | ≤ 50.0 ppb    | 1.3 ppb     |
| Trace Impurities - Lead (Pb)                                 | ≤ 0.5 ppb     | < 0.5 ppb   |
| Trace Impurities - Magnesium (Mg)                            | ≤ 7.0 ppb     | 0.8 ppb     |
| Trace Impurities - Manganese (Mn)                            | ≤ 1.0 ppb     | < 0.4 ppb   |
| Trace Impurities - Mercury (Hg)                              | ≤ 0.5 ppb     | < 0.1 ppb   |
| Trace Impurities - Nickel (Ni)                               | ≤ 2.0 ppb     | 0.3 ppb     |
| Trace Impurities - Potassium (K)                             | ≤ 500.0 ppb   | < 2.0 ppb   |
| Trace Impurities - Selenium (Se)                             | ≤ 50.0 ppb    | < 0.1 ppb   |
| Trace Impurities - Silicon (Si)                              | ≤ 100.0 ppb   | 31.5 ppb    |
| Trace Impurities - Silver (Ag)                               | ≤ 1.0 ppb     | < 0.3 ppb   |

>>> Continued on page 2 >>>

Sulfuric Acid  
BAKER INSTRA-ANALYZED® Reagent  
For Trace Metal Analysis  
Low Selenium



Material No.: 9673-33  
Batch No.: 23D2462010

| Test                              | Specification    | Result    |
|-----------------------------------|------------------|-----------|
| Trace Impurities – Sodium (Na)    | $\leq 500.0$ ppb | 5.4 ppb   |
| Trace Impurities – Strontium (Sr) | $\leq 5.0$ ppb   | < 0.2 ppb |
| Trace Impurities – Tin (Sn)       | $\leq 5.0$ ppb   | < 0.8 ppb |
| Trace Impurities – Zinc (Zn)      | $\leq 5.0$ ppb   | 0.4 ppb   |

For Laboratory, Research, or Manufacturing Use

Country of Origin: USA  
Packaging Site: Phillipsburg Mfg Ctr & DC

  
Jamie Ethier  
Vice President Global Quality

Certificate of Analysis

**ThermoFisher**  
SCIENTIFIC

## Certificate of Analysis

1 Reagent Lane

Fair Lawn, NJ 07410

201.796.7100 tel

201.796.1329 fax

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System  
Standard ISO9001:2015 by SAI Global Certificate Number CERT – 0120633

This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Certain data have been supplied by third parties. Thermo Fisher Scientific expressly disclaims all warranties, expressed or implied, including the implied warranties of merchantability and fitness for a particular purpose. Products are for research use or further manufacturing. Not for direct administration to humans or animals. It is the responsibility of the final formulator and end user to determine suitability based upon the intended use of the end product. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained.

|                   |                                 |                             |            |
|-------------------|---------------------------------|-----------------------------|------------|
| Catalog Number    | SA226                           | Quality Test / Release Date | 03/18/2024 |
| Lot Number        | 235420                          |                             |            |
| Description       | SULFURIC ACID, 0.02N, CERTIFIED |                             |            |
| Country of Origin | United States                   | Suggested Retest Date       | Mar/2029   |

| N/A                       |                        |                                   |                         |
|---------------------------|------------------------|-----------------------------------|-------------------------|
| Result Name               | Units                  | Specifications                    | Test Value              |
| APPEARANCE                |                        | REPORT                            | Clear, colorless liquid |
| COLOR                     | APHA                   | <= 5                              | <5                      |
| IDENTIFICATION            | PASS/FAIL              | = PASS TEST                       | PASS TEST               |
| NORMALITY                 |                        | Inclusive Between 0.0198 - 0.0202 | 0.0200                  |
| TRACEABLE TO NIST KHP STD | POT. ACID<br>PHTHALATE | = LOT 84L                         | SRM 84I                 |



Harout Sahagian - Quality Control Manager - Fair Lawn

Note: The data listed is valid for all package sizes of this lot of this product, expressed as an extension of this catalog number listed above.  
If there are any questions with this certificate, please call at (800) 227-6701.

\*Based on suggested storage condition.



W3163 Rec. on 12/10/24 by IZ

# Certificate of Analysis



Material BDH9284-2.5KG  
Material Description BDH SODIUM CARB ANHYD ACS 2.5KG  
Grade U S P REAGENT (ACS GRADE)

Batch 24E3156178  
Reassay Date 09/30/2027  
CAS Number 497-19-8  
Molecular Formula Na<sub>2</sub>CO<sub>3</sub>  
Molecular Mass 105.99

Date of Manufacture 09/01/2023  
Storage Room Temperature  
Material is hygroscopic. Protect from Moisture.  
Additional Product Description:

| Characteristics      | Specifications   | Measured Values            |
|----------------------|--|----------------------------|
| Appearance           | Fine white granular powder                                 | Fine white granular powder |
| Calcium              | <= 0.03 %  | 0.003 %                    |
| Chloride             | <= 0.001 %   | 0.0003 %                   |
| Heavy Metals (as Pb) | <= 0.0005 %  | 0.0001 %                   |
| Insolubles           | <= 0.01 %  | 0.001 %                    |
| Iron                 | <= 0.0005 %  | 0.0001 %                   |
| Loss on Heating      | <= 1.0 %   | 0.03 %                     |
| Magnesium            | <= 0.005 %   | 0.001 %                    |
| Phosphate            | <= 0.001 %   | 0.001 %                    |
| Potassium            | <= 0.005 %   | 0.003 %                    |
| Purity               | >= 99.5 %  | 100.0 %                    |
| Silica               | <= 0.005 %   | 0.001 %                    |
| Sulfur Compounds     | <= 0.003 %   | 0.002 %                    |
| Extra Description:   | Meets Reagent Specifications for testing USP/NF monographs |                            |

Internal ID #: 710

| Signature  | Additional Information  |
|--|---|
| We certify that this batch conforms to the specifications listed above.<br><br>This document has been electronically produced and is valid without a signature.<br><br>Leona Edwardson, Quality Control Sr. Manager - Solon<br>VWR Chemicals, LLC.<br>28600 Fountain Parkway, Solon OH 44139 USA | Analysis may have been rounded to significant digits in specification limits<br><br>Product meets analytical specifications of the grades listed. |



# Certificate of Analysis

021758 58

Buffer, Reference Standard, pH 4.00 ± 0.01 at 25°C (Color Coded Red)

Lot Number: 2411A93

Product Number: 1501

Manufacture Date: NOV 04, 2024

Expiration Date: OCT 2026

The certified value for this product is confirmed in independent testing by a second qualified chemist.  
The NIST Traceable pH value is certified to ±0.01 at 25 °C only. All other pH values at their corresponding temperatures are accurate to ± 0.05.

|    |      |      |      |      |      |      |      |      |      |      |      |
|----|------|------|------|------|------|------|------|------|------|------|------|
| °C | 0    | 5    | 10   | 15   | 20   | 25   | 30   | 35   | 40   | 45   | 50   |
| pH | 4.00 | 4.00 | 4.00 | 4.00 | 4.00 | 4.00 | 4.01 | 4.02 | 4.03 | 4.04 | 4.06 |

| Name                     | CAS#        | Grade           |
|--------------------------|-------------|-----------------|
| Water                    | 7732-18-5   | ACS/ASTM/USP/EP |
| Potassium Acid Phthalate | 877-24-7    | Buffer          |
| Preservative             | Proprietary | Commercial      |
| Red Dye                  | Proprietary | Purified        |

| Test       | Specification | Result |
|------------|---------------|--------|
| Appearance | Red liquid    | Passed |

\*Not a certified value.

| Test                                  | Certified Value | Uncertainty | NIST SRM#               |
|---------------------------------------|-----------------|-------------|-------------------------|
| pH at 25°C (Method: SQCP027, SQCP033) | 4.008           | 0.02        | 185i, 186-I-g, 186-II-g |

| Specification               | Reference       |
|-----------------------------|-----------------|
| Commercial Buffer Solutions |                 |
| Buffer B                    | ASTM (D 1293 B) |
| Buffer B                    | ASTM (D 5464)   |
| Buffer B                    | ASTM (D 5128)   |

pH measurements were performed in our Pocomoke City, MD laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387.01) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution. Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

| Part Number | Size / Package Type | Shelf Life (Unopened Container) |
|-------------|---------------------|---------------------------------|
| 1501-16     | 500 mL natural poly | 24 months                       |
| 1501-2.5    | 10 L Cubitainer®    | 24 months                       |
| 1501-5      | 20 L Cubitainer®    | 24 months                       |

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Refine your results. Redefine your industry.

# Certificate of Analysis

300 Technology Drive  
Christiansburg, VA 24073 USA  
inorganicventures.com

P: 800-669-6799/540-585-3030  
F: 540-585-3012  
info@inorganicventures.com

## 1.0 ACCREDITATION / REGISTRATION

**INORGANIC VENTURES** is accredited to ISO 17034, "General Requirements for the Competence of Reference Material Producers" and ISO/IEC 17025, "General Requirements for the Competence of Testing and Calibration Laboratories". Inorganic Ventures is also an ISO 9001 registered manufacturer (QSR Certificate Number QSR-1034).



## 2.0 PRODUCT DESCRIPTION

Product Code: Multi Analyte Ion Chromatography Solution  
Catalog Number: 300-CAL-A  
Lot Number: V2-MEB742616  
Matrix: H<sub>2</sub>O  
Value / Analyte(s):  
150 µg/mL ea:  
Sulfate,  
100 µg/mL ea:  
Bromide,  
50 µg/mL ea:  
o-Phosphate as P,  
30 µg/mL ea:  
Chloride, Nitrite as N,  
25 µg/mL ea:  
Nitrate as N,  
20 µg/mL ea:  
Fluoride

## 3.0 CERTIFIED VALUES AND UNCERTAINTIES

| ANALYTE                          | CERTIFIED VALUE    | ANALYTE                            | CERTIFIED VALUE    |
|----------------------------------|--------------------|------------------------------------|--------------------|
| Bromide, Br                      | 100.0 ± 0.5 µg/mL  | Chloride, Cl                       | 30.01 ± 0.13 µg/mL |
| Fluoride, F-                     | 20.00 ± 0.07 µg/mL | Nitrate as N, NNO <sub>3</sub> -   | 25.00 ± 0.10 µg/mL |
| Nitrite as N, NNO <sub>2</sub> - | 30.00 ± 0.10 µg/mL | o-Phosphate as P, PPO <sub>4</sub> | 50.00 ± 0.18 µg/mL |
| Sulfate, SO <sub>4</sub>         | 150.0 ± 0.8 µg/mL  |                                    |                    |

Density: 0.999 g/mL (measured at 20 ± 4 °C)

Assay Information:

| ANALYTE | METHOD     | NIST SRM#        | SRM LOT#   |
|---------|------------|------------------|------------|
| Br      | IC Assay   | 3184             | 151130     |
| Br      | Fajans     | 999c             | 999c       |
| Cl      | IC Assay   | 3182             | 190830     |
| Cl      | Fajans     | 999c             | 999c       |
| F-      | IC Assay   | 3183             | 140203     |
| NNO3-   | IC Assay   | 3185             | 170309     |
| NNO2-   | IC Assay   | Traceable to 40H | 08228TH-H2 |
| NNO2-   | Calculated | 40h              | 40h        |
| PPO4    | IC Assay   | 3186             | 170606     |
| SO4     | IC Assay   | 3181             | 080603     |

- The Calculated Value is a value calculated from the weight of a starting material that has been certified directly vs. a National Institute of Standards and Technology (NIST) SRM/RM. See Sec 4.2 for balance traceability.

The following equations are used in the calculation of the certified value and the uncertainty. Reported uncertainties represent expanded uncertainties expressed at approximately the 95% confidence level using a coverage factor of k = 2.

#### Characterization of CRM/RM by Two or More Methods

Certified Value,  $X_{CRM/RM}$ , where two or more methods of characterization are used is the weighted mean of the results:

$$X_{CRM/RM} = \sum(w_i)(X_i)$$

$X_i$  = mean of Assay Method i with standard uncertainty  $u_{char i}$

$w_i$  = the weighting factors for each method calculated using the inverse square of the variance:

$$w_i = (1/u_{char i}^2) / (\sum(1/u_{char i}^2))$$

$$CRM/RM \text{ Expanded Uncertainty } (\pm) = U_{CRM/RM} = k(u_{char}^2 + u_{bb}^2 + u_{lts}^2 + u_{ts}^2)^{1/2}$$

k = coverage factor = 2

$u_{char} = [\sum(w_i)^2(u_{char i}^2)]^{1/2}$  where  $u_{char i}$  are the errors from each characterization method

$u_{bb}$  = bottle to bottle homogeneity standard uncertainty

$u_{lts}$  = long term stability standard uncertainty (storage)

$u_{ts}$  = transport stability standard uncertainty

#### Characterization of CRM/RM by One Method

Certified Value,  $X_{CRM/RM}$ , where one method of characterization is used is the mean of individual results:

$$X_{CRM/RM} = (X_a)(u_{char a})$$

$X_a$  = mean of Assay Method A with

$u_{char a}$  = the standard uncertainty of characterization Method A

$$CRM/RM \text{ Expanded Uncertainty } (\pm) = U_{CRM/RM} = k(u_{char a}^2 + u_{bb}^2 + u_{lts}^2 + u_{ts}^2)^{1/2}$$

k = coverage factor = 2

$u_{char a}$  = the errors from characterization

$u_{bb}$  = bottle to bottle homogeneity standard uncertainty

$u_{lts}$  = long term stability standard uncertainty (storage)

$u_{ts}$  = transport stability standard uncertainty

## 4.0 TRACEABILITY TO NIST

- This product is traceable to NIST via an unbroken chain of comparisons. The uncertainties for each certified value are reported, taking into account the SRM/RM uncertainty error and the measurement, weighing and volume dilution errors. In rare cases where no NIST SRM/RM are available, the term 'in-house std.' is specified.

### 4.1 Thermometer Calibration

- All thermometers are NIST traceable through thermometers that are calibrated by an accredited calibration laboratory.

### 4.2 Balance Calibration

- All analytical balances are calibrated by an accredited calibration laboratory and procedure. The weights used for testing are annually compared to master weights and are traceable to NIST.

### 4.3 Glassware Calibration

- An in-house procedure is used to calibrate all Class A glassware used in the manufacturing and quality control of CRM/RMs.

## 5.0 CHROMATOGRAM

N/A

## 6.0 INTENDED USE

**6.1** This standard is intended for the calibration of analytical instruments and validation of analytical methods as appropriate. This CRM may be used in connection with EPA Methods 6010, 6020 (all versions), Standard Methods 3120 B and USP <232> / ICH Q3D.

**6.2** For products attaining traceability through Inorganic Ventures' Primary Certified Reference Materials (PCRM™) see the Limited License to Use PCRM™ in the Inorganic Ventures [Terms and Conditions of Sale](https://www.inorganicventures.com/terms-and-conditions-sale). The Terms and Conditions contain information on the use of materials traceable to PCRM™ certified reference materials. This Limited License agreement is especially pertinent for laboratories accredited under ISO:17034.

## **7.0 INSTRUCTIONS FOR THE CORRECT USE OF THIS REFERENCE MATERIAL**

### **7.1 Storage and Handling Recommendations**

- Store between approximately 4° - 30° C while in sealed TCT bag.
- While stored in the sealed TCT bag, transpiration of this CRM/RM is negligible. After opening the sealed TCT bag transpiration of the CRM/RM will occur, resulting in a gradual increase in the analyte concentration(s). It is the responsibility of the user to account for this effect. When the bottle is weighed both before and after being placed in storage, the mass difference observed will be a measure of transpiration mass loss.
- After opening the sealed TCT bag, keep cap tightly sealed when not in use and store between 4° - 24° C to minimize the effects of transpiration. Use at 20° ± 4° C to minimize volumetric dilution error when using the reported density. Do not pipette from the container. Do not return removed aliquots to container.

- For more information, visit [www.inorganicventures.com/TCT](http://www.inorganicventures.com/TCT)

## **8.0 HAZARDOUS INFORMATION**

- Please refer to the Safety Data Sheet for information regarding this CRM/RM.

## **9.0 HOMOGENEITY**

- This solution was mixed according to an in-house procedure and is guaranteed to be homogeneous. Homogeneity data indicate that the end user should take a minimum sample size of 0.2 mL to assure homogeneity.

## **10.0 QUALITY STANDARD DOCUMENTATION**

### **10.1 ISO 9001 Quality Management System Registration**

- QSR Certificate Number QSR-1034

### **10.2 ISO/IEC 17025 "General Requirements for the Competence of Testing and Calibration Laboratories"**

- Chemical Testing - Accredited / A2LA Certificate Number 883.01

### **10.3 ISO 17034 "General Requirements for the Competence of Reference Material Producers"**

- Reference Material Producer - Accredited / A2LA Certificate Number 883.02

Inorganic Ventures, 300 Technology Drive, Christiansburg, Va. 24073, USA; Telephone: 800.669.6799; 540.585.3030, Fax: 540.585.3012; [inorganicventures.com](http://inorganicventures.com); [info@inorganicventures.com](mailto:info@inorganicventures.com)

## **11.0 CERTIFICATION, LOT EXPIRATION AND PERIOD OF VALIDITY**

### 11.1 Certification Issue Date

April 02, 2024

- The certification is valid within the measurement uncertainty specified provided the CRM/RM is stored and handled in accordance with instructions given in Sec 7.1. This certification is nullified if instructions in Sec 7.1 are not followed or if the CRM/RM is damaged, contaminated, or otherwise modified.

### 11.2 Lot Expiration Date

- **April 02, 2029**

- The date after which this CRM/RM should not be used.

- The lot expiration date reflects the period of time that the stability of a CRM/RM can be supported by long term stability studies conducted on properly stored and handled CRM/RMs. Lot expiration is limited primarily by transpiration (loss of water from the solution) and infrequently by chemical stability.

### 11.3 Period of Validity

- Sealed TCT Bag Open Date: \_\_\_\_\_

- This CRM/RM should not be used longer than one year (or six months in the case of a 30 mL bottle) from the date of opening the aluminized bag or after the date given in Sec. 11.2, whichever comes first. This is contingent upon the CRM/RM being stored and handled in accordance with the instructions given in Sec. 7.1.

## 12.0 NAMES AND SIGNATURES OF CERTIFYING OFFICERS

### Certificate Prepared By:

Uyen Truong  
Custom Processing Supervisor



### Certificate Approved By:

Thomas Kozikowski  
Stock VS Manager



### Certifying Officer:

Paul Gaines  
Chairman / Senior Technical Director



| Ident         | Instrument IC-1 |         |         |         | Analyst: NF |          | Method: 300.0 / 9056A |            | Initial Analyst |                 |          |
|---------------|-----------------|---------|---------|---------|-------------|----------|-----------------------|------------|-----------------|-----------------|----------|
|               | Con F-          | Con CL- | Con NO2 | Con BR- | Con NO3     | Con HPO4 | Con SO4               |            |                 |                 |          |
| STD1          | 0               | 0       | 0       | 0       | 0           | 0        | 0                     | 0          | IC1-032125      | 3/21/2025 10:45 | 10 NF/IZ |
| STD2          | 0.421           | 0.619   | 0.631   | 2.075   | 0.523       | 1.052    | 3.247                 | IC1-032125 | 3/21/2025 11:07 | 10 NF/IZ        |          |
| STD3          | 0.795           | 1.199   | 1.203   | 3.994   | 1.001       | 1.993    | 5.998                 | IC1-032125 | 3/21/2025 11:28 | 10 NF/IZ        |          |
| STD4          | 0.977           | 1.475   | 1.468   | 4.904   | 1.226       | 2.407    | 7.216                 | IC1-032125 | 3/21/2025 11:50 | 10 NF/IZ        |          |
| STD5          | 1.993           | 3.009   | 2.995   | 10.03   | 2.493       | 4.968    | 14.842                | IC1-032125 | 3/21/2025 12:11 | 10 NF/IZ        |          |
| STD6          | 4.034           | 5.988   | 5.986   | 19.975  | 5.011       | 10.256   | 30.502                | IC1-032125 | 3/21/2025 12:32 | 10 NF/IZ        |          |
| STD7          | 4.979           | 7.51    | 7.517   | 25.022  | 6.247       | 12.323   | 36.695                | IC1-032125 | 3/21/2025 12:54 | 10 NF/IZ        |          |
| ICV           | 2.031           | 3.038   | 3.08    | 10.364  | 2.568       | 5.199    | 15.269                | IC1-032125 | 3/21/2025 13:37 | 10 NF/IZ        |          |
| ICB           | 0               | 0.122   | 0.08    | 0       | 0           | 0        | 0                     | IC1-032125 | 3/21/2025 13:58 | 10 NF/IZ        |          |
| CCV           | 2.034           | 3.119   | 3.093   | 10.383  | 2.559       | 5.188    | 15.179                | IC1-032125 | 4/3/2025 9:38   | 10 NF/IZ        |          |
| CCB           | 0               | 0       | 0       | 0       | 0           | 0        | 0                     | IC1-032125 | 4/3/2025 9:59   | 10 NF/IZ        |          |
| LB135296BSW   | 2.028           | 3.121   | 3.103   | 10.387  | 2.561       | 5.226    | 15.185                | IC1-032125 | 4/3/2025 10:43  | 10 NF/IZ        |          |
| LB135296BLW   | 0               | 0       | 0       | 0       | 0           | 0        | 0                     | IC1-032125 | 4/3/2025 11:04  | 10 NF/IZ        |          |
| Q1711-01      | 0.332           | 26.59   | 0       | 0.215   | 0           | 2.108    | 10.666                | IC1-032125 | 4/3/2025 12:29  | 10 NF/IZ        |          |
| Q1711-02MS    | 2.301           | 28.881  | 3.076   | 10.43   | 2.571       | 2.367    | 24.892                | IC1-032125 | 4/3/2025 12:50  | 10 NF/IZ        |          |
| Q1711-03MSD   | 2.172           | 28.857  | 2.983   | 10.101  | 2.479       | 1.573    | 24.485                | IC1-032125 | 4/3/2025 13:12  | 10 NF/IZ        |          |
| Q1711-04      | 0.275           | 20.1    | 0       | 0.221   | 0           | 0        | 4.529                 | IC1-032125 | 4/3/2025 13:55  | 10 NF/IZ        |          |
| Q1711-08      | 0               | 0.047   | 0       | 0       | 0           | 0        | 0                     | IC1-032125 | 4/3/2025 14:16  | 10 NF/IZ        |          |
| Q1716-01      | 0.433           | 427.687 | 0       | 0.63    | 0.296       | 0        | 54.697                | IC1-032125 | 4/3/2025 14:38  | 10 NF/IZ        |          |
| Q1711-01DLX10 | 0.054           | 2.364   | 0       | 0       | 0           | 0        | 1.413                 | IC1-032125 | 4/3/2025 14:59  | 10 NF/IZ        |          |
| Q1711-04DLX5  | 0.071           | 3.687   | 0       | 0       | 0           | 0        | 1.238                 | IC1-032125 | 4/3/2025 15:21  | 10 NF/IZ        |          |
| CCV           | 2.072           | 3.107   | 3.112   | 10.418  | 2.589       | 5.237    | 15.386                | IC1-032125 | 4/3/2025 15:43  | 10 NF/IZ        |          |
| CCB           | 0               | 0       | 0       | 0       | 0           | 0        | 0                     | IC1-032125 | 4/3/2025 17:16  | 10 NF/IZ        |          |

# Certificate of Analysis

**Buffer, Reference Standard, pH 7.00 ± 0.01 at 25°C (Color Coded Yellow)**

**Lot Number:** 2504D34

**Product Number:** 1551

**Manufacture Date:** APR 03, 2025

**Expiration Date:** MAR 2027

The certified value for this product is confirmed in independent testing by a second qualified chemist.

The NIST traceable pH value is certified to ±0.01 at 25 °C only. All other pH values at their corresponding temperatures are accurate to ± 0.05.

|    |      |      |      |      |      |      |      |      |      |      |      |
|----|------|------|------|------|------|------|------|------|------|------|------|
| °C | 0    | 5    | 10   | 15   | 20   | 25   | 30   | 35   | 40   | 45   | 50   |
| pH | 7.12 | 7.09 | 7.06 | 7.04 | 7.02 | 7.00 | 6.99 | 6.98 | 6.98 | 6.97 | 6.97 |

| Name                           | CAS#        | Grade              |
|--------------------------------|-------------|--------------------|
| Water                          | 7732-18-5   | ACS/ASTM/USP/EP    |
| Sodium Phosphate Dibasic       | 7558-79-4   | ACS                |
| Potassium Dihydrogen Phosphate | 7778-77-0   | ACS                |
| Preservative                   | Proprietary |                    |
| Yellow Dye                     | Proprietary |                    |
| Sodium Hydroxide               | 1310-73-2   | Reagent (from ACS) |

| Test       | Specification | Result                         |
|------------|---------------|--------------------------------|
| Appearance | Yellow liquid | Passed *Not a certified value. |

| Test                                  | Certified Value | Uncertainty | NIST SRM#               |
|---------------------------------------|-----------------|-------------|-------------------------|
| pH at 25°C (Method: SQCP027, SQCP033) | 7.003           | 0.02        | 186-I-g, 186-II-g, 191d |

| Specification               | Reference       |
|-----------------------------|-----------------|
| Commercial Buffer Solutions | ASTM (D 1293 B) |
| Buffer A                    | ASTM (D 5464)   |
| Buffer A                    | ASTM (D 5128)   |

pH measurements were performed in our Pocomoke City, MD laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387.01) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution. Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

| Part Number | Size / Package Type | Shelf Life (Unopened Container) |
|-------------|---------------------|---------------------------------|
| 1551-2.5    | 10 L Cubitainer®    | 24 months                       |
| 1551-20     | 20 x 20 mL pack     | 24 months                       |
| 1551-32     | 1 L natural poly    | 24 months                       |
| 1551-5      | 20 L Cubitainer®    | 24 months                       |

**Recommended Storage:** 15°C - 30°C (59°F - 86°F)



Jose Pena (04/03/2025)  
Operations Manager

**This product was tested in an ISO 17025 Accredited Laboratory**

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.