

Duplicate Sample Summary

Client:	G Environmental	SDG No.:	Q3737
Project:	Buff	Sample ID:	Q3737-01
Client ID:	MW3DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Alkalinity	mg/L	+/-20	268		267		1	0		12/03/2025

Duplicate Sample Summary

Client: G Environmental	SDG No.: Q3737
Project: Buff	Sample ID: Q3737-02
Client ID: MW4MSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Fluoride	mg/L	+/-20	2.30		2.30		1	0		12/02/2025
Chloride	mg/L	+/-20	19.6	OR	19.5	OR	1	1		12/02/2025
Sulfate	mg/L	+/-20	1180	OR	1160	OR	1	2		12/02/2025
Nitrate	mg/L	+/-20	3.80		3.70		1	3		12/02/2025
Nitrite	mg/L	+/-20	3.40		3.30		1	3		12/02/2025
Bromide	mg/L	+/-20	10.9		10.5		1	4		12/02/2025
Orthophosphate as P	mg/L	+/-20	5.40		5.70		1	5		12/02/2025