

Laboratory Control Sample Summary

Client: G Environmental

SDG No.: Q3737

Project: Buff

Run No.: LB138084

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138084BSW							
Bromide	mg/L	10	10.7		107	1	90-110	12/02/2025
Chloride	mg/L	3	3.20		107	1	90-110	12/02/2025
Fluoride	mg/L	2	2.10		105	1	90-110	12/02/2025
Nitrite	mg/L	3	3.20		107	1	90-110	12/02/2025
Nitrate	mg/L	2.5	2.70		108	1	90-110	12/02/2025
Sulfate	mg/L	15	16.1		107	1	90-110	12/02/2025
Orthophosphate as P	mg/L	5	5.50		110	1	90-110	12/02/2025

Laboratory Control Sample Summary

Client: G Environmental

SDG No.: Q3737

Project: Buff

Run No.: LB138090

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138090BSW							
Alkalinity	mg/L	50	46.8		94	1	80-120	12/03/2025