



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: G Environmental

SDG No.: Q3737

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Iron	86.8	100	87	65 - 135	P	12/01/2025	13:50	LB138072
	Manganese	21.5	20.0	107	65 - 135	P	12/01/2025	13:50	LB138072
	Sodium	1990	2000	100	65 - 135	P	12/01/2025	13:50	LB138072