

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>GENV01</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3738</u>
Instrument ID:	<u>MSVOA_N</u>	Calibration Date/Time:	<u>12/02/2025</u> <u>10:00</u>
Lab File ID:	<u>VN088396.D</u>	Init. Calib. Date(s):	<u>11/24/2025</u> <u>11/24/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>12:38</u> <u>14:35</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.751	0.786		4.66	20
Chloromethane	0.807	0.845	0.1	4.71	20
Vinyl Chloride	0.821	0.843		2.68	20
Bromomethane	0.395	0.429		8.61	20
Chloroethane	0.508	0.539		6.1	20
Trichlorofluoromethane	1.153	1.181		2.43	20
1,1,2-Trichlorotrifluoroethane	0.614	0.613		-0.16	20
Tert butyl alcohol	0.150	0.138		-8	20
1,1-Dichloroethene	0.610	0.615		0.82	20
Acetone	0.341	0.307		-9.97	20
Carbon Disulfide	1.993	2.027		1.71	20
Methyl tert-butyl Ether	2.285	2.558		11.95	20
Methyl Acetate	1.014	1.074		5.92	20
Methylene Chloride	0.729	0.758		3.98	20
trans-1,2-Dichloroethene	0.676	0.712		5.32	20
1,1-Dichloroethane	1.367	1.456	0.1	6.51	20
Cyclohexane	1.239	1.232		-0.56	20
2-Butanone	0.548	0.563		2.74	20
Carbon Tetrachloride	0.532	0.564		6.01	20
cis-1,2-Dichloroethene	0.785	0.832		5.99	20
Bromochloromethane	0.681	0.699		2.64	20
Chloroform	1.343	1.462		8.86	20
1,1,1-Trichloroethane	1.181	1.247		5.59	20
Methylcyclohexane	0.572	0.574		0.35	20
Benzene	1.570	1.657		5.54	20
1,2-Dichloroethane	0.599	0.667		11.35	20
Trichloroethene	0.371	0.369		-0.54	20
1,2-Dichloropropane	0.402	0.431		7.21	20
Bromodichloromethane	0.586	0.638		8.87	20
4-Methyl-2-Pentanone	0.633	0.694		9.64	20
Toluene	0.906	1.005		10.93	20
t-1,3-Dichloropropene	0.601	0.670		11.48	20
cis-1,3-Dichloropropene	0.632	0.699		10.6	20
1,1,2-Trichloroethane	0.351	0.376		7.12	20
2-Hexanone	0.399	0.450		12.78	20
Dibromochloromethane	0.400	0.439		9.75	20
1,2-Dibromoethane	0.355	0.382		7.61	20
Tetrachloroethene	0.396	0.355		-10.35	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>GENV01</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3738</u>
Instrument ID:	<u>MSVOA_N</u>	Calibration Date/Time:	<u>12/02/2025</u> <u>10:00</u>
Lab File ID:	<u>VN088396.D</u>	Init. Calib. Date(s):	<u>11/24/2025</u> <u>11/24/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>12:38</u> <u>14:35</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Chlorobenzene	1.123	1.184	0.3	5.43	20
Ethyl Benzene	1.963	2.113		7.64	20
m/p-Xylenes	0.728	0.798		9.61	20
o-Xylene	0.694	0.758		9.22	20
Styrene	1.127	1.318		16.95	20
Bromoform	0.277	0.305	0.1	10.11	20
Isopropylbenzene	3.697	4.066		9.98	20
1,1,2,2-Tetrachloroethane	1.195	1.245	0.3	4.18	20
1,3-Dichlorobenzene	1.647	1.752		6.38	20
1,4-Dichlorobenzene	1.725	1.800		4.35	20
1,2-Dichlorobenzene	1.550	1.664		7.36	20
1,2-Dibromo-3-Chloropropane	0.290	0.303		4.48	20
1,2,4-Trichlorobenzene	0.921	0.953		3.47	20
1,2,3-Trichlorobenzene	0.896	0.919		2.57	20
1,2-Dichloroethane-d4	0.942	0.990		5.1	20
Dibromofluoromethane	0.346	0.360		4.05	20
Toluene-d8	1.289	1.336		3.65	20
4-Bromofluorobenzene	0.442	0.474		7.24	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.