

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120125\
 Data File : VN088374.D
 Acq On : 01 Dec 2025 12:10
 Operator : JC\MD
 Sample : VN1201WBSD01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1201WBSD01

Manual Integrations
 APPROVED

Reviewed By : John Carlone 12/02/2025
 Supervised By : Mahesh Dadoda 12/02/2025

Quant Time: Dec 02 00:11:04 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112425W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 25 01:34:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	257123	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	472668	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	413438	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	196888	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	252803	52.192	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 104.380%			
35) Dibromofluoromethane	8.147	113	171251	52.281	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 104.560%			
50) Toluene-d8	10.541	98	648629	53.220	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 106.440%			
62) 4-Bromofluorobenzene	12.829	95	213607	51.080	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 102.160%			
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	2.142	85	72877	18.870	ug/l	98
3) Chloromethane	2.383	50	80587	19.429	ug/l	94
4) Vinyl Chloride	2.536	62	79060	18.736	ug/l	99
5) Bromomethane	2.989	94	40980	20.178	ug/l	99
6) Chloroethane	3.148	64	50269	19.250	ug/l	90
7) Trichlorofluoromethane	3.512	101	110683	18.663	ug/l	99
8) Diethyl Ether	3.965	74	42809	19.231	ug/l	98
9) 1,1,2-Trichlorotrifluo...	4.377	101	58334	18.476	ug/l	98
10) Methyl Iodide	4.583	142	60569	19.379	ug/l	96
11) Tert butyl alcohol	5.512	59	68039	87.983	ug/l #	81
12) 1,1-Dichloroethene	4.347	96	58831	18.754	ug/l	84
13) Acrolein	4.177	56	87619	118.346	ug/l	97
14) Allyl chloride	5.012	41	104505	17.974	ug/l	98
15) Acrylonitrile	5.706	53	213646	95.755	ug/l	98
16) Acetone	4.418	43	152332	86.961	ug/l	98
17) Carbon Disulfide	4.706	76	192813	18.810	ug/l	100
18) Methyl Acetate	5.018	43	99171	19.019	ug/l	99
19) Methyl tert-butyl Ether	5.783	73	228498	19.442	ug/l	95
20) Methylene Chloride	5.259	84	75693	20.196	ug/l	83
21) trans-1,2-Dichloroethene	5.771	96	68139	19.608	ug/l	92
22) Diisopropyl ether	6.653	45	244953	19.733	ug/l	96
23) Vinyl Acetate	6.588	43	972820m	97.271	ug/l	
24) 1,1-Dichloroethane	6.553	63	134006	19.063	ug/l	95
25) 2-Butanone	7.465	43	270811	96.165	ug/l	97
26) 2,2-Dichloropropane	7.477	77	112288	18.506	ug/l	95
27) cis-1,2-Dichloroethene	7.465	96	78314	19.392	ug/l	96
28) Bromochloromethane	7.794	49	64631	18.456	ug/l	99
29) Tetrahydrofuran	7.824	42	193749	95.919	ug/l	99
30) Chloroform	7.947	83	135282	19.585	ug/l	96
31) Cyclohexane	8.235	56	117687	18.475	ug/l	98
32) 1,1,1-Trichloroethane	8.147	97	115347	18.992	ug/l	95
36) 1,1-Dichloropropene	8.353	75	89956	17.794	ug/l	98
37) Ethyl Acetate	7.547	43	123427	18.569	ug/l	98
38) Carbon Tetrachloride	8.347	117	94617	18.801	ug/l	96
39) Methylcyclohexane	9.576	83	99985	18.491	ug/l	93
40) Benzene	8.588	78	284116	19.149	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120125\
 Data File : VN088374.D
 Acq On : 01 Dec 2025 12:10
 Operator : JC\MD
 Sample : VN1201WBSD01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1201WBSD01

Manual Integrations
 APPROVED

Reviewed By :John Carlone 12/02/2025
 Supervised By :Mahesh Dadoda 12/02/2025

Quant Time: Dec 02 00:11:04 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112425W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 25 01:34:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.765	41	63105	19.168	ug/l	98
42) 1,2-Dichloroethane	8.647	62	109703	19.380	ug/l	98
43) Isopropyl Acetate	8.671	43	192472	18.952	ug/l	99
44) Trichloroethene	9.329	130	66801	19.054	ug/l	90
45) 1,2-Dichloropropane	9.600	63	72843	19.162	ug/l	100
46) Dibromomethane	9.688	93	54401	20.053	ug/l	98
47) Bromodichloromethane	9.865	83	106876	19.308	ug/l #	93
48) Methyl methacrylate	9.659	41	84615	18.260	ug/l	99
49) 1,4-Dioxane	9.676	88	22017	373.614	ug/l #	99
51) 4-Methyl-2-Pentanone	10.424	43	589357	98.453	ug/l	99
52) Toluene	10.606	92	170024	19.846	ug/l	98
53) t-1,3-Dichloropropene	10.818	75	108520	19.099	ug/l	99
54) cis-1,3-Dichloropropene	10.288	75	116758	19.541	ug/l	97
55) 1,1,2-Trichloroethane	10.994	97	66233	19.982	ug/l	95
56) Ethyl methacrylate	10.853	69	104863	20.028	ug/l	97
57) 1,3-Dichloropropane	11.141	76	120668	20.179	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.141	63	276569	91.622	ug/l	99
59) 2-Hexanone	11.176	43	355826	94.358	ug/l	98
60) Dibromochloromethane	11.335	129	76699	20.298	ug/l	98
61) 1,2-Dibromoethane	11.447	107	66826	19.909	ug/l	97
64) Tetrachloroethene	11.076	164	62208	18.980	ug/l	95
65) Chlorobenzene	11.870	112	183680	19.779	ug/l	96
66) 1,1,1,2-Tetrachloroethane	11.941	131	61851	19.747	ug/l	99
67) Ethyl Benzene	11.941	91	306102	18.859	ug/l	97
68) m/p-Xylenes	12.053	106	227659	37.830	ug/l	99
69) o-Xylene	12.376	106	109896	19.155	ug/l	95
70) Styrene	12.388	104	181083	19.438	ug/l	99
71) Bromoform	12.559	173	46125	20.106	ug/l #	99
73) Isopropylbenzene	12.670	105	280946	19.299	ug/l	100
74) N-amyl acetate	12.682	43	97186m	19.603	ug/l	
75) 1,1,2,2-Tetrachloroethane	12.917	83	91931	19.534	ug/l	97
76) 1,2,3-Trichloropropane	12.970	75	93113m	20.192	ug/l	
77) Bromobenzene	12.959	156	65049	19.529	ug/l	96
78) n-propylbenzene	13.012	91	348515	19.569	ug/l	100
79) 2-Chlorotoluene	13.100	91	204279	18.783	ug/l	99
80) 1,3,5-Trimethylbenzene	13.153	105	234071	19.404	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.717	75	28199	17.348	ug/l #	96
82) 4-Chlorotoluene	13.200	91	211139	19.546	ug/l	97
83) tert-Butylbenzene	13.412	119	192736	18.795	ug/l	99
84) 1,2,4-Trimethylbenzene	13.459	105	235984	19.682	ug/l	100
85) sec-Butylbenzene	13.594	105	283289	19.230	ug/l	100
86) p-Isopropyltoluene	13.706	119	228290	18.898	ug/l	98
87) 1,3-Dichlorobenzene	13.712	146	126657	19.534	ug/l	98
88) 1,4-Dichlorobenzene	13.788	146	129333	19.042	ug/l	98
89) n-Butylbenzene	14.029	91	219415	18.707	ug/l	99
90) Hexachloroethane	14.306	117	41810	18.946	ug/l	96
91) 1,2-Dichlorobenzene	14.082	146	115702	18.962	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.700	75	21103	18.481	ug/l	98
93) 1,2,4-Trichlorobenzene	15.370	180	65316	18.016	ug/l	95
94) Hexachlorobutadiene	15.470	225	24556	18.849	ug/l	94
95) Naphthalene	15.611	128	231154	18.400	ug/l	99
96) 1,2,3-Trichlorobenzene	15.811	180	67438	19.114	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120125\
Data File : VN088374.D
Acq On : 01 Dec 2025 12:10
Operator : JC\MD
Sample : VN1201WBSD01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1201WBSD01

Manual Integrations
APPROVED

Reviewed By :John Carlone 12/02/2025
Supervised By :Mahesh Dadoda 12/02/2025

Quant Time: Dec 02 00:11:04 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112425W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 25 01:34:52 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Manual Integrations

Reviewed By :John Carlone 12/02/2025
Supervised By :Mahesh Dadoda 12/02/2025

[illegible]