

Surrogate Summary

SDG No.: **Q3738**

Client: **G Environmental**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3738-01	MW4	1,2-Dichloroethane-d4	50	56.4	113		70 (74)	130 (125)
		Dibromofluoromethane	50	50.9	102		70 (75)	130 (124)
		Toluene-d8	50	50.6	101		70 (86)	130 (113)
		4-Bromofluorobenzene	50	53.3	107		70 (77)	130 (121)
Q3738-02	MW5	1,2-Dichloroethane-d4	50	58.9	118		70 (74)	130 (125)
		Dibromofluoromethane	50	51.6	103		70 (75)	130 (124)
		Toluene-d8	50	51.3	103		70 (86)	130 (113)
		4-Bromofluorobenzene	50	56.6	113		70 (77)	130 (121)
Q3738-03	Field	1,2-Dichloroethane-d4	50	57.8	116		70 (74)	130 (125)
		Dibromofluoromethane	50	52.3	105		70 (75)	130 (124)
		Toluene-d8	50	50.3	101		70 (86)	130 (113)
		4-Bromofluorobenzene	50	51.7	103		70 (77)	130 (121)
VN1201WBL01	VN1201WBL01	1,2-Dichloroethane-d4	50	57.8	116		70 (74)	130 (125)
		Dibromofluoromethane	50	52.2	104		70 (75)	130 (124)
		Toluene-d8	50	50.0	100		70 (86)	130 (113)
		4-Bromofluorobenzene	50	52.4	105		70 (77)	130 (121)
VN1201WBS01	VN1201WBS01	1,2-Dichloroethane-d4	50	56.0	112		70 (74)	130 (125)
		Dibromofluoromethane	50	55.1	110		70 (75)	130 (124)
		Toluene-d8	50	54.9	110		70 (86)	130 (113)
		4-Bromofluorobenzene	50	53.3	107		70 (77)	130 (121)
VN1201WBSD0	VN1201WBSD01	1,2-Dichloroethane-d4	50	52.2	104		70 (74)	130 (125)
		Dibromofluoromethane	50	52.3	105		70 (75)	130 (124)
		Toluene-d8	50	53.2	106		70 (86)	130 (113)
		4-Bromofluorobenzene	50	51.1	102		70 (77)	130 (121)
VN1202WBL01	VN1202WBL01	1,2-Dichloroethane-d4	50	61.1	122		70 (74)	130 (125)
		Dibromofluoromethane	50	52.4	105		70 (75)	130 (124)
		Toluene-d8	50	50.1	100		70 (86)	130 (113)
		4-Bromofluorobenzene	50	51.8	104		70 (77)	130 (121)
VN1202WBS02	VN1202WBS02	1,2-Dichloroethane-d4	50	52.1	104		70 (74)	130 (125)
		Dibromofluoromethane	50	52.1	104		70 (75)	130 (124)
		Toluene-d8	50	50.5	101		70 (86)	130 (113)
		4-Bromofluorobenzene	50	49.3	99		70 (77)	130 (121)

