

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01
 Lab Code: ACE SDG NO.: Q3738
 Lab File ID: VN088396.D Date Analyzed: 12/02/2025
 Instrument ID: MSVOA_N Time Analyzed: 10:00
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	256486	8.20	469714	9.08	423596	11.84
UPPER LIMIT	512972	8.7	939428	9.582	847192	12.341
LOWER LIMIT	128243	7.7	234857	8.582	211798	11.341
EPA SAMPLE NO.						
MW4	183621	8.21	418628	9.08	407692	11.84
VN1202WBL01	181798	8.21	415636	9.08	401801	11.84
VN1202WBS02	245156	8.21	446464	9.08	396309	11.84

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.