

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>GENV01</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3738</u>
Lab File ID:	<u>VN088370.D</u>	Date Analyzed:	<u>12/01/2025</u>
Instrument ID:	<u>MSVOA_N</u>	Time Analyzed:	<u>09:58</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)
		Heated Purge:	(Y/N) <u>N</u>

	IS4 AREA #	RT #				
12 HOUR STD	197460	13.765				
UPPER LIMIT	394920	14.265				
LOWER LIMIT	98730	13.265				
EPA SAMPLE NO.						
MW5	246573	13.77				
Field	185899	13.77				
VN1201WBL01	198799	13.77				
VN1201WBS01	207303	13.77				
VN1201WBSD01	196888	13.77				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.