

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01  
 Lab Code: ACE SDG NO.: Q3738  
 Lab File ID: VN088396.D Date Analyzed: 12/02/2025  
 Instrument ID: MSVOA\_N Time Analyzed: 10:00  
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

|                | IS4<br>AREA # | RT #   |  |  |  |  |
|----------------|---------------|--------|--|--|--|--|
| 12 HOUR STD    | 206439        | 13.764 |  |  |  |  |
| UPPER LIMIT    | 412878        | 14.264 |  |  |  |  |
| LOWER LIMIT    | 103220        | 13.264 |  |  |  |  |
| EPA SAMPLE NO. |               |        |  |  |  |  |
| MW4            | 190700        | 13.77  |  |  |  |  |
| VN1202WBL01    | 185686        | 13.77  |  |  |  |  |
| VN1202WBS02    | 188408        | 13.77  |  |  |  |  |

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.