

Initial and Continuing Calibration Verification

Client: G Environmental

SDG No.: Q3739

Project: Diamond

RunNo.: LB138067

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.472	0.5	94	90-110	12/01/2025
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.476	0.5	95	90-110	12/01/2025
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.474	0.5	95	90-110	12/01/2025
Sample ID: CCV3 Hexavalent Chromium	mg/L	0.471	0.5	94	90-110	12/01/2025

Initial and Continuing Calibration Verification

Client: G Environmental

SDG No.: Q3739

Project: Diamond

RunNo.: LB138067

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------