

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB138062

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	12/01/2025	08:10
2	CAL2	1	Water	NA	NA	20.2	7.00	12/01/2025	08:11
3	CAL3	1	Water	NA	NA	20.3	10.02	12/01/2025	08:14
4	ICV	1	Water	NA	NA	20.3	7.00	12/01/2025	08:15
5	CCV1	1	Water	NA	NA	20.2	2.01	12/01/2025	08:19
6	Q3720-01	1	Solid	20.02	20	22.6	6.78	12/01/2025	08:35
7	Q3735-01	1	Solid	20.03	20	22.9	6.80	12/01/2025	08:37
8	Q3740-01	1	Solid	20.02	20	23.2	6.06	12/01/2025	08:47
9	Q3740-01DUP	1	Solid	20.03	20	23.5	6.08	12/01/2025	08:50
10	CCV2	1	Water	NA	NA	20.3	12.02	12/01/2025	08:52

WORKLIST(Hardcopy Internal Chain)

VP 138062

WorkList Name : PH W Q3731 WorkList ID : 193376 Department : Wet-Chemistry Date : 12-01-2025 07:47:46

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3720-01	BUR-25-0059	Solid	pH	Cool 4 deg C	PSEG03	E42	11/25/2025	9045D
Q3735-01	BU-3-112625	Solid	pH	Cool 4 deg C	PSEG05	A11	11/26/2025	9045D
Q3740-01	WC1	Solid	pH	Cool 4 deg C	GENV01	--Sele	11/26/2025	9045D

Date/Time 12/01/25 07:55
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 12/01/25 12:30
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]