

Initial and Continuing Calibration Verification

Client: G Environmental

SDG No.: Q3740

Project: Diamond

RunNo.: LB138062

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV pH	pH	7.00	7	100	90-110	12/01/2025
Sample ID: CCV1 pH	pH	2.01	2.00	101	90-110	12/01/2025
Sample ID: CCV2 pH	pH	12.02	12.00	100	90-110	12/01/2025

Initial and Continuing Calibration Verification

Client: G Environmental

SDG No.: Q3740

Project: Diamond

RunNo.: LB138063

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
Corrosivity		pH	7.00	7	100	90-110	12/01/2025
Sample ID:	CCV1						
Corrosivity		pH	2.01	2.00	101	90-110	12/01/2025
Sample ID:	CCV2						
Corrosivity		pH	12.02	12.00	100	90-110	12/01/2025

Initial and Continuing Calibration Verification

Client: G Environmental

SDG No.: Q3740

Project: Diamond

RunNo.: LB138067

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.472	0.5	94	90-110	12/01/2025
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.476	0.5	95	90-110	12/01/2025
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.474	0.5	95	90-110	12/01/2025
Sample ID: CCV3 Hexavalent Chromium	mg/L	0.471	0.5	94	90-110	12/01/2025

Initial and Continuing Calibration Verification

Client: G Environmental

SDG No.: Q3740

Project: Diamond

RunNo.: LB138094

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Reactive Cyanide	mg/L	0.094	0.099	95	85-115	12/03/2025
Sample ID: CCV1 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	12/03/2025
Sample ID: CCV2 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	12/03/2025

Initial and Continuing Calibration Verification

Client: G Environmental

SDG No.: Q3740

Project: Diamond

RunNo.: LB138094

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------