

Duplicate Sample Summary

Client:	G Environmental	SDG No.:	Q3740
Project:	Diamond	Sample ID:	Q3733-01
Client ID:	SU-04-11-26-2025DUP	Percent Solids for Spike Sample:	91.5

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Hexavalent Chromium	mg/Kg	+/-20	0.075	U	0.075	U	1	0		12/01/2025

Duplicate Sample Summary

Client: G Environmental	SDG No.: Q3740
Project: Diamond	Sample ID: Q3740-01
Client ID: WC1DUP	Percent Solids for Spike Sample: 100

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
pH	pH	+/-20	6.06		6.08		1	0.33		12/01/2025
Corrosivity	pH	+/-20	6.06		6.08		1	0.33		12/01/2025
Reactive Sulfide	mg/Kg	+/-20	1.57	J	1.57	J	1	0		12/01/2025
Ignitability	oC	+/-20	NO		NO		1	0		12/03/2025
Reactive Cyanide	mg/Kg	+/-20	0.016	J	0.016	J	1	0		12/03/2025