

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: G Environmental

SDG No.: Q3740

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Aluminum	116	100	116	65 - 135	P	12/01/2025	13:50	LB138072
	Antimony	51.9	50.0	104	65 - 135	P	12/01/2025	13:50	LB138072
	Arsenic	22.7	20.0	114	65 - 135	P	12/01/2025	13:50	LB138072
	Barium	102	100	102	65 - 135	P	12/01/2025	13:50	LB138072
	Beryllium	6.43	6.0	107	65 - 135	P	12/01/2025	13:50	LB138072
	Cadmium	6.21	6.0	104	65 - 135	P	12/01/2025	13:50	LB138072
	Calcium	2090	2000	104	65 - 135	P	12/01/2025	13:50	LB138072
	Chromium	9.87	10.0	99	65 - 135	P	12/01/2025	13:50	LB138072
	Cobalt	30.5	30.0	102	65 - 135	P	12/01/2025	13:50	LB138072
	Copper	22.0	20.0	110	65 - 135	P	12/01/2025	13:50	LB138072
	Iron	86.8	100	87	65 - 135	P	12/01/2025	13:50	LB138072
	Lead	13.2	12.0	110	65 - 135	P	12/01/2025	13:50	LB138072
	Magnesium	2170	2000	108	65 - 135	P	12/01/2025	13:50	LB138072
	Manganese	21.5	20.0	107	65 - 135	P	12/01/2025	13:50	LB138072
	Nickel	41.6	40.0	104	65 - 135	P	12/01/2025	13:50	LB138072
	Potassium	2140	2000	107	65 - 135	P	12/01/2025	13:50	LB138072
	Selenium	22.4	20.0	112	65 - 135	P	12/01/2025	13:50	LB138072
	Silver	10.2	10.0	102	65 - 135	P	12/01/2025	13:50	LB138072
	Sodium	1990	2000	100	65 - 135	P	12/01/2025	13:50	LB138072
	Thallium	37.8	40.0	95	65 - 135	P	12/01/2025	13:50	LB138072
	Vanadium	39.7	40.0	99	65 - 135	P	12/01/2025	13:50	LB138072
	Zinc	44.3	40.0	111	65 - 135	P	12/01/2025	13:50	LB138072
CRI01	Aluminum	108	100	108	65 - 135	P	12/02/2025	11:42	LB138085
	Antimony	52.1	50.0	104	65 - 135	P	12/02/2025	11:42	LB138085
	Arsenic	20.2	20.0	101	65 - 135	P	12/02/2025	11:42	LB138085
	Barium	102	100	102	65 - 135	P	12/02/2025	11:42	LB138085
	Beryllium	6.67	6.0	111	65 - 135	P	12/02/2025	11:42	LB138085
	Cadmium	6.34	6.0	106	65 - 135	P	12/02/2025	11:42	LB138085
	Calcium	2050	2000	102	65 - 135	P	12/02/2025	11:42	LB138085
	Chromium	10.6	10.0	106	65 - 135	P	12/02/2025	11:42	LB138085
	Cobalt	30.0	30.0	100	65 - 135	P	12/02/2025	11:42	LB138085
	Copper	21.5	20.0	108	65 - 135	P	12/02/2025	11:42	LB138085
	Iron	110	100	110	65 - 135	P	12/02/2025	11:42	LB138085
	Lead	10.2	12.0	85	65 - 135	P	12/02/2025	11:42	LB138085
	Magnesium	2160	2000	108	65 - 135	P	12/02/2025	11:42	LB138085
	Manganese	21.6	20.0	108	65 - 135	P	12/02/2025	11:42	LB138085

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: G Environmental

SDG No.: Q3740

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Nickel	41.1	40.0	103	65 - 135	P	12/02/2025	11:42	LB138085
	Potassium	1980	2000	99	65 - 135	P	12/02/2025	11:42	LB138085
	Selenium	19.3	20.0	96	65 - 135	P	12/02/2025	11:42	LB138085
	Silver	10.1	10.0	101	65 - 135	P	12/02/2025	11:42	LB138085
	Sodium	1990	2000	100	65 - 135	P	12/02/2025	11:42	LB138085
	Thallium	49.9	40.0	125	65 - 135	P	12/02/2025	11:42	LB138085
	Vanadium	42.0	40.0	105	65 - 135	P	12/02/2025	11:42	LB138085
	Zinc	41.5	40.0	104	65 - 135	P	12/02/2025	11:42	LB138085
CRA	Mercury	0.20	0.2	99	70 - 130	CV	12/02/2025	12:48	LB138081