

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

SDG No.: Q3740

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV132	Mercury	3.68	4.0	92	90 - 110	CV	12/01/2025	14:54	1b138068

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

SDG No.: Q3740

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV44	Mercury	4.92	5.0	98	90 - 110	CV	12/01/2025	15:02	1b138068
CCV45	Mercury	4.99	5.0	100	90 - 110	CV	12/01/2025	15:34	1b138068
CCV46	Mercury	4.99	5.0	100	90 - 110	CV	12/01/2025	16:06	1b138068
CCV47	Mercury	4.92	5.0	98	90 - 110	CV	12/01/2025	16:24	1b138068

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

SDG No.: Q3740

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Arsenic	4020	4000	101	90- 110	P	12/01/2025	13:16	LB138072
	Barium	7850	8000	98	90- 110	P	12/01/2025	13:16	LB138072
	Cadmium	1940	2000	97	90- 110	P	12/01/2025	13:16	LB138072
	Chromium	791	800	99	90- 110	P	12/01/2025	13:16	LB138072
	Lead	3830	4000	96	90- 110	P	12/01/2025	13:16	LB138072
	Selenium	4190	4000	105	90- 110	P	12/01/2025	13:16	LB138072
	Silver	1000	1000	100	90- 110	P	12/01/2025	13:16	LB138072

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

SDG No.: Q3740

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Arsenic	23.8	20.0	119	80 - 120	P	12/01/2025	13:42	LB138072
	Barium	105	100	105	80 - 120	P	12/01/2025	13:42	LB138072
	Cadmium	6.23	6.0	104	80 - 120	P	12/01/2025	13:42	LB138072
	Chromium	9.35	10.0	94	80 - 120	P	12/01/2025	13:42	LB138072
	Lead	11.5	12.0	96	80 - 120	P	12/01/2025	13:42	LB138072
	Selenium	20.6	20.0	103	80 - 120	P	12/01/2025	13:42	LB138072
	Silver	9.97	10.0	100	80 - 120	P	12/01/2025	13:42	LB138072

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

SDG No.: Q3740

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Arsenic	5040	5000	101	90 - 110	P	12/01/2025	14:30	LB138072
	Barium	9960	10000	100	90 - 110	P	12/01/2025	14:30	LB138072
	Cadmium	2470	2500	99	90 - 110	P	12/01/2025	14:30	LB138072
	Chromium	994	1000	99	90 - 110	P	12/01/2025	14:30	LB138072
	Lead	4970	5000	100	90 - 110	P	12/01/2025	14:30	LB138072
	Selenium	5140	5000	103	90 - 110	P	12/01/2025	14:30	LB138072
	Silver	1270	1250	102	90 - 110	P	12/01/2025	14:30	LB138072
CCV02	Arsenic	4990	5000	100	90 - 110	P	12/01/2025	15:26	LB138072
	Barium	10200	10000	102	90 - 110	P	12/01/2025	15:26	LB138072
	Cadmium	2410	2500	97	90 - 110	P	12/01/2025	15:26	LB138072
	Chromium	958	1000	96	90 - 110	P	12/01/2025	15:26	LB138072
	Lead	4870	5000	98	90 - 110	P	12/01/2025	15:26	LB138072
	Selenium	5090	5000	102	90 - 110	P	12/01/2025	15:26	LB138072
	Silver	1250	1250	100	90 - 110	P	12/01/2025	15:26	LB138072
CCV03	Arsenic	5100	5000	102	90 - 110	P	12/01/2025	16:33	LB138072
	Barium	10000	10000	100	90 - 110	P	12/01/2025	16:33	LB138072
	Cadmium	2450	2500	98	90 - 110	P	12/01/2025	16:33	LB138072
	Chromium	995	1000	100	90 - 110	P	12/01/2025	16:33	LB138072
	Lead	4940	5000	99	90 - 110	P	12/01/2025	16:33	LB138072
	Selenium	5180	5000	104	90 - 110	P	12/01/2025	16:33	LB138072
	Silver	1270	1250	102	90 - 110	P	12/01/2025	16:33	LB138072
CCV04	Arsenic	5140	5000	103	90 - 110	P	12/01/2025	17:25	LB138072
	Barium	10500	10000	104	90 - 110	P	12/01/2025	17:25	LB138072
	Cadmium	2500	2500	100	90 - 110	P	12/01/2025	17:25	LB138072
	Chromium	1010	1000	100	90 - 110	P	12/01/2025	17:25	LB138072
	Lead	5050	5000	101	90 - 110	P	12/01/2025	17:25	LB138072
	Selenium	5140	5000	103	90 - 110	P	12/01/2025	17:25	LB138072
	Silver	1310	1250	105	90 - 110	P	12/01/2025	17:25	LB138072
CCV05	Arsenic	5110	5000	102	90 - 110	P	12/01/2025	18:16	LB138072
	Barium	9960	10000	100	90 - 110	P	12/01/2025	18:16	LB138072
	Cadmium	2500	2500	100	90 - 110	P	12/01/2025	18:16	LB138072
	Chromium	1030	1000	103	90 - 110	P	12/01/2025	18:16	LB138072
	Lead	5020	5000	100	90 - 110	P	12/01/2025	18:16	LB138072
	Selenium	5160	5000	103	90 - 110	P	12/01/2025	18:16	LB138072

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

SDG No.: Q3740

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Silver	1300	1250	104	90 - 110	P	12/01/2025	18:16	LB138072
CCV06	Arsenic	5230	5000	105	90 - 110	P	12/01/2025	19:07	LB138072
	Barium	10300	10000	102	90 - 110	P	12/01/2025	19:07	LB138072
	Cadmium	2540	2500	102	90 - 110	P	12/01/2025	19:07	LB138072
	Chromium	1020	1000	102	90 - 110	P	12/01/2025	19:07	LB138072
	Lead	5110	5000	102	90 - 110	P	12/01/2025	19:07	LB138072
	Selenium	5320	5000	106	90 - 110	P	12/01/2025	19:07	LB138072
	Silver	1310	1250	105	90 - 110	P	12/01/2025	19:07	LB138072
CCV07	Arsenic	5410	5000	108	90 - 110	P	12/01/2025	19:24	LB138072
	Barium	10300	10000	103	90 - 110	P	12/01/2025	19:24	LB138072
	Cadmium	2630	2500	105	90 - 110	P	12/01/2025	19:24	LB138072
	Chromium	1040	1000	104	90 - 110	P	12/01/2025	19:24	LB138072
	Lead	5280	5000	106	90 - 110	P	12/01/2025	19:24	LB138072
	Selenium	5530	5000	110	90 - 110	P	12/01/2025	19:24	LB138072
	Silver	1340	1250	107	90 - 110	P	12/01/2025	19:24	LB138072

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

SDG No.: Q3740

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Arsenic	3880	4000	97	90 - 110	P	12/02/2025	11:29	LB138085
	Barium	8280	8000	104	90 - 110	P	12/02/2025	11:29	LB138085
	Cadmium	1920	2000	96	90 - 110	P	12/02/2025	11:29	LB138085
	Chromium	781	800	98	90 - 110	P	12/02/2025	11:29	LB138085
	Lead	3790	4000	95	90 - 110	P	12/02/2025	11:29	LB138085
	Selenium	3970	4000	99	90 - 110	P	12/02/2025	11:29	LB138085
	Silver	1080	1000	108	90 - 110	P	12/02/2025	11:29	LB138085

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

SDG No.: Q3740

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Arsenic	22.6	20.0	113	80 - 120	P	12/02/2025	11:34	LB138085
	Barium	104	100	104	80 - 120	P	12/02/2025	11:34	LB138085
	Cadmium	6.22	6.0	104	80 - 120	P	12/02/2025	11:34	LB138085
	Chromium	9.91	10.0	99	80 - 120	P	12/02/2025	11:34	LB138085
	Lead	11.3	12.0	94	80 - 120	P	12/02/2025	11:34	LB138085
	Selenium	18.8	20.0	94	80 - 120	P	12/02/2025	11:34	LB138085
	Silver	10.3	10.0	103	80 - 120	P	12/02/2025	11:34	LB138085

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

SDG No.: Q3740

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Arsenic	4950	5000	99	90 - 110	P	12/02/2025	12:10	LB138085
	Barium	10200	10000	102	90 - 110	P	12/02/2025	12:10	LB138085
	Cadmium	2470	2500	99	90 - 110	P	12/02/2025	12:10	LB138085
	Chromium	1000	1000	100	90 - 110	P	12/02/2025	12:10	LB138085
	Lead	4950	5000	99	90 - 110	P	12/02/2025	12:10	LB138085
	Selenium	4980	5000	100	90 - 110	P	12/02/2025	12:10	LB138085
	Silver	1250	1250	100	90 - 110	P	12/02/2025	12:10	LB138085
CCV02	Arsenic	5050	5000	101	90 - 110	P	12/02/2025	13:13	LB138085
	Barium	9890	10000	99	90 - 110	P	12/02/2025	13:13	LB138085
	Cadmium	2500	2500	100	90 - 110	P	12/02/2025	13:13	LB138085
	Chromium	1020	1000	102	90 - 110	P	12/02/2025	13:13	LB138085
	Lead	5020	5000	100	90 - 110	P	12/02/2025	13:13	LB138085
	Selenium	5180	5000	104	90 - 110	P	12/02/2025	13:13	LB138085
	Silver	1280	1250	103	90 - 110	P	12/02/2025	13:13	LB138085
CCV03	Arsenic	5040	5000	101	90 - 110	P	12/02/2025	14:11	LB138085
	Barium	10100	10000	102	90 - 110	P	12/02/2025	14:11	LB138085
	Cadmium	2470	2500	99	90 - 110	P	12/02/2025	14:11	LB138085
	Chromium	987	1000	99	90 - 110	P	12/02/2025	14:11	LB138085
	Lead	4970	5000	100	90 - 110	P	12/02/2025	14:11	LB138085
	Selenium	5230	5000	105	90 - 110	P	12/02/2025	14:11	LB138085
	Silver	1260	1250	100	90 - 110	P	12/02/2025	14:11	LB138085
CCV04	Arsenic	5140	5000	103	90 - 110	P	12/02/2025	15:15	LB138085
	Barium	10500	10000	105	90 - 110	P	12/02/2025	15:15	LB138085
	Cadmium	2500	2500	100	90 - 110	P	12/02/2025	15:15	LB138085
	Chromium	990	1000	99	90 - 110	P	12/02/2025	15:15	LB138085
	Lead	5050	5000	101	90 - 110	P	12/02/2025	15:15	LB138085
	Selenium	5330	5000	107	90 - 110	P	12/02/2025	15:15	LB138085
	Silver	1280	1250	102	90 - 110	P	12/02/2025	15:15	LB138085
CCV05	Arsenic	5040	5000	101	90 - 110	P	12/02/2025	16:05	LB138085
	Barium	10600	10000	106	90 - 110	P	12/02/2025	16:05	LB138085
	Cadmium	2460	2500	98	90 - 110	P	12/02/2025	16:05	LB138085
	Chromium	968	1000	97	90 - 110	P	12/02/2025	16:05	LB138085
	Lead	4950	5000	99	90 - 110	P	12/02/2025	16:05	LB138085
	Selenium	5220	5000	104	90 - 110	P	12/02/2025	16:05	LB138085

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

SDG No.: Q3740

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Silver	1250	1250	100	90 - 110	P	12/02/2025	16:05	LB138085
CCV06	Arsenic	5050	5000	101	90 - 110	P	12/02/2025	16:55	LB138085
	Barium	10400	10000	104	90 - 110	P	12/02/2025	16:55	LB138085
	Cadmium	2470	2500	99	90 - 110	P	12/02/2025	16:55	LB138085
	Chromium	983	1000	98	90 - 110	P	12/02/2025	16:55	LB138085
	Lead	5000	5000	100	90 - 110	P	12/02/2025	16:55	LB138085
	Selenium	5220	5000	104	90 - 110	P	12/02/2025	16:55	LB138085
	Silver	1260	1250	101	90 - 110	P	12/02/2025	16:55	LB138085