

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: G Environmental

SDG No.: Q3740

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Arsenic	22.7	20.0	114	65 - 135	P	12/01/2025	13:50	LB138072
	Barium	102	100	102	65 - 135	P	12/01/2025	13:50	LB138072
	Cadmium	6.21	6.0	104	65 - 135	P	12/01/2025	13:50	LB138072
	Chromium	9.87	10.0	99	65 - 135	P	12/01/2025	13:50	LB138072
	Lead	13.2	12.0	110	65 - 135	P	12/01/2025	13:50	LB138072
	Selenium	22.4	20.0	112	65 - 135	P	12/01/2025	13:50	LB138072
	Silver	10.2	10.0	102	65 - 135	P	12/01/2025	13:50	LB138072
CRA	Mercury	0.15	0.2	73	70 - 130	CV	12/01/2025	15:09	lb138068
CRI01	Arsenic	20.2	20.0	101	65 - 135	P	12/02/2025	11:42	LB138085
	Barium	102	100	102	65 - 135	P	12/02/2025	11:42	LB138085
	Cadmium	6.34	6.0	106	65 - 135	P	12/02/2025	11:42	LB138085
	Chromium	10.6	10.0	106	65 - 135	P	12/02/2025	11:42	LB138085
	Lead	10.2	12.0	85	65 - 135	P	12/02/2025	11:42	LB138085
	Selenium	19.3	20.0	96	65 - 135	P	12/02/2025	11:42	LB138085
	Silver	10.1	10.0	101	65 - 135	P	12/02/2025	11:42	LB138085