

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: G Environmental

SDG No.: Q3740

Instrument: CV1

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB170754TB		WATER		Batch Number:	PB170770		Prep Date:	12/01/2025	
	Mercury	0.76	<2	U	2.00	CV	12/01/2025	16:13	lb138068
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB170770BL		WATER		Batch Number:	PB170770		Prep Date:	12/01/2025	
	Mercury	0.076	<0.2	U	0.20	CV	12/01/2025	15:19	lb138068

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: G Environmental

SDG No.: Q3740

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB170754TB		WATER		Batch Number:	PB170766		Prep Date:	12/01/2025	
	Arsenic	25.6	<50	U	100	P	12/02/2025	16:46	LB138085
	Barium	72.8	<250	U	500	P	12/02/2025	16:46	LB138085
	Cadmium	2.50	<15	U	30.0	P	12/02/2025	16:46	LB138085
	Chromium	24.4	<25	J	50.0	P	12/02/2025	16:46	LB138085
	Lead	11.5	<30	U	60.0	P	12/02/2025	16:46	LB138085
	Selenium	48.2	<50	U	100	P	12/02/2025	16:46	LB138085
	Silver	8.10	<25	U	50.0	P	12/02/2025	16:46	LB138085
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB170766BL		WATER		Batch Number:	PB170766		Prep Date:	12/01/2025	
	Arsenic	25.6	<50	U	100	P	12/01/2025	16:52	LB138072
	Barium	72.8	<250	U	500	P	12/01/2025	16:52	LB138072
	Cadmium	2.50	<15	U	30.0	P	12/01/2025	16:52	LB138072
	Chromium	10.6	<25	U	50.0	P	12/01/2025	16:52	LB138072
	Lead	11.5	<30	U	60.0	P	12/01/2025	16:52	LB138072
	Selenium	48.2	<50	U	100	P	12/01/2025	16:52	LB138072
	Silver	8.10	<25	U	50.0	P	12/01/2025	16:52	LB138072