

Continuing Calibration Report for SequenceID : FE120125AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE057097.D

Aliphatic C9-C12	11763898.000	60.000	3.201	6.825	196064.967	200655.681	2.288
Aliphatic C12-C16	7880968.000	40.000	6.826	10.270	197024.200	203177.346	3.028
Aliphatic C16-C21	11472211.000	60.000	10.271	13.644	191203.517	198315.768	3.586
Aliphatic C21-C28	14290009.000	80.000	13.645	17.313	178625.113	198747.045	10.124
Aliphatic C28-C40	18814402.000	120.000	17.314	22.242	156786.683	199899.890	21.567
Aliphatic EPH	64221488.000	360.000	3.201	22.242	178393.022	199869.809	10.745