

Continuing Calibration Report for SequenceID : FE120125AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE057115.D

Aliphatic C9-C12	13599714.000	60.000	3.201	6.825	226661.900	200655.681	-12.961
Aliphatic C12-C16	9082136.000	40.000	6.826	10.270	227053.400	203177.346	-11.751
Aliphatic C16-C21	13143542.000	60.000	10.271	13.644	219059.033	198315.768	-10.460
Aliphatic C21-C28	16552280.000	80.000	13.645	17.313	206903.500	198747.045	-4.104
Aliphatic C28-C40	27010035.000	120.000	17.314	22.242	225083.625	199899.890	-12.598
Aliphatic EPH	79387707.000	360.000	3.201	22.242	220521.408	199869.809	-10.333