

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** GENV01  
**Lab Code:** ACE **SDG NO.:** Q3740  
**GC Column:** ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

**Client Sample No.:** CCAL02 **Date Analyzed:** 12/01/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PO115479.D **Time Analyzed:** 16:17

| COMPOUND             | RT    | RT WINDOW |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|----------------------|-------|-----------|--------|--------------------|-------------------|-------|
|                      |       | FROM      | TO     |                    |                   |       |
| Aroclor-1016-1       | 5.473 | 5.387     | 5.587  | 484.180            | 500.000           | -3.2  |
| Aroclor-1016-2       | 5.496 | 5.408     | 5.608  | 474.620            | 500.000           | -5.1  |
| Aroclor-1016-3       | 5.557 | 5.470     | 5.670  | 478.340            | 500.000           | -4.3  |
| Aroclor-1016-4       | 5.653 | 5.566     | 5.766  | 463.340            | 500.000           | -7.3  |
| Aroclor-1016-5       | 5.944 | 5.858     | 6.058  | 482.710            | 500.000           | -3.5  |
| Aroclor-1260-1       | 7.057 | 6.971     | 7.171  | 443.840            | 500.000           | -11.2 |
| Aroclor-1260-2       | 7.311 | 7.225     | 7.425  | 492.280            | 500.000           | -1.5  |
| Aroclor-1260-3       | 7.669 | 7.582     | 7.782  | 480.450            | 500.000           | -3.9  |
| Aroclor-1260-4       | 7.890 | 7.803     | 8.003  | 468.020            | 500.000           | -6.4  |
| Aroclor-1260-5       | 8.197 | 8.110     | 8.310  | 460.700            | 500.000           | -7.9  |
| Decachlorobiphenyl   | 9.910 | 9.830     | 10.030 | 57.050             | 50.000            | 14.1  |
| Tetrachloro-m-xylene | 4.336 | 4.248     | 4.448  | 57.490             | 50.000            | 15.0  |

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** GENV01  
**Lab Code:** ACE **SDG NO.:** Q3740  
**GC Column:** ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

**Client Sample No.:** CCAL02 **Date Analyzed:** 12/01/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PO115479.D **Time Analyzed:** 16:17

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|----------------------|-------|-----------|-------|--------------------|-------------------|-------|
|                      |       | FROM      | TO    |                    |                   |       |
| Aroclor-1016-1       | 4.715 | 4.628     | 4.828 | 470.130            | 500.000           | -6.0  |
| Aroclor-1016-2       | 4.734 | 4.647     | 4.847 | 480.510            | 500.000           | -3.9  |
| Aroclor-1016-3       | 4.908 | 4.821     | 5.021 | 479.840            | 500.000           | -4.0  |
| Aroclor-1016-4       | 4.951 | 4.864     | 5.064 | 478.130            | 500.000           | -4.4  |
| Aroclor-1016-5       | 5.162 | 5.076     | 5.276 | 485.820            | 500.000           | -2.8  |
| Aroclor-1260-1       | 6.191 | 6.106     | 6.306 | 422.330            | 500.000           | -15.5 |
| Aroclor-1260-2       | 6.379 | 6.294     | 6.494 | 392.810            | 500.000           | -21.4 |
| Aroclor-1260-3       | 6.531 | 6.446     | 6.646 | 402.610            | 500.000           | -19.5 |
| Aroclor-1260-4       | 7.001 | 6.917     | 7.117 | 401.900            | 500.000           | -19.6 |
| Aroclor-1260-5       | 7.243 | 7.158     | 7.358 | 402.560            | 500.000           | -19.5 |
| Decachlorobiphenyl   | 8.622 | 8.541     | 8.741 | 52.890             | 50.000            | 5.8   |
| Tetrachloro-m-xylene | 3.640 | 3.548     | 3.748 | 54.200             | 50.000            | 8.4   |