

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3740

Client ID : SSTDCCC040

Date Analyzed: 12/01/2025

Lab File ID: BF144382.D

Time Analyzed: 12:41

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	58717	6.863	221118	8.15	119295	9.90
UPPER LIMIT	117434	7.363	442236	8.645	238590	10.404
LOWER LIMIT	29358.5	6.363	110559	7.645	59647.5	9.404
EPA SAMPLE NO.						
01 BU-3-112625MS	49538	6.86	162841	8.15	75668	9.90
02 BU-3-112625MSD	47026	6.86	155006	8.15	78597	9.90
03 WC1	52921	6.86	191478	8.14	101201	9.90
04 PB170763BL	58345	6.86	230427	8.15	129664	9.90
05 PB170763BS	56637	6.86	215029	8.15	112964	9.90

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.