

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>GENV01</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3740</u>
Instrument ID:	<u>MSVOA_Y</u>	Calibration Date/Time:	<u>12/01/2025 07:55</u>
Lab File ID:	<u>VY023829.D</u>	Init. Calib. Date(s):	<u>11/26/2025 11/26/2025</u>
Heated Purge: (Y/N)	<u>Y</u>	Init. Calib. Time(s):	<u>08:04 10:02</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.377	0.352		-6.63	20
Chloromethane	0.576	0.473	0.1	-17.88	20
Vinyl Chloride	0.624	0.570		-8.65	20
Bromomethane	0.440	0.360		-18.18	20
Chloroethane	0.406	0.380		-6.4	20
Trichlorofluoromethane	0.843	0.848		0.59	20
1,1,2-Trichlorotrifluoroethane	0.492	0.526		6.91	20
1,1-Dichloroethene	0.484	0.491		1.45	20
Acetone	0.140	0.158		12.86	20
Carbon Disulfide	1.562	1.340		-14.21	20
Methyl tert-butyl Ether	1.237	1.330		7.52	20
Methyl Acetate	0.271	0.304		12.18	20
Methylene Chloride	0.593	0.531		-10.45	20
trans-1,2-Dichloroethene	0.536	0.539		0.56	20
1,1-Dichloroethane	0.976	1.008	0.1	3.28	20
Cyclohexane	0.936	0.916		-2.14	20
2-Butanone	0.171	0.187		9.36	20
Carbon Tetrachloride	0.498	0.548		10.04	20
cis-1,2-Dichloroethene	0.622	0.639		2.73	20
Bromochloromethane	0.406	0.389		-4.19	20
Chloroform	0.981	1.019		3.87	20
1,1,1-Trichloroethane	0.871	0.916		5.17	20
Methylcyclohexane	0.600	0.648		8	20
Benzene	1.398	1.495		6.94	20
1,2-Dichloroethane	0.360	0.396		10	20
Trichloroethene	0.359	0.392		9.19	20
1,2-Dichloropropane	0.330	0.363		10	20
Bromodichloromethane	0.467	0.516		10.49	20
4-Methyl-2-Pentanone	0.210	0.252		20	20
Toluene	0.866	0.957		10.51	20
t-1,3-Dichloropropene	0.428	0.474		10.75	20
cis-1,3-Dichloropropene	0.512	0.561		9.57	20
1,1,2-Trichloroethane	0.240	0.275		14.58	20
2-Hexanone	0.155	0.188		21.29	20
Dibromochloromethane	0.321	0.361		12.46	20
1,2-Dibromoethane	0.222	0.249		12.16	20
Tetrachloroethene	0.474	0.532		12.24	20
Chlorobenzene	1.098	1.206	0.3	9.84	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>GENV01</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3740</u>
Instrument ID:	<u>MSVOA_Y</u>	Calibration Date/Time:	<u>12/01/2025</u> <u>07:55</u>
Lab File ID:	<u>VY023829.D</u>	Init. Calib. Date(s):	<u>11/26/2025</u> <u>11/26/2025</u>
Heated Purge: (Y/N)	<u>Y</u>	Init. Calib. Time(s):	<u>08:04</u> <u>10:02</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.875	2.141		14.19	20
m/p-Xylenes	0.722	0.815		12.88	20
o-Xylene	0.677	0.762		12.56	20
Styrene	1.116	1.275		14.25	20
Bromoform	0.217	0.246	0.1	13.36	20
Isopropylbenzene	3.549	4.062		14.45	20
1,1,2,2-Tetrachloroethane	0.584	0.664	0.3	13.7	20
1,3-Dichlorobenzene	1.673	1.834		9.62	20
1,4-Dichlorobenzene	1.670	1.788		7.07	20
1,2-Dichlorobenzene	1.463	1.606		9.77	20
1,2-Dibromo-3-Chloropropane	0.098	0.107		9.18	20
1,2,4-Trichlorobenzene	0.872	0.955		9.52	20
1,2,3-Trichlorobenzene	0.743	0.826		11.17	20
1,2-Dichloroethane-d4	0.487	0.517		6.16	20
Dibromofluoromethane	0.296	0.332		12.16	20
Toluene-d8	1.176	1.313		11.65	20
4-Bromofluorobenzene	0.387	0.437		12.92	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.