

Instrument ID: **MSVOA_Y**

Daily Analysis Runlog For Sequence/QC Batch ID # VY112625

Review By	Semsettin Yesilyurt	Review On	12/1/2025 9:50:11 AM
Supervise By	John Carlone	Supervise On	12/1/2025 2:45:01 PM
SubDirectory	VY112625	HP Acquire Method	MSVOA_Y
		HP Processing Method	82y112625s.m
STD. NAME	STD REF.#		
Tune/Reschk	VP136432		
Initial Calibration Stds	VP136433,VP136434,VP136435,VP136436,VP136437,VP136438		
CCC	VP136440		
Internal Standard/PEM	VP136146		
ICV/I.BLK	VP136439		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VY023808.D	26 Nov 2025 07:11	SY/MD	Ok
2	VSTDIC005	VY023809.D	26 Nov 2025 08:04	SY/MD	Ok,M
3	VSTDIC010	VY023810.D	26 Nov 2025 08:27	SY/MD	Ok,M
4	VSTDIC020	VY023811.D	26 Nov 2025 08:49	SY/MD	Ok,M
5	VSTDIC050	VY023812.D	26 Nov 2025 09:17	SY/MD	Ok,M
6	VSTDIC100	VY023813.D	26 Nov 2025 09:39	SY/MD	Ok,M
7	VSTDIC150	VY023814.D	26 Nov 2025 10:02	SY/MD	Ok,M
8	VIBLK	VY023815.D	26 Nov 2025 10:27	SY/MD	Ok
9	VSTDICV050	VY023816.D	26 Nov 2025 10:50	SY/MD	Ok,M
10	VY1126SBL01	VY023817.D	26 Nov 2025 11:13	SY/MD	Ok
11	VY1126SBS01	VY023818.D	26 Nov 2025 11:47	SY/MD	Ok,M
12	VY1126SBS01	VY023819.D	26 Nov 2025 12:09	SY/MD	Ok,M
13	Q3604-04	VY023820.D	26 Nov 2025 12:34	SY/MD	Not Ok
14	Q3604-05	VY023821.D	26 Nov 2025 12:57	SY/MD	Not Ok
15	VIBLK	VY023822.D	26 Nov 2025 13:21	SY/MD	Ok
16	Q3721-01	VY023823.D	26 Nov 2025 13:44	SY/MD	Ok
17	Q3723-01	VY023824.D	26 Nov 2025 14:08	SY/MD	Ok
18	Q3725-04	VY023825.D	26 Nov 2025 14:31	SY/MD	Ok
19	VIBLK	VY023826.D	26 Nov 2025 14:54	SY/MD	Ok
20	VSTDIC050	VY023827.D	26 Nov 2025 15:17	SY/MD	Ok,M

M : Manual Integration

Instrument ID: **MSVOA_Y**

Daily Analysis Runlog For Sequence/QCBatch ID # VY120125

Review By	John Carlone	Review On	12/2/2025 9:00:18 AM		
Supervise By	Mahesh Dadoda	Supervise On	12/2/2025 10:16:40 AM		
SubDirectory	VY120125	HP Acquire Method	MSVOA_Y	HP Processing Method	82y112625s.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP136454			
CCC		VP136455,VP136456			
Internal Standard/PEM		VP136146			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VY023828.D	01 Dec 2025 07:24	SY/MD	Ok
2	VSTDCCC050	VY023829.D	01 Dec 2025 07:55	SY/MD	Ok,M
3	VY1201SBL01	VY023830.D	01 Dec 2025 09:13	SY/MD	Ok
4	VY1201SBS01	VY023831.D	01 Dec 2025 09:37	SY/MD	Ok,M
5	VY1201SBSD01	VY023832.D	01 Dec 2025 10:00	SY/MD	Ok,M
6	Q3719-09RE	VY023833.D	01 Dec 2025 10:38	SY/MD	Confirms
7	Q3719-13	VY023834.D	01 Dec 2025 11:25	SY/MD	Ok
8	Q3735-01	VY023835.D	01 Dec 2025 12:02	SY/MD	ReRun
9	Q3740-01	VY023836.D	01 Dec 2025 12:26	SY/MD	Ok
10	Q3741-02	VY023837.D	01 Dec 2025 12:49	SY/MD	Ok
11	Q3741-03	VY023838.D	01 Dec 2025 13:12	SY/MD	Ok
12	Q3741-04	VY023839.D	01 Dec 2025 13:36	SY/MD	ReRun
13	Q3741-05	VY023840.D	01 Dec 2025 13:59	SY/MD	ReRun
14	Q3741-01	VY023841.D	01 Dec 2025 14:23	SY/MD	Ok
15	Q3741-06	VY023842.D	01 Dec 2025 14:46	SY/MD	ReRun
16	Q3742-01	VY023843.D	01 Dec 2025 15:10	SY/MD	Ok
17	Q3742-02	VY023844.D	01 Dec 2025 15:33	SY/MD	Ok
18	Q3742-03	VY023845.D	01 Dec 2025 15:57	SY/MD	Ok
19	Q3742-04	VY023846.D	01 Dec 2025 16:20	SY/MD	Ok
20	VSTDCCC050	VY023847.D	01 Dec 2025 17:47	SY/MD	Ok,M

M : Manual Integration