



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q3740
Client: G Environmental

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
-----------	-----------	--------	-----------	---------------	---	-----	-----	-------

Client ID: 0

Total Voc :
Total Concentration:

- A
- B
- C
- D
- E
- F
- G
- H
- I
- J



SAMPLE DATA

Report of Analysis

Client: G Environmental
Project: Diamond
Client Sample ID: WC1
Lab Sample ID: Q3740-01
Analytical Method: 8260D
Sample Wt/Vol: 4.98 g

Level: LOW
Final Vol: 5000 uL

Date Collected: 11/26/25
Date Received: 11/26/25
SDG No.: Q3740
Matrix: SOIL
% Solid: 90
Test: VOC-TCLVOA-10

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
75-71-8	Dichlorodifluoromethane	1.30	U	1	1.30	5.60	ug/Kg	12/01/25 12:26	VY120125
74-87-3	Chloromethane	1.30	U	1	1.30	5.60	ug/Kg	12/01/25 12:26	VY120125
75-01-4	Vinyl Chloride	0.88	U	1	0.88	5.60	ug/Kg	12/01/25 12:26	VY120125
74-83-9	Bromomethane	1.20	U	1	1.20	5.60	ug/Kg	12/01/25 12:26	VY120125
75-00-3	Chloroethane	1.40	U	1	1.40	5.60	ug/Kg	12/01/25 12:26	VY120125
75-69-4	Trichlorofluoromethane	1.30	U	1	1.30	5.60	ug/Kg	12/01/25 12:26	VY120125
76-13-1	1,1,2-Trichlorotrifluoroethane	1.20	U	1	1.20	5.60	ug/Kg	12/01/25 12:26	VY120125
75-35-4	1,1-Dichloroethene	1.10	U	1	1.10	5.60	ug/Kg	12/01/25 12:26	VY120125
67-64-1	Acetone	5.30	U	1	5.30	27.9	ug/Kg	12/01/25 12:26	VY120125
75-15-0	Carbon Disulfide	1.20	U	1	1.20	5.60	ug/Kg	12/01/25 12:26	VY120125
1634-04-4	Methyl tert-butyl Ether	0.81	U	1	0.81	5.60	ug/Kg	12/01/25 12:26	VY120125
79-20-9	Methyl Acetate	1.70	U	1	1.70	5.60	ug/Kg	12/01/25 12:26	VY120125
75-09-2	Methylene Chloride	3.90	U	1	3.90	11.2	ug/Kg	12/01/25 12:26	VY120125
156-60-5	trans-1,2-Dichloroethene	0.96	U	1	0.96	5.60	ug/Kg	12/01/25 12:26	VY120125
75-34-3	1,1-Dichloroethane	0.89	U	1	0.89	5.60	ug/Kg	12/01/25 12:26	VY120125
110-82-7	Cyclohexane	0.88	U	1	0.88	5.60	ug/Kg	12/01/25 12:26	VY120125
78-93-3	2-Butanone	7.30	U	1	7.30	27.9	ug/Kg	12/01/25 12:26	VY120125
56-23-5	Carbon Tetrachloride	1.10	U	1	1.10	5.60	ug/Kg	12/01/25 12:26	VY120125
156-59-2	cis-1,2-Dichloroethene	0.84	U	1	0.84	5.60	ug/Kg	12/01/25 12:26	VY120125
74-97-5	Bromochloromethane	1.30	U	1	1.30	5.60	ug/Kg	12/01/25 12:26	VY120125
67-66-3	Chloroform	0.94	U	1	0.94	5.60	ug/Kg	12/01/25 12:26	VY120125
71-55-6	1,1,1-Trichloroethane	1.00	U	1	1.00	5.60	ug/Kg	12/01/25 12:26	VY120125
108-87-2	Methylcyclohexane	1.00	U	1	1.00	5.60	ug/Kg	12/01/25 12:26	VY120125
71-43-2	Benzene	0.88	U	1	0.88	5.60	ug/Kg	12/01/25 12:26	VY120125
107-06-2	1,2-Dichloroethane	0.88	U	1	0.88	5.60	ug/Kg	12/01/25 12:26	VY120125
79-01-6	Trichloroethene	0.90	U	1	0.90	5.60	ug/Kg	12/01/25 12:26	VY120125
78-87-5	1,2-Dichloropropane	1.00	U	1	1.00	5.60	ug/Kg	12/01/25 12:26	VY120125
75-27-4	Bromodichloromethane	0.87	U	1	0.87	5.60	ug/Kg	12/01/25 12:26	VY120125
108-10-1	4-Methyl-2-Pentanone	4.00	U	1	4.00	27.9	ug/Kg	12/01/25 12:26	VY120125
108-88-3	Toluene	0.87	U	1	0.87	5.60	ug/Kg	12/01/25 12:26	VY120125
10061-02-6	t-1,3-Dichloropropene	0.73	U	1	0.73	5.60	ug/Kg	12/01/25 12:26	VY120125
10061-01-5	cis-1,3-Dichloropropene	0.69	U	1	0.69	5.60	ug/Kg	12/01/25 12:26	VY120125
79-00-5	1,1,2-Trichloroethane	1.00	U	1	1.00	5.60	ug/Kg	12/01/25 12:26	VY120125
591-78-6	2-Hexanone	4.10	U	1	4.10	27.9	ug/Kg	12/01/25 12:26	VY120125
124-48-1	Dibromochloromethane	0.97	U	1	0.97	5.60	ug/Kg	12/01/25 12:26	VY120125
106-93-4	1,2-Dibromoethane	0.98	U	1	0.98	5.60	ug/Kg	12/01/25 12:26	VY120125
127-18-4	Tetrachloroethene	1.20	U	1	1.20	5.60	ug/Kg	12/01/25 12:26	VY120125
108-90-7	Chlorobenzene	1.00	U	1	1.00	5.60	ug/Kg	12/01/25 12:26	VY120125
100-41-4	Ethyl Benzene	0.75	U	1	0.75	5.60	ug/Kg	12/01/25 12:26	VY120125
179601-23-1	m/p-Xylenes	1.40	U	1	1.40	11.2	ug/Kg	12/01/25 12:26	VY120125
95-47-6	o-Xylene	0.91	U	1	0.91	5.60	ug/Kg	12/01/25 12:26	VY120125
100-42-5	Styrene	0.79	U	1	0.79	5.60	ug/Kg	12/01/25 12:26	VY120125
75-25-2	Bromoform	0.96	U	1	0.96	5.60	ug/Kg	12/01/25 12:26	VY120125

Report of Analysis

Client: G Environmental
 Project: Diamond
 Client Sample ID: WC1
 Lab Sample ID: Q3740-01
 Analytical Method: 8260D
 Sample Wt/Vol: 4.98 g

Level: LOW
 Final Vol: 5000 uL

Date Collected: 11/26/25
 Date Received: 11/26/25
 SDG No.: Q3740
 Matrix: SOIL
 % Solid: 90
 Test: VOC-TCLVOA-10

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
98-82-8	Isopropylbenzene	0.87	U	1	0.87	5.60	ug/Kg	12/01/25 12:26	VY120125
79-34-5	1,1,2,2-Tetrachloroethane	1.30	U	1	1.30	5.60	ug/Kg	12/01/25 12:26	VY120125
541-73-1	1,3-Dichlorobenzene	1.90	U	1	1.90	5.60	ug/Kg	12/01/25 12:26	VY120125
106-46-7	1,4-Dichlorobenzene	1.70	U	1	1.70	5.60	ug/Kg	12/01/25 12:26	VY120125
95-50-1	1,2-Dichlorobenzene	1.60	U	1	1.60	5.60	ug/Kg	12/01/25 12:26	VY120125
96-12-8	1,2-Dibromo-3-Chloropropane	2.10	U	1	2.10	5.60	ug/Kg	12/01/25 12:26	VY120125
120-82-1	1,2,4-Trichlorobenzene	3.30	U	1	3.30	5.60	ug/Kg	12/01/25 12:26	VY120125
87-61-6	1,2,3-Trichlorobenzene	3.50	U	1	3.50	5.60	ug/Kg	12/01/25 12:26	VY120125

SURROGATES

17060-07-0	1,2-Dichloroethane-d4	63.0			63 - 155	126%	SPK: 50
1868-53-7	Dibromofluoromethane	54.2			70 - 134	108%	SPK: 50
2037-26-5	Toluene-d8	50.5			74 - 123	101%	SPK: 50
460-00-4	4-Bromofluorobenzene	45.3			52 - 138	91%	SPK: 50

INTERNAL STANDARDS

INTERNAL STANDARDS		Area Count
363-72-4	Pentafluorobenzene	676000
540-36-3	1,4-Difluorobenzene	1280000
3114-55-4	Chlorobenzene-d5	1110000
3855-82-1	1,4-Dichlorobenzene-d4	445000

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



QC SUMMARY

Surrogate Summary

SDG No.: **Q3740**

Client: **G Environmental**

Analytical Method: **SW8260D**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3740-01	WC1	1,2-Dichloroethane-d4	50	63.0	126		63	155
		Dibromofluoromethane	50	54.2	108		70	134
		Toluene-d8	50	50.5	101		74	123
		4-Bromofluorobenzene	50	45.3	91		52	138
VY1201SBL01	VY1201SBL01	1,2-Dichloroethane-d4	50	58.3	117		63	155
		Dibromofluoromethane	50	51.0	102		70	134
		Toluene-d8	50	49.5	99		74	123
		4-Bromofluorobenzene	50	42.2	84		52	138
VY1201SBS01	VY1201SBS01	1,2-Dichloroethane-d4	50	47.7	95		63	155
		Dibromofluoromethane	50	50.5	101		70	134
		Toluene-d8	50	50.2	100		74	123
		4-Bromofluorobenzene	50	49.2	98		52	138
VY1201SBSD01	VY1201SBSD01	1,2-Dichloroethane-d4	50	48.0	96		63	155
		Dibromofluoromethane	50	50.5	101		70	134
		Toluene-d8	50	50.0	100		74	123
		4-Bromofluorobenzene	50	48.3	97		52	138

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SDG No.:	Q3740	SW-846	Analytical Method:	SW8260D
Client:	G Environmental	Datafile :	VY023831.D	

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY1201SBS01	Dichlorodifluoromethane	20	22.3	ug/Kg	112			64	136	
	Chloromethane	20	18.5	ug/Kg	93			73	129	
	Vinyl chloride	20	19.8	ug/Kg	99			73	133	
	Bromomethane	20	19.4	ug/Kg	97			77	137	
	Chloroethane	20	20.4	ug/Kg	102			69	130	
	Trichlorofluoromethane	20	21.9	ug/Kg	110			69	134	
	1,1,2-Trichlorotrifluoroethane	20	23.0	ug/Kg	115			81	123	
	1,1-Dichloroethene	20	21.2	ug/Kg	106			79	121	
	Acetone	100	120	ug/Kg	120			60	174	
	Carbon disulfide	20	18.1	ug/Kg	91			73	125	
	Methyl tert-butyl Ether	20	20.6	ug/Kg	103			77	129	
	Methyl Acetate	20	18.4	ug/Kg	92			51	140	
	Methylene Chloride	20	18.8	ug/Kg	94			54	158	
	trans-1,2-Dichloroethene	20	21.1	ug/Kg	106			80	123	
	1,1-Dichloroethane	20	21.7	ug/Kg	109			82	123	
	Cyclohexane	20	21.0	ug/Kg	105			76	122	
	2-Butanone	100	110	ug/Kg	110			69	131	
	Carbon Tetrachloride	20	23.1	ug/Kg	116			76	129	
	cis-1,2-Dichloroethene	20	21.0	ug/Kg	105			82	123	
	Bromochloromethane	20	19.2	ug/Kg	96			80	127	
	Chloroform	20	22.0	ug/Kg	110			82	125	
	1,1,1-Trichloroethane	20	22.3	ug/Kg	112			80	126	
	Methylcyclohexane	20	22.0	ug/Kg	110			77	123	
	Benzene	20	22.0	ug/Kg	110			84	121	
	1,2-Dichloroethane	20	22.3	ug/Kg	112			81	126	
	Trichloroethene	20	22.6	ug/Kg	113			83	122	
	1,2-Dichloropropane	20	22.6	ug/Kg	113			83	122	
	Bromodichloromethane	20	22.2	ug/Kg	111			82	123	
	4-Methyl-2-Pentanone	100	110	ug/Kg	110			70	135	
	Toluene	20	22.5	ug/Kg	113			83	122	
	t-1,3-Dichloropropene	20	21.4	ug/Kg	107			78	124	
	cis-1,3-Dichloropropene	20	22.0	ug/Kg	110			81	122	
	1,1,2-Trichloroethane	20	22.0	ug/Kg	110			82	125	
	2-Hexanone	100	110	ug/Kg	110			66	138	
	Dibromochloromethane	20	22.0	ug/Kg	110			79	125	
	1,2-Dibromoethane	20	21.2	ug/Kg	106			80	125	
	Tetrachloroethene	20	23.3	ug/Kg	117			83	125	
	Chlorobenzene	20	22.6	ug/Kg	113			84	122	
	Ethyl Benzene	20	22.9	ug/Kg	115			82	124	
	m/p-Xylenes	40	46.2	ug/Kg	116			83	124	
	o-Xylene	20	22.5	ug/Kg	113			83	123	
	Styrene	20	22.7	ug/Kg	114			82	124	
	Bromoform	20	21.9	ug/Kg	110			75	127	
	Isopropylbenzene	20	22.6	ug/Kg	113			82	124	
	1,1,2,2-Tetrachloroethane	20	20.9	ug/Kg	104			77	127	
	1,3-Dichlorobenzene	20	21.6	ug/Kg	108			83	122	
	1,4-Dichlorobenzene	20	21.6	ug/Kg	108			84	121	
	1,2-Dichlorobenzene	20	21.9	ug/Kg	110			83	124	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3740</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>G Environmental</u>	Datafile :	<u>VY023831.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	
									High	RPD
VY1201SBS01	1,2-Dibromo-3-Chloropropane	20	19.6	ug/Kg	98			66	134	
	1,2,4-Trichlorobenzene	20	20.8	ug/Kg	104			78	127	
	1,2,3-Trichlorobenzene	20	21.2	ug/Kg	106			70	137	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SDG No.:	Q3740	SW-846	Analytical Method:	SW8260D
Client:	G Environmental	Datafile :	VY023832.D	

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY1201SBSD01	Dichlorodifluoromethane	20	22.3	ug/Kg	112	0		64	136	20
	Chloromethane	20	18.7	ug/Kg	94	1		73	129	15
	Vinyl chloride	20	19.8	ug/Kg	99	0		73	133	15
	Bromomethane	20	19.4	ug/Kg	97	0		77	137	15
	Chloroethane	20	20.2	ug/Kg	101	1		69	130	15
	Trichlorofluoromethane	20	21.9	ug/Kg	110	0		69	134	27
	1,1,2-Trichlorotrifluoroethane	20	23.1	ug/Kg	116	1		81	123	20
	1,1-Dichloroethene	20	21.4	ug/Kg	107	1		79	121	16
	Acetone	100	130	ug/Kg	130	8		60	174	28
	Carbon disulfide	20	18.0	ug/Kg	90	1		73	125	15
	Methyl tert-butyl Ether	20	21.8	ug/Kg	109	6		77	129	20
	Methyl Acetate	20	20.3	ug/Kg	102	10		51	140	30
	Methylene Chloride	20	19.2	ug/Kg	96	2		54	158	20
	trans-1,2-Dichloroethene	20	21.3	ug/Kg	106	0		80	123	15
	1,1-Dichloroethane	20	22.0	ug/Kg	110	1		82	123	15
	Cyclohexane	20	21.3	ug/Kg	106	1		76	122	15
	2-Butanone	100	120	ug/Kg	120	9		69	131	29
	Carbon Tetrachloride	20	22.3	ug/Kg	112	4		76	129	15
	cis-1,2-Dichloroethene	20	21.5	ug/Kg	108	3		82	123	15
	Bromochloromethane	20	19.5	ug/Kg	98	2		80	127	15
	Chloroform	20	22.1	ug/Kg	111	1		82	125	15
	1,1,1-Trichloroethane	20	22.5	ug/Kg	113	1		80	126	15
	Methylcyclohexane	20	21.8	ug/Kg	109	1		77	123	15
	Benzene	20	21.8	ug/Kg	109	1		84	121	15
	1,2-Dichloroethane	20	22.5	ug/Kg	113	1		81	126	15
	Trichloroethene	20	22.3	ug/Kg	112	1		83	122	15
	1,2-Dichloropropane	20	22.7	ug/Kg	114	1		83	122	15
	Bromodichloromethane	20	22.4	ug/Kg	112	1		82	123	15
	4-Methyl-2-Pentanone	100	110	ug/Kg	110	0		70	135	26
	Toluene	20	22.4	ug/Kg	112	1		83	122	15
	t-1,3-Dichloropropene	20	21.8	ug/Kg	109	2		78	124	15
	cis-1,3-Dichloropropene	20	21.8	ug/Kg	109	1		81	122	15
	1,1,2-Trichloroethane	20	23.0	ug/Kg	115	4		82	125	15
	2-Hexanone	100	120	ug/Kg	120	9		66	138	27
	Dibromochloromethane	20	22.4	ug/Kg	112	2		79	125	15
	1,2-Dibromoethane	20	22.1	ug/Kg	111	5		80	125	16
	Tetrachloroethene	20	23.1	ug/Kg	116	1		83	125	15
	Chlorobenzene	20	22.2	ug/Kg	111	2		84	122	15
	Ethyl Benzene	20	22.6	ug/Kg	113	2		82	124	15
	m/p-Xylenes	40	45.5	ug/Kg	114	2		83	124	15
	o-Xylene	20	22.3	ug/Kg	112	1		83	123	15
	Styrene	20	22.7	ug/Kg	114	0		82	124	15
	Bromoform	20	22.2	ug/Kg	111	1		75	127	16
	Isopropylbenzene	20	22.7	ug/Kg	114	1		82	124	15
	1,1,2,2-Tetrachloroethane	20	22.6	ug/Kg	113	8		77	127	21
	1,3-Dichlorobenzene	20	22.3	ug/Kg	112	4		83	122	15
	1,4-Dichlorobenzene	20	21.9	ug/Kg	110	2		84	121	15
	1,2-Dichlorobenzene	20	22.5	ug/Kg	113	3		83	124	15

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3740</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>G Environmental</u>	Datafile :	<u>VY023832.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY1201SBSD01	1,2-Dibromo-3-Chloropropane	20	21.8	ug/Kg	109	11		66	134	20
	1,2,4-Trichlorobenzene	20	21.7	ug/Kg	109	5		78	127	20
	1,2,3-Trichlorobenzene	20	22.2	ug/Kg	111	5		70	137	16

VOLATILE METHOD BLANK SUMMARY

Client ID

VY1201SBL01

Lab Name: Alliance

Contract: GENV01

Lab Code: ACE

SDG NO.: Q3740

Lab File ID: VY023830.D

Lab Sample ID: VY1201SBL01

Date Analyzed: 12/01/2025

Time Analyzed: 09:13

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: (Y/N) Y

Instrument ID: MSVOA_Y

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VY1201SBS01	VY1201SBS01	VY023831.D	12/01/2025
VY1201SBSD01	VY1201SBSD01	VY023832.D	12/01/2025
WC1	Q3740-01	VY023836.D	12/01/2025

COMMENTS: _____

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: GENV01

Lab Code: ACE

SDG NO.: Q3740

Lab File ID: VY023808.D

BFB Injection Date: 11/26/2025

Instrument ID: MSVOA_Y

BFB Injection Time: 07:11

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	18.5
75	30.0 - 60.0% of mass 95	51.4
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.3
173	Less than 2.0% of mass 174	0.9 (1.1) 1
174	50.0 - 100.0% of mass 95	87.3
175	5.0 - 9.0% of mass 174	6.4 (7.4) 1
176	95.0 - 101.0% of mass 174	83.8 (96) 1
177	5.0 - 9.0% of mass 176	5.6 (6.6) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC005	VSTDIC005	VY023809.D	11/26/2025	08:04
VSTDIC010	VSTDIC010	VY023810.D	11/26/2025	08:27
VSTDIC020	VSTDIC020	VY023811.D	11/26/2025	08:49
VSTDIC050	VSTDIC050	VY023812.D	11/26/2025	09:17
VSTDIC100	VSTDIC100	VY023813.D	11/26/2025	09:39
VSTDIC150	VSTDIC150	VY023814.D	11/26/2025	10:02

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: GENV01

Lab Code: ACE

SDG NO.: Q3740

Lab File ID: VY023828.D

BFB Injection Date: 12/01/2025

Instrument ID: MSVOA_Y

BFB Injection Time: 07:24

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	18.9
75	30.0 - 60.0% of mass 95	50.5
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.6
173	Less than 2.0% of mass 174	0.8 (1) 1
174	50.0 - 100.0% of mass 95	82.5
175	5.0 - 9.0% of mass 174	5.8 (7) 1
176	95.0 - 101.0% of mass 174	79.4 (96.3) 1
177	5.0 - 9.0% of mass 176	5.2 (6.5) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VY023829.D	12/01/2025	07:55
VY1201SBL01	VY1201SBL01	VY023830.D	12/01/2025	09:13
VY1201SBS01	VY1201SBS01	VY023831.D	12/01/2025	09:37
VY1201SBSD01	VY1201SBSD01	VY023832.D	12/01/2025	10:00
WC1	Q3740-01	VY023836.D	12/01/2025	12:26

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01
 Lab Code: ACE SDG NO.: Q3740
 Lab File ID: VY023829.D Date Analyzed: 12/01/2025
 Instrument ID: MSVOA_Y Time Analyzed: 07:55
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	565841	7.71	855261	8.62	736930	11.42
UPPER LIMIT	1131680	8.213	1710520	9.122	1473860	11.92
LOWER LIMIT	282921	7.213	427631	8.122	368465	10.92
EPA SAMPLE NO.						
WC1	675919	7.71	1275968	8.62	1114794	11.42
VY1201SBL01	719047	7.71	1221329	8.62	995220	11.42
VY1201SBS01	535284	7.71	821167	8.62	693280	11.42
VY1201SBSD01	526036	7.71	818117	8.62	695554	11.42

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>GENV01</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3740</u>
Lab File ID:	<u>VY023829.D</u>	Date Analyzed:	<u>12/01/2025</u>
Instrument ID:	<u>MSVOA_Y</u>	Time Analyzed:	<u>07:55</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)
		Heated Purge:	(Y/N) <u>Y</u>

	IS4 AREA #	RT #				
12 HOUR STD	375940	13.352				
UPPER LIMIT	751880	13.852				
LOWER LIMIT	187970	12.852				
EPA SAMPLE NO.						
WC1	445014	13.35				
VY1201SBL01	399677	13.35				
VY1201SBS01	360194	13.35				
VY1201SBSD01	353062	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client: G Environmental
Project: Diamond
Client Sample ID: VY1201SBL01
Lab Sample ID: VY1201SBL01
Analytical Method: 8260D
Sample Wt/Vol: 5 g

Level: LOW
Final Vol: 5000 uL

Date Collected:
Date Received:
SDG No.: Q3740
Matrix: SOIL
% Solid: 100
Test: VOC-TCLVOA-10

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
75-71-8	Dichlorodifluoromethane	1.10	U	1	1.10	5.00	ug/Kg	12/01/25 09:13	VY120125
74-87-3	Chloromethane	1.10	U	1	1.10	5.00	ug/Kg	12/01/25 09:13	VY120125
75-01-4	Vinyl Chloride	0.79	U	1	0.79	5.00	ug/Kg	12/01/25 09:13	VY120125
74-83-9	Bromomethane	1.10	U	1	1.10	5.00	ug/Kg	12/01/25 09:13	VY120125
75-00-3	Chloroethane	1.30	U	1	1.30	5.00	ug/Kg	12/01/25 09:13	VY120125
75-69-4	Trichlorofluoromethane	1.20	U	1	1.20	5.00	ug/Kg	12/01/25 09:13	VY120125
76-13-1	1,1,2-Trichlorotrifluoroethane	1.10	U	1	1.10	5.00	ug/Kg	12/01/25 09:13	VY120125
75-35-4	1,1-Dichloroethene	1.00	U	1	1.00	5.00	ug/Kg	12/01/25 09:13	VY120125
67-64-1	Acetone	4.70	U	1	4.70	25.0	ug/Kg	12/01/25 09:13	VY120125
75-15-0	Carbon Disulfide	1.10	U	1	1.10	5.00	ug/Kg	12/01/25 09:13	VY120125
1634-04-4	Methyl tert-butyl Ether	0.73	U	1	0.73	5.00	ug/Kg	12/01/25 09:13	VY120125
79-20-9	Methyl Acetate	1.50	U	1	1.50	5.00	ug/Kg	12/01/25 09:13	VY120125
75-09-2	Methylene Chloride	3.50	U	1	3.50	10.0	ug/Kg	12/01/25 09:13	VY120125
156-60-5	trans-1,2-Dichloroethene	0.86	U	1	0.86	5.00	ug/Kg	12/01/25 09:13	VY120125
75-34-3	1,1-Dichloroethane	0.80	U	1	0.80	5.00	ug/Kg	12/01/25 09:13	VY120125
110-82-7	Cyclohexane	0.79	U	1	0.79	5.00	ug/Kg	12/01/25 09:13	VY120125
78-93-3	2-Butanone	6.50	U	1	6.50	25.0	ug/Kg	12/01/25 09:13	VY120125
56-23-5	Carbon Tetrachloride	0.97	U	1	0.97	5.00	ug/Kg	12/01/25 09:13	VY120125
156-59-2	cis-1,2-Dichloroethene	0.75	U	1	0.75	5.00	ug/Kg	12/01/25 09:13	VY120125
74-97-5	Bromochloromethane	1.20	U	1	1.20	5.00	ug/Kg	12/01/25 09:13	VY120125
67-66-3	Chloroform	0.84	U	1	0.84	5.00	ug/Kg	12/01/25 09:13	VY120125
71-55-6	1,1,1-Trichloroethane	0.93	U	1	0.93	5.00	ug/Kg	12/01/25 09:13	VY120125
108-87-2	Methylcyclohexane	0.91	U	1	0.91	5.00	ug/Kg	12/01/25 09:13	VY120125
71-43-2	Benzene	0.79	U	1	0.79	5.00	ug/Kg	12/01/25 09:13	VY120125
107-06-2	1,2-Dichloroethane	0.79	U	1	0.79	5.00	ug/Kg	12/01/25 09:13	VY120125
79-01-6	Trichloroethene	0.81	U	1	0.81	5.00	ug/Kg	12/01/25 09:13	VY120125
78-87-5	1,2-Dichloropropane	0.91	U	1	0.91	5.00	ug/Kg	12/01/25 09:13	VY120125
75-27-4	Bromodichloromethane	0.78	U	1	0.78	5.00	ug/Kg	12/01/25 09:13	VY120125
108-10-1	4-Methyl-2-Pentanone	3.60	U	1	3.60	25.0	ug/Kg	12/01/25 09:13	VY120125
108-88-3	Toluene	0.78	U	1	0.78	5.00	ug/Kg	12/01/25 09:13	VY120125
10061-02-6	t-1,3-Dichloropropene	0.65	U	1	0.65	5.00	ug/Kg	12/01/25 09:13	VY120125
10061-01-5	cis-1,3-Dichloropropene	0.62	U	1	0.62	5.00	ug/Kg	12/01/25 09:13	VY120125
79-00-5	1,1,2-Trichloroethane	0.92	U	1	0.92	5.00	ug/Kg	12/01/25 09:13	VY120125
591-78-6	2-Hexanone	3.70	U	1	3.70	25.0	ug/Kg	12/01/25 09:13	VY120125
124-48-1	Dibromochloromethane	0.87	U	1	0.87	5.00	ug/Kg	12/01/25 09:13	VY120125
106-93-4	1,2-Dibromoethane	0.88	U	1	0.88	5.00	ug/Kg	12/01/25 09:13	VY120125
127-18-4	Tetrachloroethene	1.10	U	1	1.10	5.00	ug/Kg	12/01/25 09:13	VY120125
108-90-7	Chlorobenzene	0.91	U	1	0.91	5.00	ug/Kg	12/01/25 09:13	VY120125
100-41-4	Ethyl Benzene	0.67	U	1	0.67	5.00	ug/Kg	12/01/25 09:13	VY120125
179601-23-1	m/p-Xylenes	1.20	U	1	1.20	10.0	ug/Kg	12/01/25 09:13	VY120125
95-47-6	o-Xylene	0.82	U	1	0.82	5.00	ug/Kg	12/01/25 09:13	VY120125
100-42-5	Styrene	0.71	U	1	0.71	5.00	ug/Kg	12/01/25 09:13	VY120125
75-25-2	Bromoform	0.86	U	1	0.86	5.00	ug/Kg	12/01/25 09:13	VY120125

Report of Analysis

Client: G Environmental
 Project: Diamond
 Client Sample ID: VY1201SBL01
 Lab Sample ID: VY1201SBL01
 Analytical Method: 8260D
 Sample Wt/Vol: 5 g

Level: LOW
 Final Vol: 5000 uL

Date Collected:
 Date Received:
 SDG No.: Q3740
 Matrix: SOIL
 % Solid: 100
 Test: VOC-TCLVOA-10

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
98-82-8	Isopropylbenzene	0.78	U	1	0.78	5.00	ug/Kg	12/01/25 09:13	VY120125
79-34-5	1,1,2,2-Tetrachloroethane	1.20	U	1	1.20	5.00	ug/Kg	12/01/25 09:13	VY120125
541-73-1	1,3-Dichlorobenzene	1.70	U	1	1.70	5.00	ug/Kg	12/01/25 09:13	VY120125
106-46-7	1,4-Dichlorobenzene	1.60	U	1	1.60	5.00	ug/Kg	12/01/25 09:13	VY120125
95-50-1	1,2-Dichlorobenzene	1.50	U	1	1.50	5.00	ug/Kg	12/01/25 09:13	VY120125
96-12-8	1,2-Dibromo-3-Chloropropane	1.80	U	1	1.80	5.00	ug/Kg	12/01/25 09:13	VY120125
120-82-1	1,2,4-Trichlorobenzene	3.00	U	1	3.00	5.00	ug/Kg	12/01/25 09:13	VY120125
87-61-6	1,2,3-Trichlorobenzene	3.20	U	1	3.20	5.00	ug/Kg	12/01/25 09:13	VY120125

SURROGATES

17060-07-0	1,2-Dichloroethane-d4	58.3			63 - 155	117%	SPK: 50
1868-53-7	Dibromofluoromethane	51.0			70 - 134	102%	SPK: 50
2037-26-5	Toluene-d8	49.5			74 - 123	99%	SPK: 50
460-00-4	4-Bromofluorobenzene	42.2			52 - 138	84%	SPK: 50

INTERNAL STANDARDS

INTERNAL STANDARDS		Area Count
363-72-4	Pentafluorobenzene	719000
540-36-3	1,4-Difluorobenzene	1220000
3114-55-4	Chlorobenzene-d5	995000
3855-82-1	1,4-Dichlorobenzene-d4	400000

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client: G Environmental
Project: Diamond
Client Sample ID: VY1201SBS01
Lab Sample ID: VY1201SBS01
Analytical Method: 8260D
Sample Wt/Vol: 5 g

Level: LOW
Final Vol: 5000 uL

Date Collected:
Date Received:
SDG No.: Q3740
Matrix: SOIL
% Solid: 100
Test: VOC-TCLVOA-10

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
75-71-8	Dichlorodifluoromethane	22.3		1	1.10	5.00	ug/Kg	12/01/25 09:37	VY120125
74-87-3	Chloromethane	18.5		1	1.10	5.00	ug/Kg	12/01/25 09:37	VY120125
75-01-4	Vinyl Chloride	19.8		1	0.79	5.00	ug/Kg	12/01/25 09:37	VY120125
74-83-9	Bromomethane	19.4		1	1.10	5.00	ug/Kg	12/01/25 09:37	VY120125
75-00-3	Chloroethane	20.4		1	1.30	5.00	ug/Kg	12/01/25 09:37	VY120125
75-69-4	Trichlorofluoromethane	21.9		1	1.20	5.00	ug/Kg	12/01/25 09:37	VY120125
76-13-1	1,1,2-Trichlorotrifluoroethane	23.0		1	1.10	5.00	ug/Kg	12/01/25 09:37	VY120125
75-35-4	1,1-Dichloroethene	21.2		1	1.00	5.00	ug/Kg	12/01/25 09:37	VY120125
67-64-1	Acetone	120		1	4.70	25.0	ug/Kg	12/01/25 09:37	VY120125
75-15-0	Carbon Disulfide	18.1		1	1.10	5.00	ug/Kg	12/01/25 09:37	VY120125
1634-04-4	Methyl tert-butyl Ether	20.6		1	0.73	5.00	ug/Kg	12/01/25 09:37	VY120125
79-20-9	Methyl Acetate	18.4		1	1.50	5.00	ug/Kg	12/01/25 09:37	VY120125
75-09-2	Methylene Chloride	18.8		1	3.50	10.0	ug/Kg	12/01/25 09:37	VY120125
156-60-5	trans-1,2-Dichloroethene	21.1		1	0.86	5.00	ug/Kg	12/01/25 09:37	VY120125
75-34-3	1,1-Dichloroethane	21.7		1	0.80	5.00	ug/Kg	12/01/25 09:37	VY120125
110-82-7	Cyclohexane	21.0		1	0.79	5.00	ug/Kg	12/01/25 09:37	VY120125
78-93-3	2-Butanone	110		1	6.50	25.0	ug/Kg	12/01/25 09:37	VY120125
56-23-5	Carbon Tetrachloride	23.1		1	0.97	5.00	ug/Kg	12/01/25 09:37	VY120125
156-59-2	cis-1,2-Dichloroethene	21.0		1	0.75	5.00	ug/Kg	12/01/25 09:37	VY120125
74-97-5	Bromochloromethane	19.2		1	1.20	5.00	ug/Kg	12/01/25 09:37	VY120125
67-66-3	Chloroform	22.0		1	0.84	5.00	ug/Kg	12/01/25 09:37	VY120125
71-55-6	1,1,1-Trichloroethane	22.3		1	0.93	5.00	ug/Kg	12/01/25 09:37	VY120125
108-87-2	Methylcyclohexane	22.0		1	0.91	5.00	ug/Kg	12/01/25 09:37	VY120125
71-43-2	Benzene	22.0		1	0.79	5.00	ug/Kg	12/01/25 09:37	VY120125
107-06-2	1,2-Dichloroethane	22.3		1	0.79	5.00	ug/Kg	12/01/25 09:37	VY120125
79-01-6	Trichloroethene	22.6		1	0.81	5.00	ug/Kg	12/01/25 09:37	VY120125
78-87-5	1,2-Dichloropropane	22.6		1	0.91	5.00	ug/Kg	12/01/25 09:37	VY120125
75-27-4	Bromodichloromethane	22.2		1	0.78	5.00	ug/Kg	12/01/25 09:37	VY120125
108-10-1	4-Methyl-2-Pentanone	110		1	3.60	25.0	ug/Kg	12/01/25 09:37	VY120125
108-88-3	Toluene	22.5		1	0.78	5.00	ug/Kg	12/01/25 09:37	VY120125
10061-02-6	t-1,3-Dichloropropene	21.4		1	0.65	5.00	ug/Kg	12/01/25 09:37	VY120125
10061-01-5	cis-1,3-Dichloropropene	22.0		1	0.62	5.00	ug/Kg	12/01/25 09:37	VY120125
79-00-5	1,1,2-Trichloroethane	22.0		1	0.92	5.00	ug/Kg	12/01/25 09:37	VY120125
591-78-6	2-Hexanone	110		1	3.70	25.0	ug/Kg	12/01/25 09:37	VY120125
124-48-1	Dibromochloromethane	22.0		1	0.87	5.00	ug/Kg	12/01/25 09:37	VY120125
106-93-4	1,2-Dibromoethane	21.2		1	0.88	5.00	ug/Kg	12/01/25 09:37	VY120125
127-18-4	Tetrachloroethene	23.3		1	1.10	5.00	ug/Kg	12/01/25 09:37	VY120125
108-90-7	Chlorobenzene	22.6		1	0.91	5.00	ug/Kg	12/01/25 09:37	VY120125
100-41-4	Ethyl Benzene	22.9		1	0.67	5.00	ug/Kg	12/01/25 09:37	VY120125
179601-23-1	m/p-Xylenes	46.2		1	1.20	10.0	ug/Kg	12/01/25 09:37	VY120125
95-47-6	o-Xylene	22.5		1	0.82	5.00	ug/Kg	12/01/25 09:37	VY120125
100-42-5	Styrene	22.7		1	0.71	5.00	ug/Kg	12/01/25 09:37	VY120125
75-25-2	Bromoform	21.9		1	0.86	5.00	ug/Kg	12/01/25 09:37	VY120125

Report of Analysis

Client: G Environmental
 Project: Diamond
 Client Sample ID: VY1201SBS01
 Lab Sample ID: VY1201SBS01
 Analytical Method: 8260D
 Sample Wt/Vol: 5 g

Level: LOW
 Final Vol: 5000 uL

Date Collected:
 Date Received:
 SDG No.: Q3740
 Matrix: SOIL
 % Solid: 100
 Test: VOC-TCLVOA-10

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
98-82-8	Isopropylbenzene	22.6		1	0.78	5.00	ug/Kg	12/01/25 09:37	VY120125
79-34-5	1,1,2,2-Tetrachloroethane	20.9		1	1.20	5.00	ug/Kg	12/01/25 09:37	VY120125
541-73-1	1,3-Dichlorobenzene	21.6		1	1.70	5.00	ug/Kg	12/01/25 09:37	VY120125
106-46-7	1,4-Dichlorobenzene	21.6		1	1.60	5.00	ug/Kg	12/01/25 09:37	VY120125
95-50-1	1,2-Dichlorobenzene	21.9		1	1.50	5.00	ug/Kg	12/01/25 09:37	VY120125
96-12-8	1,2-Dibromo-3-Chloropropane	19.6		1	1.80	5.00	ug/Kg	12/01/25 09:37	VY120125
120-82-1	1,2,4-Trichlorobenzene	20.8		1	3.00	5.00	ug/Kg	12/01/25 09:37	VY120125
87-61-6	1,2,3-Trichlorobenzene	21.2		1	3.20	5.00	ug/Kg	12/01/25 09:37	VY120125
SURROGATES									
17060-07-0	1,2-Dichloroethane-d4	47.7			63 - 155	95%	SPK: 50		
1868-53-7	Dibromofluoromethane	50.5			70 - 134	101%	SPK: 50		
2037-26-5	Toluene-d8	50.2			74 - 123	100%	SPK: 50		
460-00-4	4-Bromofluorobenzene	49.2			52 - 138	98%	SPK: 50		
INTERNAL STANDARDS									
		Area Count							
363-72-4	Pentafluorobenzene	535000							
540-36-3	1,4-Difluorobenzene	821000							
3114-55-4	Chlorobenzene-d5	693000							
3855-82-1	1,4-Dichlorobenzene-d4	360000							

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client: G Environmental
 Project: Diamond
 Client Sample ID: VY1201SBSD01
 Lab Sample ID: VY1201SBSD01
 Analytical Method: 8260D
 Sample Wt/Vol: 5 g

Level: LOW
 Final Vol: 5000 uL

Date Collected:
 Date Received:
 SDG No.: Q3740
 Matrix: SOIL
 % Solid: 100
 Test: VOC-TCLVOA-10

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
75-71-8	Dichlorodifluoromethane	22.3		1	1.10	5.00	ug/Kg	12/01/25 10:00	VY120125
74-87-3	Chloromethane	18.7		1	1.10	5.00	ug/Kg	12/01/25 10:00	VY120125
75-01-4	Vinyl Chloride	19.8		1	0.79	5.00	ug/Kg	12/01/25 10:00	VY120125
74-83-9	Bromomethane	19.4		1	1.10	5.00	ug/Kg	12/01/25 10:00	VY120125
75-00-3	Chloroethane	20.2		1	1.30	5.00	ug/Kg	12/01/25 10:00	VY120125
75-69-4	Trichlorofluoromethane	21.9		1	1.20	5.00	ug/Kg	12/01/25 10:00	VY120125
76-13-1	1,1,2-Trichlorotrifluoroethane	23.1		1	1.10	5.00	ug/Kg	12/01/25 10:00	VY120125
75-35-4	1,1-Dichloroethene	21.4		1	1.00	5.00	ug/Kg	12/01/25 10:00	VY120125
67-64-1	Acetone	130		1	4.70	25.0	ug/Kg	12/01/25 10:00	VY120125
75-15-0	Carbon Disulfide	18.0		1	1.10	5.00	ug/Kg	12/01/25 10:00	VY120125
1634-04-4	Methyl tert-butyl Ether	21.8		1	0.73	5.00	ug/Kg	12/01/25 10:00	VY120125
79-20-9	Methyl Acetate	20.3		1	1.50	5.00	ug/Kg	12/01/25 10:00	VY120125
75-09-2	Methylene Chloride	19.2		1	3.50	10.0	ug/Kg	12/01/25 10:00	VY120125
156-60-5	trans-1,2-Dichloroethene	21.3		1	0.86	5.00	ug/Kg	12/01/25 10:00	VY120125
75-34-3	1,1-Dichloroethane	22.0		1	0.80	5.00	ug/Kg	12/01/25 10:00	VY120125
110-82-7	Cyclohexane	21.3		1	0.79	5.00	ug/Kg	12/01/25 10:00	VY120125
78-93-3	2-Butanone	120		1	6.50	25.0	ug/Kg	12/01/25 10:00	VY120125
56-23-5	Carbon Tetrachloride	22.3		1	0.97	5.00	ug/Kg	12/01/25 10:00	VY120125
156-59-2	cis-1,2-Dichloroethene	21.5		1	0.75	5.00	ug/Kg	12/01/25 10:00	VY120125
74-97-5	Bromochloromethane	19.5		1	1.20	5.00	ug/Kg	12/01/25 10:00	VY120125
67-66-3	Chloroform	22.1		1	0.84	5.00	ug/Kg	12/01/25 10:00	VY120125
71-55-6	1,1,1-Trichloroethane	22.5		1	0.93	5.00	ug/Kg	12/01/25 10:00	VY120125
108-87-2	Methylcyclohexane	21.8		1	0.91	5.00	ug/Kg	12/01/25 10:00	VY120125
71-43-2	Benzene	21.8		1	0.79	5.00	ug/Kg	12/01/25 10:00	VY120125
107-06-2	1,2-Dichloroethane	22.5		1	0.79	5.00	ug/Kg	12/01/25 10:00	VY120125
79-01-6	Trichloroethene	22.3		1	0.81	5.00	ug/Kg	12/01/25 10:00	VY120125
78-87-5	1,2-Dichloropropane	22.7		1	0.91	5.00	ug/Kg	12/01/25 10:00	VY120125
75-27-4	Bromodichloromethane	22.4		1	0.78	5.00	ug/Kg	12/01/25 10:00	VY120125
108-10-1	4-Methyl-2-Pentanone	110		1	3.60	25.0	ug/Kg	12/01/25 10:00	VY120125
108-88-3	Toluene	22.4		1	0.78	5.00	ug/Kg	12/01/25 10:00	VY120125
10061-02-6	t-1,3-Dichloropropene	21.8		1	0.65	5.00	ug/Kg	12/01/25 10:00	VY120125
10061-01-5	cis-1,3-Dichloropropene	21.8		1	0.62	5.00	ug/Kg	12/01/25 10:00	VY120125
79-00-5	1,1,2-Trichloroethane	23.0		1	0.92	5.00	ug/Kg	12/01/25 10:00	VY120125
591-78-6	2-Hexanone	120		1	3.70	25.0	ug/Kg	12/01/25 10:00	VY120125
124-48-1	Dibromochloromethane	22.4		1	0.87	5.00	ug/Kg	12/01/25 10:00	VY120125
106-93-4	1,2-Dibromoethane	22.1		1	0.88	5.00	ug/Kg	12/01/25 10:00	VY120125
127-18-4	Tetrachloroethene	23.1		1	1.10	5.00	ug/Kg	12/01/25 10:00	VY120125
108-90-7	Chlorobenzene	22.2		1	0.91	5.00	ug/Kg	12/01/25 10:00	VY120125
100-41-4	Ethyl Benzene	22.6		1	0.67	5.00	ug/Kg	12/01/25 10:00	VY120125
179601-23-1	m/p-Xylenes	45.5		1	1.20	10.0	ug/Kg	12/01/25 10:00	VY120125
95-47-6	o-Xylene	22.3		1	0.82	5.00	ug/Kg	12/01/25 10:00	VY120125
100-42-5	Styrene	22.7		1	0.71	5.00	ug/Kg	12/01/25 10:00	VY120125
75-25-2	Bromoform	22.2		1	0.86	5.00	ug/Kg	12/01/25 10:00	VY120125

Report of Analysis

Client: G Environmental
 Project: Diamond
 Client Sample ID: VY1201SBSD01
 Lab Sample ID: VY1201SBSD01
 Analytical Method: 8260D
 Sample Wt/Vol: 5 g

Level: LOW
 Final Vol: 5000 uL

Date Collected:
 Date Received:
 SDG No.: Q3740
 Matrix: SOIL
 % Solid: 100
 Test: VOC-TCLVOA-10

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
98-82-8	Isopropylbenzene	22.7		1	0.78	5.00	ug/Kg	12/01/25 10:00	VY120125
79-34-5	1,1,2,2-Tetrachloroethane	22.6		1	1.20	5.00	ug/Kg	12/01/25 10:00	VY120125
541-73-1	1,3-Dichlorobenzene	22.3		1	1.70	5.00	ug/Kg	12/01/25 10:00	VY120125
106-46-7	1,4-Dichlorobenzene	21.9		1	1.60	5.00	ug/Kg	12/01/25 10:00	VY120125
95-50-1	1,2-Dichlorobenzene	22.5		1	1.50	5.00	ug/Kg	12/01/25 10:00	VY120125
96-12-8	1,2-Dibromo-3-Chloropropane	21.8		1	1.80	5.00	ug/Kg	12/01/25 10:00	VY120125
120-82-1	1,2,4-Trichlorobenzene	21.7		1	3.00	5.00	ug/Kg	12/01/25 10:00	VY120125
87-61-6	1,2,3-Trichlorobenzene	22.2		1	3.20	5.00	ug/Kg	12/01/25 10:00	VY120125

SURROGATES

17060-07-0	1,2-Dichloroethane-d4	48.0			63 - 155	96%	SPK: 50
1868-53-7	Dibromofluoromethane	50.5			70 - 134	101%	SPK: 50
2037-26-5	Toluene-d8	50.0			74 - 123	100%	SPK: 50
460-00-4	4-Bromofluorobenzene	48.3			52 - 138	97%	SPK: 50

INTERNAL STANDARDS

INTERNAL STANDARDS		Area Count
363-72-4	Pentafluorobenzene	526000
540-36-3	1,4-Difluorobenzene	818000
3114-55-4	Chlorobenzene-d5	696000
3855-82-1	1,4-Dichlorobenzene-d4	353000

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: <u>Alliance</u>	Contract: <u>GENV01</u>
Lab Code: <u>ACE</u>	SDG No.: <u>Q3740</u>
Instrument ID: <u>MSVOA_Y</u>	Calibration Date(s): <u>11/26/2025</u> <u>11/26/2025</u>
Heated Purge: (Y/N) <u>Y</u>	Calibration Time(s): <u>08:04</u> <u>10:02</u>
GC Column: <u>RXI-624</u> ID: <u>0.25</u> (mm)	

LAB FILE ID: RRF005 = VY023809.D RRF010 = VY023810.D RRF020 = VY023811.D RRF050 = VY023812.D RRF100 = VY023813.D RRF150 = VY023814.D								
COMPOUND	RRF005	RRF010	RRF020	RRF050	RRF100	RRF150	RRF	% RSD
Dichlorodifluoromethane	0.422	0.412	0.414	0.340	0.339	0.334	0.377	11.4
Chloromethane	0.640	0.632	0.607	0.531	0.532	0.515	0.576	9.8
Vinyl Chloride	0.662	0.655	0.640	0.604	0.602	0.580	0.624	5.3
Bromomethane	0.543	0.480	0.439	0.386	0.402	0.390	0.440	14
Chloroethane	0.437	0.420	0.420	0.395	0.394	0.372	0.406	5.8
Trichlorofluoromethane	0.906	0.865	0.874	0.815	0.805	0.793	0.843	5.3
1,1,2-Trichlorotrifluoroethane	0.537	0.509	0.501	0.473	0.477	0.455	0.492	6
1,1-Dichloroethene	0.508	0.487	0.491	0.471	0.476	0.471	0.484	3
Acetone	0.148	0.138	0.137	0.135	0.150	0.133	0.140	5.1
Carbon Disulfide	1.644	1.610	1.618	1.519	1.521	1.461	1.562	4.6
Methyl tert-butyl Ether	1.182	1.158	1.200	1.242	1.355	1.287	1.237	6
Methyl Acetate	0.273	0.264	0.262	0.257	0.299	0.270	0.271	5.5
Methylene Chloride	0.755	0.662	0.582	0.530	0.530	0.500	0.593	16.5
trans-1,2-Dichloroethene	0.553	0.537	0.543	0.526	0.539	0.518	0.536	2.3
1,1-Dichloroethane	1.012	0.982	0.989	0.955	0.968	0.947	0.976	2.4
Cyclohexane	1.060	0.953	0.937	0.896	0.903	0.867	0.936	7.2
2-Butanone	0.169	0.163	0.163	0.169	0.192	0.171	0.171	6.4
Carbon Tetrachloride	0.514	0.490	0.502	0.494	0.507	0.483	0.498	2.3
cis-1,2-Dichloroethene	0.636	0.620	0.623	0.611	0.632	0.613	0.622	1.6
Bromochloromethane	0.437	0.400	0.405	0.399	0.408	0.386	0.406	4.2
Chloroform	1.025	0.991	0.996	0.961	0.969	0.943	0.981	3
1,1,1-Trichloroethane	0.893	0.885	0.882	0.851	0.869	0.844	0.871	2.3
Methylcyclohexane	0.598	0.572	0.591	0.604	0.631	0.603	0.600	3.2
Benzene	1.415	1.388	1.434	1.383	1.410	1.355	1.398	2
1,2-Dichloroethane	0.365	0.354	0.358	0.359	0.372	0.354	0.360	2
Trichloroethene	0.366	0.347	0.371	0.355	0.362	0.350	0.359	2.6
1,2-Dichloropropane	0.324	0.325	0.336	0.329	0.339	0.324	0.330	2
Bromodichloromethane	0.468	0.460	0.468	0.463	0.484	0.460	0.467	1.9
4-Methyl-2-Pentanone	0.183	0.185	0.199	0.219	0.252	0.223	0.210	12.6
Toluene	0.818	0.831	0.883	0.892	0.908	0.866	0.866	4.1

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name:	<u>Alliance</u>	Contract:	<u>GENV01</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3740</u>
Instrument ID:	<u>MSVOA_Y</u>	Calibration Date(s):	<u>11/26/2025</u> <u>11/26/2025</u>
Heated Purge: (Y/N)	<u>Y</u>	Calibration Time(s):	<u>08:04</u> <u>10:02</u>
GC Column:	<u>RXI-624</u> ID: <u>0.25</u> (mm)		

LAB FILE ID:								
RRF005 = VY023809.D			RRF010 = VY023810.D			RRF020 = VY023811.D		
RRF050 = VY023812.D			RRF100 = VY023813.D			RRF150 = VY023814.D		
COMPOUND	RRF005	RRF010	RRF020	RRF050	RRF100	RRF150	RRF	% RSD
t-1,3-Dichloropropene	0.399	0.394	0.415	0.438	0.473	0.450	0.428	7.2
cis-1,3-Dichloropropene	0.488	0.489	0.509	0.516	0.548	0.526	0.512	4.5
1,1,2-Trichloroethane	0.233	0.230	0.235	0.242	0.257	0.240	0.240	4
2-Hexanone	0.131	0.135	0.147	0.162	0.190	0.167	0.155	14.2
Dibromochloromethane	0.305	0.308	0.321	0.322	0.347	0.326	0.321	4.7
1,2-Dibromoethane	0.218	0.207	0.219	0.220	0.243	0.226	0.222	5.4
Tetrachloroethene	0.525	0.455	0.510	0.470	0.442	0.445	0.474	7.4
Chlorobenzene	1.100	1.091	1.118	1.093	1.109	1.074	1.098	1.4
Ethyl Benzene	1.779	1.789	1.887	1.918	1.967	1.907	1.875	4
m/p-Xylenes	0.676	0.695	0.739	0.743	0.754	0.728	0.722	4.2
o-Xylene	0.633	0.643	0.679	0.690	0.724	0.695	0.677	5
Styrene	0.995	1.047	1.117	1.162	1.210	1.163	1.116	7.2
Bromoform	0.214	0.209	0.209	0.216	0.237	0.221	0.217	4.9
Isopropylbenzene	3.390	3.431	3.571	3.610	3.701	3.590	3.549	3.3
1,1,2,2-Tetrachloroethane	0.580	0.573	0.524	0.576	0.654	0.599	0.584	7.2
1,3-Dichlorobenzene	1.696	1.674	1.670	1.654	1.698	1.645	1.673	1.3
1,4-Dichlorobenzene	1.798	1.678	1.645	1.635	1.662	1.601	1.670	4.1
1,2-Dichlorobenzene	1.489	1.439	1.456	1.460	1.501	1.431	1.463	1.9
1,2-Dibromo-3-Chloropropane	0.104	0.094	0.085	0.095	0.110	0.100	0.098	8.7
1,2,4-Trichlorobenzene	0.855	0.777	0.835	0.877	0.949	0.940	0.872	7.5
1,2,3-Trichlorobenzene	0.711	0.658	0.710	0.740	0.817	0.820	0.743	8.7
1,2-Dichloroethane-d4	0.508	0.478	0.483	0.492	0.486	0.475	0.487	2.5
Dibromofluoromethane	0.299	0.285	0.297	0.303	0.299	0.295	0.296	2.1
Toluene-d8	1.161	1.125	1.183	1.212	1.209	1.170	1.176	2.8
4-Bromofluorobenzene	0.388	0.362	0.381	0.395	0.406	0.390	0.387	3.8

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>GENV01</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3740</u>
Instrument ID:	<u>MSVOA_Y</u>	Calibration Date/Time:	<u>12/01/2025 07:55</u>
Lab File ID:	<u>VY023829.D</u>	Init. Calib. Date(s):	<u>11/26/2025 11/26/2025</u>
Heated Purge: (Y/N)	<u>Y</u>	Init. Calib. Time(s):	<u>08:04 10:02</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.377	0.352		-6.63	20
Chloromethane	0.576	0.473	0.1	-17.88	20
Vinyl Chloride	0.624	0.570		-8.65	20
Bromomethane	0.440	0.360		-18.18	20
Chloroethane	0.406	0.380		-6.4	20
Trichlorofluoromethane	0.843	0.848		0.59	20
1,1,2-Trichlorotrifluoroethane	0.492	0.526		6.91	20
1,1-Dichloroethene	0.484	0.491		1.45	20
Acetone	0.140	0.158		12.86	20
Carbon Disulfide	1.562	1.340		-14.21	20
Methyl tert-butyl Ether	1.237	1.330		7.52	20
Methyl Acetate	0.271	0.304		12.18	20
Methylene Chloride	0.593	0.531		-10.45	20
trans-1,2-Dichloroethene	0.536	0.539		0.56	20
1,1-Dichloroethane	0.976	1.008	0.1	3.28	20
Cyclohexane	0.936	0.916		-2.14	20
2-Butanone	0.171	0.187		9.36	20
Carbon Tetrachloride	0.498	0.548		10.04	20
cis-1,2-Dichloroethene	0.622	0.639		2.73	20
Bromochloromethane	0.406	0.389		-4.19	20
Chloroform	0.981	1.019		3.87	20
1,1,1-Trichloroethane	0.871	0.916		5.17	20
Methylcyclohexane	0.600	0.648		8	20
Benzene	1.398	1.495		6.94	20
1,2-Dichloroethane	0.360	0.396		10	20
Trichloroethene	0.359	0.392		9.19	20
1,2-Dichloropropane	0.330	0.363		10	20
Bromodichloromethane	0.467	0.516		10.49	20
4-Methyl-2-Pentanone	0.210	0.252		20	20
Toluene	0.866	0.957		10.51	20
t-1,3-Dichloropropene	0.428	0.474		10.75	20
cis-1,3-Dichloropropene	0.512	0.561		9.57	20
1,1,2-Trichloroethane	0.240	0.275		14.58	20
2-Hexanone	0.155	0.188		21.29	20
Dibromochloromethane	0.321	0.361		12.46	20
1,2-Dibromoethane	0.222	0.249		12.16	20
Tetrachloroethene	0.474	0.532		12.24	20
Chlorobenzene	1.098	1.206	0.3	9.84	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>GENV01</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3740</u>
Instrument ID:	<u>MSVOA_Y</u>	Calibration Date/Time:	<u>12/01/2025</u> <u>07:55</u>
Lab File ID:	<u>VY023829.D</u>	Init. Calib. Date(s):	<u>11/26/2025</u> <u>11/26/2025</u>
Heated Purge: (Y/N)	<u>Y</u>	Init. Calib. Time(s):	<u>08:04</u> <u>10:02</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.875	2.141		14.19	20
m/p-Xylenes	0.722	0.815		12.88	20
o-Xylene	0.677	0.762		12.56	20
Styrene	1.116	1.275		14.25	20
Bromoform	0.217	0.246	0.1	13.36	20
Isopropylbenzene	3.549	4.062		14.45	20
1,1,2,2-Tetrachloroethane	0.584	0.664	0.3	13.7	20
1,3-Dichlorobenzene	1.673	1.834		9.62	20
1,4-Dichlorobenzene	1.670	1.788		7.07	20
1,2-Dichlorobenzene	1.463	1.606		9.77	20
1,2-Dibromo-3-Chloropropane	0.098	0.107		9.18	20
1,2,4-Trichlorobenzene	0.872	0.955		9.52	20
1,2,3-Trichlorobenzene	0.743	0.826		11.17	20
1,2-Dichloroethane-d4	0.487	0.517		6.16	20
Dibromofluoromethane	0.296	0.332		12.16	20
Toluene-d8	1.176	1.313		11.65	20
4-Bromofluorobenzene	0.387	0.437		12.92	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.



SAMPLE RAW DATA

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120125\
 Data File : VY023836.D
 Acq On : 01 Dec 2025 12:26
 Operator : SY/MD
 Sample : Q3740-01
 Misc : 4.98g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC1

A
B
C
D
E
F
G
H
I
J

Quant Time: Dec 02 03:15:43 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y112625S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 27 01:14:36 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	675919	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	1275968	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	1114794	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	445014	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	414869	63.027	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 126.060%	
35) Dibromofluoromethane	7.640	113	409711	54.204	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 108.400%	
50) Toluene-d8	10.109	98	1516112	50.498	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 101.000%	
62) 4-Bromofluorobenzene	12.408	95	447050	45.254	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 90.500%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120125\
Data File : VY023836.D
Acq On : 01 Dec 2025 12:26
Operator : SY/MD
Sample : Q3740-01
Misc : 4.98g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC1

Quant Time: Dec 02 03:15:43 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y112625S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 27 01:14:36 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.713 min Scan# 983
Delta R.T. 0.006 min
Lab File: VY023836.D
Acq: 01 Dec 2025 12:26




Instrument :
MSVOA_Y
ClientSampleId :
WC1


Tgt Ion:168 Resp: 675919
Ion Ratio Lower Upper
168 100
99 55.1 44.0 66.0



#33
1,2-Dichloroethane-d4
Concen: 63.027 ug/l
RT: 8.061 min Scan# 1040
Delta R.T. 0.000 min
Lab File: VY023836.D
Acq: 01 Dec 2025 12:26

Tgt Ion: 65 Resp: 414869
Ion Ratio Lower Upper
65 100
67 53.7 0.0 107.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.616 min Scan# 1131

Delta R.T. 0.000 min

Lab File: VY023836.D

Acq: 01 Dec 2025 12:26

Instrument :

MSVOA_Y

ClientSampleId :

WC1





Tgt Ion:114 Resp: 1275968

Ion Ratio Lower Upper

114 100

63 19.7 0.0 39.4

88 15.6 0.0 30.8



#35

Dibromofluoromethane

Concen: 54.204 ug/l

RT: 7.640 min Scan# 971

Delta R.T. 0.006 min

Lab File: VY023836.D

Acq: 01 Dec 2025 12:26




Tgt Ion:113 Resp: 409711


Ion Ratio Lower Upper

113 100

111 102.8 81.4 122.0

192 19.9 15.8 23.8









#50
Toluene-d8
Concen: 50.498 ug/l
RT: 10.109 min Scan# 1375
Delta R.T. 0.006 min
Lab File: VY023836.D
Acq: 01 Dec 2025 12:26




Instrument :
MSVOA_Y
ClientSampleId :
WC1


Tgt Ion: 98 Resp: 1516112
Ion Ratio Lower Upper
98 100
100 65.4 51.9 77.9



#62
4-Bromofluorobenzene
Concen: 45.254 ug/l
RT: 12.408 min Scan# 1753
Delta R.T. 0.000 min
Lab File: VY023836.D
Acq: 01 Dec 2025 12:26

Tgt Ion: 95 Resp: 447050
Ion Ratio Lower Upper
95 100
174 85.5 0.0 168.6
176 82.4 0.0 164.6





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.420 min Scan# 11

Delta R.T. 0.006 min

Lab File: VY023836.D

Acq: 01 Dec 2025 12:26

Instrument :

MSVOA_Y

ClientSampleId :

WC1





Tgt Ion:117 Resp: 1114794

Ion Ratio Lower Upper

117 100

82 55.0 45.4 68.0

119 32.0 26.0 39.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.347 min Scan# 1907

Delta R.T. 0.000 min

Lab File: VY023836.D

Acq: 01 Dec 2025 12:26




Tgt Ion:152 Resp: 445014

Ion Ratio Lower Upper

152 100

115 57.7 28.7 86.3

150 157.1 0.0 345.6



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120125\
 Data File : VY023836.D
 Acq On : 01 Dec 2025 12:26
 Operator : SY/MD
 Sample : Q3740-01
 Misc : 4.98g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC1

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Filtering: 5

Sampling : 1

Min Area: 3 % of largest Peak

Start Thrs: 0.2

Max Peaks: 100

Stop Thrs : 0

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y112625S.M

Title : SW846 8260

Signal : TIC: VY023836.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	4.616	464	475	486	rBV2	24415	73300	1.84%	0.356%
2	6.890	835	848	862	rBV	52285	171708	4.30%	0.833%
3	7.640	960	971	977	rBV	566863	1373573	34.40%	6.664%
4	7.713	977	983	998	rVB	847794	1950356	48.85%	9.463%
5	8.061	1026	1040	1054	rBV2	551747	1343838	33.66%	6.520%
6	8.616	1121	1131	1145	rBV	1499583	2922268	73.19%	14.179%
7	10.109	1368	1376	1396	rBV	2356435	3992719	100.00%	19.372%
8	10.512	1434	1442	1448	rBV2	100211	181001	4.53%	0.878%
9	10.878	1495	1502	1511	rBV	79330	139415	3.49%	0.676%
10	11.249	1557	1563	1571	rVB	23696	40292	1.01%	0.195%
11	11.420	1583	1591	1604	rBV	2047717	3383140	84.73%	16.415%
12	12.408	1746	1753	1763	rBV	1535166	2389756	59.85%	11.595%
13	13.347	1900	1907	1922	rBV	1774020	2649094	66.35%	12.853%


Sum of corrected areas: 20610460

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120125\
Data File : VY023836.D
Acq On : 01 Dec 2025 12:26
Operator : SY/MD
Sample : Q3740-01
Misc : 4.98g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC1

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y112625S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120125\
Data File : VY023836.D
Acq On : 01 Dec 2025 12:26
Operator : SY/MD
Sample : Q3740-01
Misc : 4.98g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC1

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y112625S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected

A

B

C

D

E

F

G

H

I

J

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120125\
Data File : VY023836.D
Acq On : 01 Dec 2025 12:26
Operator : SY/MD
Sample : Q3740-01
Misc : 4.98g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC1

A

B

C

D

E

F

G

H

I

J

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y112625S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	--Internal Standard--			
					#	RT	Resp	Conc

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120125\
 Data File : VY023830.D
 Acq On : 01 Dec 2025 09:13
 Operator : SY/MD
 Sample : VY1201SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1201SBL01

A
B
C
D
E
F
G
H
I
J

Quant Time: Dec 02 03:12:33 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y112625S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 27 01:14:36 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	719047	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.622	114	1221329	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	995220	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	399677	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.067	65	408084	58.278	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 116.560%	
35) Dibromofluoromethane	7.640	113	368921	50.991	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 101.980%	
50) Toluene-d8	10.109	98	1423543	49.536	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 99.080%	
62) 4-Bromofluorobenzene	12.408	95	399287	42.228	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 84.460%	


Target Compounds Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120125\
Data File : VY023830.D
Acq On : 01 Dec 2025 09:13
Operator : SY/MD
Sample : VY1201SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1201SBL01

Quant Time: Dec 02 03:12:33 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y112625S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 27 01:14:36 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.713 min Scan# 91
Delta R.T. 0.006 min
Lab File: VY023830.D
Acq: 01 Dec 2025 09:13




Instrument :
MSVOA_Y
ClientSampleId :
VY1201SBL01


Tgt Ion:168 Resp: 719047
Ion Ratio Lower Upper
168 100
99 54.1 44.0 66.0



#33
1,2-Dichloroethane-d4
Concen: 58.278 ug/l
RT: 8.067 min Scan# 1041
Delta R.T. 0.006 min
Lab File: VY023830.D
Acq: 01 Dec 2025 09:13

Tgt Ion: 65 Resp: 408084
Ion Ratio Lower Upper
65 100
67 53.0 0.0 107.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.622 min Scan# 1132

Delta R.T. 0.006 min

Lab File: VY023830.D

Acq: 01 Dec 2025 09:13

Instrument :

MSVOA_Y

ClientSampleId :

VY1201SBL01





Tgt Ion:114 Resp: 1221329

Ion Ratio Lower Upper

114 100

63 19.6 0.0 39.4

88 15.2 0.0 30.8



#35

Dibromofluoromethane

Concen: 50.991 ug/l

RT: 7.640 min Scan# 971

Delta R.T. 0.006 min

Lab File: VY023830.D

Acq: 01 Dec 2025 09:13




Tgt Ion:113 Resp: 368921


Ion Ratio Lower Upper

113 100

111 102.6 81.4 122.0

192 19.9 15.8 23.8









#50
Toluene-d8
Concen: 49.536 ug/l
RT: 10.109 min Scan# 1375
Delta R.T. 0.006 min
Lab File: VY023830.D
Acq: 01 Dec 2025 09:13




Instrument :
MSVOA_Y
ClientSampleId :
VY1201SBL01


Tgt Ion: 98 Resp: 1423543
Ion Ratio Lower Upper
98 100
100 64.8 51.9 77.9



#62
4-Bromofluorobenzene
Concen: 42.228 ug/l
RT: 12.408 min Scan# 1753
Delta R.T. 0.000 min
Lab File: VY023830.D
Acq: 01 Dec 2025 09:13

Tgt Ion: 95 Resp: 399287
Ion Ratio Lower Upper
95 100
174 86.7 0.0 168.6
176 83.0 0.0 164.6





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.420 min Scan# 11

Delta R.T. 0.006 min

Lab File: VY023830.D

Acq: 01 Dec 2025 09:13

Instrument :

MSVOA_Y

ClientSampleId :

VY1201SBL01





Tgt Ion:117 Resp: 995220

Ion Ratio Lower Upper

117 100

82 54.7 45.4 68.0

119 31.1 26.0 39.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.346 min Scan# 1907

Delta R.T. 0.000 min

Lab File: VY023830.D

Acq: 01 Dec 2025 09:13




Tgt Ion:152 Resp: 399677

Ion Ratio Lower Upper

152 100

115 57.5 28.7 86.3

150 157.0 0.0 345.6



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120125\
Data File : VY023830.D
Acq On : 01 Dec 2025 09:13
Operator : SY/MD
Sample : VY1201SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1201SBL01

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Sampling : 1

Start Thrs: 0.2

Stop Thrs : 0

Filtering: 5

Min Area: 3 % of largest Peak

Max Peaks: 100

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y112625S.M

Title : SW846 8260

Signal : TIC: VY023830.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	7.640	959	971	977	rBV	512914	1222015	32.91%	6.638%
2	7.713	977	983	997	rVB	922989	2090149	56.29%	11.354%
3	8.067	1028	1041	1054	rBV	512329	1158612	31.20%	6.294%
4	8.622	1122	1132	1151	rBV	1402835	2785988	75.03%	15.134%
5	10.109	1368	1376	1385	rBV	2235133	3713240	100.00%	20.171%
6	11.420	1583	1591	1605	rBV	1839514	3003024	80.87%	16.313%
7	12.408	1745	1753	1766	rBV	1331174	2060538	55.49%	11.193%
8	13.346	1900	1907	1917	rBV	1581696	2375670	63.98%	12.905%




Sum of corrected areas: 18409236

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120125\
Data File : VY023830.D
Acq On : 01 Dec 2025 09:13
Operator : SY/MD
Sample : VY1201SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1201SBL01

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y112625S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120125\
Data File : VY023830.D
Acq On : 01 Dec 2025 09:13
Operator : SY/MD
Sample : VY1201SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1201SBL01

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y112625S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected

A
B
C
D
E
F
G
H
I
J

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120125\
Data File : VY023830.D
Acq On : 01 Dec 2025 09:13
Operator : SY/MD
Sample : VY1201SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1201SBL01

A

B

C

D

E

F

G

H

I

J

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y112625S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST0.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	--Internal Standard--			
					#	RT	Resp	Conc

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120125\
 Data File : VY023831.D
 Acq On : 01 Dec 2025 09:37
 Operator : SY/MD
 Sample : VY1201SBS01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1201SBS01

Manual Integrations APPROVED

Reviewed By : John Carlone 12/02/2025
 Supervised By : Mahesh Dadoda 12/02/2025

Quant Time: Dec 02 03:12:53 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y112625S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 27 01:14:36 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	535284	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	821167	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	693280	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	360194	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	248830	47.734	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	95.460%		
35) Dibromofluoromethane	7.640	113	245782	50.526	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	101.060%		
50) Toluene-d8	10.109	98	970281	50.217	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery =	100.440%		
62) 4-Bromofluorobenzene	12.408	95	312712	49.188	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery =	98.380%		
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.867	85	89977	22.304	ug/l	98
3) Chloromethane	2.074	50	113977	18.468	ug/l	98
4) Vinyl Chloride	2.214	62	132305	19.818	ug/l	97
5) Bromomethane	2.598	94	91332	19.393	ug/l	98
6) Chloroethane	2.739	64	88723	20.392	ug/l	99
7) Trichlorofluoromethane	3.068	101	197869	21.922	ug/l	97
8) Diethyl Ether	3.458	74	57657	20.282	ug/l	98
9) 1,1,2-Trichlorotrifluo...	3.824	101	121248	23.016	ug/l	100
10) Methyl Iodide	4.013	142	111306	19.282	ug/l	100
11) Tert butyl alcohol	4.872	59	34349	80.230	ug/l	99
12) 1,1-Dichloroethene	3.793	96	109976	21.220	ug/l	98
13) Acrolein	3.659	56	58063	94.341	ug/l	99
14) Allyl chloride	4.391	41	175485	21.374	ug/l	100
15) Acrylonitrile	5.074	53	130380	106.262	ug/l	99
16) Acetone	3.879	43	178820	119.062	ug/l	99
17) Carbon Disulfide	4.110	76	302314	18.075	ug/l	99
18) Methyl Acetate	4.391	43	53409	18.414	ug/l	98
19) Methyl tert-butyl Ether	5.122	73	272508	20.576	ug/l	98
20) Methylene Chloride	4.629	84	119171	18.768	ug/l	98
21) trans-1,2-Dichloroethene	5.122	96	121125	21.109	ug/l	99
22) Diisopropyl ether	6.025	45	393709	22.177	ug/l	99
23) Vinyl Acetate	5.970	43	1063545	107.732	ug/l	100
24) 1,1-Dichloroethane	5.927	63	226397	21.678	ug/l	97
25) 2-Butanone	6.903	43	202591	110.530	ug/l	98
26) 2,2-Dichloropropane	6.890	77	197627	21.349	ug/l	99
27) cis-1,2-Dichloroethene	6.896	96	140195	21.039	ug/l	99
28) Bromochloromethane	7.250	49	83277	19.170	ug/l	99
29) Tetrahydrofuran	7.268	42	103454	101.198	ug/l	99
30) Chloroform	7.427	83	230773	21.975	ug/l	100
31) Cyclohexane	7.707	56	210277	20.982	ug/l	96
32) 1,1,1-Trichloroethane	7.622	97	208075	22.322	ug/l	99
36) 1,1-Dichloropropene	7.841	75	170161	22.814	ug/l	98
37) Ethyl Acetate	6.988	43	76926	21.927	ug/l	98
38) Carbon Tetrachloride	7.823	117	189158	23.105	ug/l	99
39) Methylcyclohexane	9.110	83	216857	22.015	ug/l	98
40) Benzene	8.085	78	505947	22.043	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120125\
 Data File : VY023831.D
 Acq On : 01 Dec 2025 09:37
 Operator : SY/MD
 Sample : VY1201SBS01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1201SBS01

Manual Integrations APPROVED

Reviewed By :John Carlone 12/02/2025
 Supervised By :Mahesh Dadoda 12/02/2025

Quant Time: Dec 02 03:12:53 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y112625S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 27 01:14:36 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.220	41	36625	20.911	ug/l	96
42) 1,2-Dichloroethane	8.165	62	131688	22.255	ug/l	99
43) Isopropyl Acetate	8.201	43	135465	20.662	ug/l	98
44) Trichloroethene	8.866	130	133143	22.609	ug/l	99
45) 1,2-Dichloropropane	9.146	63	122233	22.587	ug/l	99
46) Dibromomethane	9.238	93	63760	21.713	ug/l	98
47) Bromodichloromethane	9.427	83	170078	22.162	ug/l	98
48) Methyl methacrylate	9.225	41	65192	21.120	ug/l	99
49) 1,4-Dioxane	9.231	88	13950	423.887	ug/l #	89
51) 4-Methyl-2-Pentanone	10.000	43	369864	107.166	ug/l	99
52) Toluene	10.170	92	320816	22.547	ug/l	100
53) t-1,3-Dichloropropene	10.396	75	150627	21.427	ug/l	97
54) cis-1,3-Dichloropropene	9.859	75	185024	21.988	ug/l	96
55) 1,1,2-Trichloroethane	10.573	97	86396	21.964	ug/l	95
56) Ethyl methacrylate	10.439	69	105292	20.386	ug/l	98
57) 1,3-Dichloropropane	10.719	76	149278	21.755	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.713	63	212318	85.570	ug/l	99
59) 2-Hexanone	10.762	43	291270	114.126	ug/l	98
60) Dibromochloromethane	10.914	129	116254	22.024	ug/l	99
61) 1,2-Dibromoethane	11.018	107	77457	21.214	ug/l	100
64) Tetrachloroethene	10.652	164	153349	23.317	ug/l	96
65) Chlorobenzene	11.444	112	344594	22.644	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.518	131	118503	22.922	ug/l	99
67) Ethyl Benzene	11.524	91	594212	22.860	ug/l	97
68) m/p-Xylenes	11.627	106	462889	46.213	ug/l	99
69) o-Xylene	11.957	106	211631	22.537	ug/l	99
70) Styrene	11.969	104	350904	22.684	ug/l	100
71) Bromoform	12.133	173	66096	21.918	ug/l	99
73) Isopropylbenzene	12.255	105	577632	22.596	ug/l	100
74) N-amyl acetate	12.072	43	113778	19.922	ug/l	98
75) 1,1,2,2-Tetrachloroethane	12.505	83	87761	20.852	ug/l	99
76) 1,2,3-Trichloropropane	12.560	75	61026m	19.483	ug/l	
77) Bromobenzene	12.530	156	130000	21.620	ug/l	98
78) n-propylbenzene	12.597	91	707533	22.988	ug/l	99
79) 2-Chlorotoluene	12.682	91	391060	22.077	ug/l	98
80) 1,3,5-Trimethylbenzene	12.737	105	470287	22.535	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.304	75	30845	20.814	ug/l	97
82) 4-Chlorotoluene	12.780	91	404596	22.240	ug/l	100
83) tert-Butylbenzene	12.999	119	423201	22.574	ug/l	99
84) 1,2,4-Trimethylbenzene	13.042	105	466685	22.376	ug/l	100
85) sec-Butylbenzene	13.176	105	637239	23.027	ug/l	99
86) p-Isopropyltoluene	13.292	119	528395	22.843	ug/l	99
87) 1,3-Dichlorobenzene	13.286	146	260011	21.576	ug/l	100
88) 1,4-Dichlorobenzene	13.365	146	259834	21.601	ug/l	99
89) n-Butylbenzene	13.615	91	482726	22.760	ug/l	99
90) Hexachloroethane	13.877	117	108829	22.393	ug/l	99
91) 1,2-Dichlorobenzene	13.657	146	231193	21.939	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.273	75	13788	19.572	ug/l	98
93) 1,2,4-Trichlorobenzene	14.919	180	130560	20.779	ug/l	99
94) Hexachlorobutadiene	15.023	225	88094	22.514	ug/l	97
95) Naphthalene	15.145	128	208532	19.680	ug/l	99
96) 1,2,3-Trichlorobenzene	15.328	180	113544	21.218	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120125\
Data File : VY023831.D
Acq On : 01 Dec 2025 09:37
Operator : SY/MD
Sample : VY1201SBS01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1201SBS01

Manual Integrations
APPROVED

Reviewed By :John Carlone 12/02/2025
Supervised By :Mahesh Dadoda 12/02/2025

Quant Time: Dec 02 03:12:53 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y112625S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 27 01:14:36 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120125\
Data File : VY023831.D
Acq On : 01 Dec 2025 09:37
Operator : SY/MD
Sample : VY1201SBS01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1201SBS01

Manual Integrations APPROVED

Reviewed By :John Carlone 12/02/2025
Supervised By :Mahesh Dadoda 12/02/2025

Quant Time: Dec 02 03:12:53 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y112625S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 27 01:14:36 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120125\
 Data File : VY023832.D
 Acq On : 01 Dec 2025 10:00
 Operator : SY/MD
 Sample : VY1201SBS01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1201SBS01

Manual Integrations APPROVED

Reviewed By : John Carlone 12/02/2025
 Supervised By : Mahesh Dadoda 12/02/2025

Quant Time: Dec 02 03:13:44 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y112625S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 27 01:14:36 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	526036	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	818117	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	695554	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	353062	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.067	65	245781	47.978	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	95.960%
35) Dibromofluoromethane	7.640	113	244714	50.494	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.980%
50) Toluene-d8	10.109	98	961906	49.969	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.940%
62) 4-Bromofluorobenzene	12.408	95	306085	48.325	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	96.640%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.867	85	88398	22.298	ug/l	100
3) Chloromethane	2.074	50	113584	18.728	ug/l	98
4) Vinyl Chloride	2.208	62	129763	19.779	ug/l	98
5) Bromomethane	2.598	94	89913	19.427	ug/l	96
6) Chloroethane	2.739	64	86497	20.230	ug/l	98
7) Trichlorofluoromethane	3.062	101	193966	21.867	ug/l	97
8) Diethyl Ether	3.458	74	59159	21.176	ug/l	95
9) 1,1,2-Trichlorotrifluo...	3.824	101	119754	23.132	ug/l	99
10) Methyl Iodide	4.007	142	114483	20.181	ug/l	100
11) Tert butyl alcohol	4.878	59	38994	92.681	ug/l	98
12) 1,1-Dichloroethene	3.793	96	109088	21.419	ug/l	100
13) Acrolein	3.659	56	52436	86.696	ug/l	99
14) Allyl chloride	4.391	41	173297	21.478	ug/l	100
15) Acrylonitrile	5.067	53	135769	112.600	ug/l	99
16) Acetone	3.872	43	190658	129.176	ug/l	97
17) Carbon Disulfide	4.110	76	296016	18.010	ug/l	99
18) Methyl Acetate	4.397	43	57954	20.332	ug/l	100
19) Methyl tert-butyl Ether	5.122	73	283581	21.788	ug/l	97
20) Methylene Chloride	4.628	84	120089	19.246	ug/l	99
21) trans-1,2-Dichloroethene	5.122	96	120293	21.332	ug/l	96
22) Diisopropyl ether	6.031	45	395326	22.660	ug/l	98
23) Vinyl Acetate	5.970	43	1093650	112.730	ug/l	99
24) 1,1-Dichloroethane	5.927	63	225791	22.000	ug/l	97
25) 2-Butanone	6.896	43	220919	122.648	ug/l	99
26) 2,2-Dichloropropane	6.890	77	196626	21.614	ug/l	99
27) cis-1,2-Dichloroethene	6.896	96	140774	21.498	ug/l	100
28) Bromochloromethane	7.256	49	83127	19.472	ug/l	99
29) Tetrahydrofuran	7.268	42	110786	110.275	ug/l	99
30) Chloroform	7.427	83	228113	22.104	ug/l	95
31) Cyclohexane	7.707	56	209574	21.279	ug/l	98
32) 1,1,1-Trichloroethane	7.622	97	205882	22.475	ug/l	99
36) 1,1-Dichloropropene	7.841	75	167201	22.501	ug/l	99
37) Ethyl Acetate	6.994	43	82518	23.609	ug/l	100
38) Carbon Tetrachloride	7.823	117	181727	22.280	ug/l	97
39) Methylcyclohexane	9.109	83	214053	21.811	ug/l	98
40) Benzene	8.085	78	498139	21.784	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120125\
 Data File : VY023832.D
 Acq On : 01 Dec 2025 10:00
 Operator : SY/MD
 Sample : VY1201SBS01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1201SBS01

Manual Integrations APPROVED

Reviewed By : John Carlone 12/02/2025
 Supervised By : Mahesh Dadoda 12/02/2025

Quant Time: Dec 02 03:13:44 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y112625S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 27 01:14:36 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.232	41	38160	21.869	ug/l	95
42) 1,2-Dichloroethane	8.158	62	132848	22.535	ug/l	99
43) Isopropyl Acetate	8.201	43	144201	22.077	ug/l #	87
44) Trichloroethene	8.866	130	131042	22.335	ug/l	99
45) 1,2-Dichloropropane	9.140	63	122189	22.663	ug/l	98
46) Dibromomethane	9.231	93	65343	22.335	ug/l	99
47) Bromodichloromethane	9.426	83	171533	22.435	ug/l	97
48) Methyl methacrylate	9.225	41	68408	22.244	ug/l	99
49) 1,4-Dioxane	9.237	88	15875	484.179	ug/l	97
51) 4-Methyl-2-Pentanone	9.999	43	394930	114.855	ug/l	99
52) Toluene	10.170	92	317325	22.385	ug/l	99
53) t-1,3-Dichloropropene	10.396	75	152796	21.816	ug/l	95
54) cis-1,3-Dichloropropene	9.859	75	183114	21.842	ug/l	97
55) 1,1,2-Trichloroethane	10.573	97	90140	23.001	ug/l	96
56) Ethyl methacrylate	10.438	69	112544	21.871	ug/l	98
57) 1,3-Dichloropropane	10.719	76	153060	22.389	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.713	63	246607	99.760	ug/l	99
59) 2-Hexanone	10.762	43	309938	121.893	ug/l	99
60) Dibromochloromethane	10.914	129	117865	22.413	ug/l	99
61) 1,2-Dibromoethane	11.018	107	80328	22.082	ug/l	99
64) Tetrachloroethene	10.646	164	152218	23.069	ug/l	99
65) Chlorobenzene	11.444	112	338867	22.195	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.517	131	120164	23.167	ug/l	98
67) Ethyl Benzene	11.517	91	590343	22.637	ug/l	99
68) m/p-Xylenes	11.627	106	457077	45.484	ug/l	100
69) o-Xylene	11.956	106	209946	22.284	ug/l	99
70) Styrene	11.969	104	352673	22.724	ug/l	99
71) Bromoform	12.133	173	67224	22.219	ug/l #	99
73) Isopropylbenzene	12.255	105	569849	22.742	ug/l	100
74) N-amyl acetate	12.072	43	121531	21.709	ug/l	98
75) 1,1,2,2-Tetrachloroethane	12.505	83	93304	22.617	ug/l	99
76) 1,2,3-Trichloropropane	12.554	75	61775m	20.120	ug/l	
77) Bromobenzene	12.536	156	132780	22.529	ug/l	99
78) n-propylbenzene	12.597	91	693288	22.981	ug/l	99
79) 2-Chlorotoluene	12.682	91	390507	22.491	ug/l	99
80) 1,3,5-Trimethylbenzene	12.737	105	469631	22.958	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.304	75	32372	22.286	ug/l	97
82) 4-Chlorotoluene	12.779	91	406161	22.777	ug/l	99
83) tert-Butylbenzene	12.999	119	417792	22.735	ug/l	100
84) 1,2,4-Trimethylbenzene	13.042	105	464199	22.707	ug/l	100
85) sec-Butylbenzene	13.176	105	626233	23.086	ug/l	100
86) p-Isopropyltoluene	13.292	119	517504	22.824	ug/l	99
87) 1,3-Dichlorobenzene	13.285	146	263200	22.282	ug/l	99
88) 1,4-Dichlorobenzene	13.365	146	258566	21.930	ug/l	99
89) n-Butylbenzene	13.615	91	477523	22.969	ug/l	98
90) Hexachloroethane	13.883	117	107367	22.538	ug/l	100
91) 1,2-Dichlorobenzene	13.657	146	232245	22.484	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.273	75	15026	21.760	ug/l	98
93) 1,2,4-Trichlorobenzene	14.919	180	133887	21.739	ug/l	98
94) Hexachlorobutadiene	15.023	225	86872	22.650	ug/l	99
95) Naphthalene	15.145	128	219911	21.173	ug/l	99
96) 1,2,3-Trichlorobenzene	15.328	180	116198	22.153	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120125\
Data File : VY023832.D
Acq On : 01 Dec 2025 10:00
Operator : SY/MD
Sample : VY1201SBSD01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1201SBSD01

Manual Integrations
APPROVED

Reviewed By :John Carlone 12/02/2025
Supervised By :Mahesh Dadoda 12/02/2025

Quant Time: Dec 02 03:13:44 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y112625S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 27 01:14:36 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_Y
ClientSampleId :
VY1201SBSD01

Manual Integrations APPROVED

Reviewed By :John Carlone 12/02/2025
Supervised By :Mahesh Dadoda 12/02/2025



Manual Integration Report

Sequence:	VY112625	Instrument	MSVOA_y
-----------	----------	------------	---------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDIC005	VY023809.D	1,2,3-Trichloropropane	sam	12/1/2025 9:09:40 AM	MMDadoda	12/2/2025 8:05:29 AM	Peak Integrated by Software
VSTDIC005	VY023809.D	Methacrylonitrile	sam	12/1/2025 9:09:40 AM	MMDadoda	12/2/2025 8:05:29 AM	Peak Integrated by Software
VSTDIC005	VY023809.D	Tert butyl alcohol	sam	12/1/2025 9:09:40 AM	MMDadoda	12/2/2025 8:05:29 AM	Peak Integrated by Software
VSTDIC010	VY023810.D	1,2,3-Trichloropropane	sam	12/1/2025 9:09:45 AM	MMDadoda	12/2/2025 8:05:32 AM	Peak Integrated by Software
VSTDIC020	VY023811.D	1,2,3-Trichloropropane	sam	12/1/2025 9:44:03 AM	MMDadoda	12/2/2025 8:05:35 AM	Peak Integrated by Software
VSTDICCC050	VY023812.D	1,2,3-Trichloropropane	sam	12/1/2025 9:09:50 AM	MMDadoda	12/2/2025 8:05:40 AM	Peak Integrated by Software
VSTDIC100	VY023813.D	1,2,3-Trichloropropane	sam	12/1/2025 9:46:48 AM	MMDadoda	12/2/2025 8:05:43 AM	Peak Integrated by Software
VSTDIC150	VY023814.D	1,2,3-Trichloropropane	sam	12/1/2025 9:46:53 AM	MMDadoda	12/2/2025 8:05:46 AM	Peak Integrated by Software
VSTDICV050	VY023816.D	1,2,3-Trichloropropane	sam	12/1/2025 9:46:58 AM	MMDadoda	12/2/2025 8:05:50 AM	Peak Integrated by Software
VSTDCCC050	VY023827.D	1,2,3-Trichloropropane	sam	12/1/2025 9:44:22 AM	MMDadoda	12/2/2025 8:06:02 AM	Peak Integrated by Software

Manual Integration Report

Sequence:	VY120125	Instrument	MSVOA_y
-----------	----------	------------	---------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDCCC050	VY023829.D	1,2,3-Trichloropropane	JOHN	12/2/2025 8:59:44 AM	MMDadoda	12/2/2025 10:15:52 AM	Peak Integrated by Software
VY1201SBS01	VY023831.D	1,2,3-Trichloropropane	JOHN	12/2/2025 8:59:49 AM	MMDadoda	12/2/2025 10:15:59 AM	Peak Integrated by Software
VY1201SBSD0 1	VY023832.D	1,2,3-Trichloropropane	JOHN	12/2/2025 8:59:53 AM	MMDadoda	12/2/2025 10:16:03 AM	Peak Integrated by Software
VSTDCCC050	VY023847.D	1,2,3-Trichloropropane	JOHN	12/2/2025 9:00:02 AM	MMDadoda	12/2/2025 10:16:12 AM	Peak Integrated by Software

Instrument ID: **MSVOA_Y**

Daily Analysis Runlog For Sequence/QCBatch ID # VY112625

Review By	Semsettin Yesilyurt	Review On	12/1/2025 9:50:11 AM		
Supervise By	John Carlone	Supervise On	12/1/2025 2:45:01 PM		
SubDirectory	VY112625	HP Acquire Method	MSVOA_Y	HP Processing Method	82y112625s.m
STD. NAME		STD REF.#			
Tune/Reschk		VP136432			
Initial Calibration Stds		VP136433,VP136434,VP136435,VP136436,VP136437,VP136438			
CCC		VP136440			
Internal Standard/PEM		VP136146			
ICV/I.BLK		VP136439			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VY023808.D	26 Nov 2025 07:11	SY/MD	Ok
2	VSTDIC005	VY023809.D	26 Nov 2025 08:04	SY/MD	Ok,M
3	VSTDIC010	VY023810.D	26 Nov 2025 08:27	SY/MD	Ok,M
4	VSTDIC020	VY023811.D	26 Nov 2025 08:49	SY/MD	Ok,M
5	VSTDIC050	VY023812.D	26 Nov 2025 09:17	SY/MD	Ok,M
6	VSTDIC100	VY023813.D	26 Nov 2025 09:39	SY/MD	Ok,M
7	VSTDIC150	VY023814.D	26 Nov 2025 10:02	SY/MD	Ok,M
8	VIBLK	VY023815.D	26 Nov 2025 10:27	SY/MD	Ok
9	VSTDICV050	VY023816.D	26 Nov 2025 10:50	SY/MD	Ok,M
10	VY1126SBL01	VY023817.D	26 Nov 2025 11:13	SY/MD	Ok
11	VY1126SBS01	VY023818.D	26 Nov 2025 11:47	SY/MD	Ok,M
12	VY1126SBSD01	VY023819.D	26 Nov 2025 12:09	SY/MD	Ok,M
13	Q3604-04	VY023820.D	26 Nov 2025 12:34	SY/MD	Not Ok
14	Q3604-05	VY023821.D	26 Nov 2025 12:57	SY/MD	Not Ok
15	VIBLK	VY023822.D	26 Nov 2025 13:21	SY/MD	Ok
16	Q3721-01	VY023823.D	26 Nov 2025 13:44	SY/MD	Ok
17	Q3723-01	VY023824.D	26 Nov 2025 14:08	SY/MD	Ok
18	Q3725-04	VY023825.D	26 Nov 2025 14:31	SY/MD	Ok
19	VIBLK	VY023826.D	26 Nov 2025 14:54	SY/MD	Ok
20	VSTDIC050	VY023827.D	26 Nov 2025 15:17	SY/MD	Ok,M

M : Manual Integration

Instrument ID: **MSVOA_Y**

Daily Analysis Runlog For Sequence/QCBatch ID # VY120125

Review By	John Carlone	Review On	12/2/2025 9:00:18 AM		
Supervise By	Mahesh Dadoda	Supervise On	12/2/2025 10:16:40 AM		
SubDirectory	VY120125	HP Acquire Method	MSVOA_Y	HP Processing Method	82y112625s.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP136454			
CCC		VP136455,VP136456			
Internal Standard/PEM		VP136146			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VY023828.D	01 Dec 2025 07:24	SY/MD	Ok
2	VSTDCCC050	VY023829.D	01 Dec 2025 07:55	SY/MD	Ok,M
3	VY1201SBL01	VY023830.D	01 Dec 2025 09:13	SY/MD	Ok
4	VY1201SBS01	VY023831.D	01 Dec 2025 09:37	SY/MD	Ok,M
5	VY1201SBSD01	VY023832.D	01 Dec 2025 10:00	SY/MD	Ok,M
6	Q3719-09RE	VY023833.D	01 Dec 2025 10:38	SY/MD	Confirms
7	Q3719-13	VY023834.D	01 Dec 2025 11:25	SY/MD	Ok
8	Q3735-01	VY023835.D	01 Dec 2025 12:02	SY/MD	ReRun
9	Q3740-01	VY023836.D	01 Dec 2025 12:26	SY/MD	Ok
10	Q3741-02	VY023837.D	01 Dec 2025 12:49	SY/MD	Ok
11	Q3741-03	VY023838.D	01 Dec 2025 13:12	SY/MD	Ok
12	Q3741-04	VY023839.D	01 Dec 2025 13:36	SY/MD	ReRun
13	Q3741-05	VY023840.D	01 Dec 2025 13:59	SY/MD	ReRun
14	Q3741-01	VY023841.D	01 Dec 2025 14:23	SY/MD	Ok
15	Q3741-06	VY023842.D	01 Dec 2025 14:46	SY/MD	ReRun
16	Q3742-01	VY023843.D	01 Dec 2025 15:10	SY/MD	Ok
17	Q3742-02	VY023844.D	01 Dec 2025 15:33	SY/MD	Ok
18	Q3742-03	VY023845.D	01 Dec 2025 15:57	SY/MD	Ok
19	Q3742-04	VY023846.D	01 Dec 2025 16:20	SY/MD	Ok
20	VSTDCCC050	VY023847.D	01 Dec 2025 17:47	SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY112625

Review By	Semsettin Yesilyurt	Review On	12/1/2025 9:50:11 AM		
Supervise By	John Carlone	Supervise On	12/1/2025 2:45:01 PM		
SubDirectory	VY112625	HP Acquire Method	MSVOA_Y	HP Processing Method	82y112625s.m
STD. NAME		STD REF.#			
Tune/Reschk		VP136432			
Initial Calibration Stds		VP136433,VP136434,VP136435,VP136436,VP136437,VP136438			
CCC		VP136440			
Internal Standard/PEM		VP136146			
ICV/I.BLK		VP136439			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY023808.D	26 Nov 2025 07:11		SY/MD	Ok
2	VSTDIC005	VSTDIC005	VY023809.D	26 Nov 2025 08:04		SY/MD	Ok,M
3	VSTDIC010	VSTDIC010	VY023810.D	26 Nov 2025 08:27		SY/MD	Ok,M
4	VSTDIC020	VSTDIC020	VY023811.D	26 Nov 2025 08:49		SY/MD	Ok,M
5	VSTDIC050	VSTDIC050	VY023812.D	26 Nov 2025 09:17		SY/MD	Ok,M
6	VSTDIC100	VSTDIC100	VY023813.D	26 Nov 2025 09:39		SY/MD	Ok,M
7	VSTDIC150	VSTDIC150	VY023814.D	26 Nov 2025 10:02		SY/MD	Ok,M
8	VIBLK	VIBLK	VY023815.D	26 Nov 2025 10:27		SY/MD	Ok
9	VSTDICV050	ICVVY112625	VY023816.D	26 Nov 2025 10:50		SY/MD	Ok,M
10	VY1126SBL01	VY1126SBL01	VY023817.D	26 Nov 2025 11:13		SY/MD	Ok
11	VY1126SBS01	VY1126SBS01	VY023818.D	26 Nov 2025 11:47		SY/MD	Ok,M
12	VY1126SBS01	VY1126SBS01	VY023819.D	26 Nov 2025 12:09		SY/MD	Ok,M
13	Q3604-04	S-1	VY023820.D	26 Nov 2025 12:34	not reported,vial A	SY/MD	Not Ok
14	Q3604-05	S-2	VY023821.D	26 Nov 2025 12:57	not reported,vial A	SY/MD	Not Ok
15	VIBLK	VIBLK	VY023822.D	26 Nov 2025 13:21		SY/MD	Ok
16	Q3721-01	EO-2-112525	VY023823.D	26 Nov 2025 13:44	vial A	SY/MD	Ok
17	Q3723-01	CC-1-112525	VY023824.D	26 Nov 2025 14:08	vial A	SY/MD	Ok
18	Q3725-04	STOCKPILE-SAMPLES	VY023825.D	26 Nov 2025 14:31	vial A	SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY112625

Review By	Semsettin Yesilyurt	Review On	12/1/2025 9:50:11 AM		
Supervise By	John Carlone	Supervise On	12/1/2025 2:45:01 PM		
SubDirectory	VY112625	HP Acquire Method	MSVOA_Y	HP Processing Method	82y112625s.m

STD. NAME	STD REF.#
Tune/Reschk	VP136432
Initial Calibration Stds	VP136433,VP136434,VP136435,VP136436,VP136437,VP136438
CCC	VP136440
Internal Standard/PEM	VP136146
ICV/I.BLK	VP136439
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	VIBLK	VIBLK	VY023826.D	26 Nov 2025 14:54		SY/MD	Ok
20	VSTDCCC050	VSTDCCC050EC	VY023827.D	26 Nov 2025 15:17		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY120125

Review By	John Carlone	Review On	12/2/2025 9:00:18 AM
Supervise By	Maresh Dadoda	Supervise On	12/2/2025 10:16:40 AM
SubDirectory	VY120125	HP Acquire Method	MSVOA_Y HP Processing Method 82y112625s.m
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP136454		
CCC	VP136455,VP136456		
Internal Standard/PEM	VP136146		
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY023828.D	01 Dec 2025 07:24		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY023829.D	01 Dec 2025 07:55		SY/MD	Ok,M
3	VY1201SBL01	VY1201SBL01	VY023830.D	01 Dec 2025 09:13		SY/MD	Ok
4	VY1201SBS01	VY1201SBS01	VY023831.D	01 Dec 2025 09:37		SY/MD	Ok,M
5	VY1201SBSD01	VY1201SBSD01	VY023832.D	01 Dec 2025 10:00		SY/MD	Ok,M
6	Q3719-09RE	SB-1125-15ARE	VY023833.D	01 Dec 2025 10:38	Internal standard fail;Surrogate fail,vial B	SY/MD	Confirms
7	Q3719-13	SB-1125-17A	VY023834.D	01 Dec 2025 11:25	vial B	SY/MD	Ok
8	Q3735-01	BU-3-112625	VY023835.D	01 Dec 2025 12:02	Internal Standard Fail,vial A	SY/MD	ReRun
9	Q3740-01	WC1	VY023836.D	01 Dec 2025 12:26	vial A	SY/MD	Ok
10	Q3741-02	B50-SP13-112025	VY023837.D	01 Dec 2025 12:49	vial A	SY/MD	Ok
11	Q3741-03	B50-SP14-112025	VY023838.D	01 Dec 2025 13:12	vial A	SY/MD	Ok
12	Q3741-04	B50-SP9-112425	VY023839.D	01 Dec 2025 13:36	Surrogate fail,vial A	SY/MD	ReRun
13	Q3741-05	B50-SP11-112525	VY023840.D	01 Dec 2025 13:59	Internal standard fail;Surrogate fail,vial A	SY/MD	ReRun
14	Q3741-01	B50-SP3-111925	VY023841.D	01 Dec 2025 14:23	vial A	SY/MD	Ok
15	Q3741-06	B50-SP4-112525	VY023842.D	01 Dec 2025 14:46	Internal standard fail;Surrogate fail,vial A	SY/MD	ReRun
16	Q3742-01	SB-1125-23	VY023843.D	01 Dec 2025 15:10	vial A	SY/MD	Ok
17	Q3742-02	SB-1125-19	VY023844.D	01 Dec 2025 15:33	vial A	SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY120125

Review By	John Carlone	Review On	12/2/2025 9:00:18 AM		
Supervise By	Mahesh Dadoda	Supervise On	12/2/2025 10:16:40 AM		
SubDirectory	VY120125	HP Acquire Method	MSVOA_Y	HP Processing Method	82y112625s.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP136454			
CCC		VP136455,VP136456			
Internal Standard/PEM		VP136146			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	Q3742-03	SB-1125-8	VY023845.D	01 Dec 2025 15:57	vial A	SY/MD	Ok
19	Q3742-04	SB-1125-9	VY023846.D	01 Dec 2025 16:20	vial A	SY/MD	Ok
20	VSTDCCC050	VSTDCCC050EC	VY023847.D	01 Dec 2025 17:47		SY/MD	Ok,M

M : Manual Integration

LAB CHRONICLE

OrderID:	Q3740	OrderDate:	11/26/2025 4:30:00 PM
Client:	G Environmental	Project:	Diamond
Contact:	Gary Landis	Location:	--Select--,A11,VOA Lab

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3740-01	WC1	SOIL	VOC-TCLVOA-10	8260D	11/26/25		12/01/25	11/26/25