

Surrogate Summary

SDG No.: **Q3740**

Client: **G Environmental**

Analytical Method: **SW8260D**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3740-01	WC1	1,2-Dichloroethane-d4	50	63.0	126		63	155
		Dibromofluoromethane	50	54.2	108		70	134
		Toluene-d8	50	50.5	101		74	123
		4-Bromofluorobenzene	50	45.3	91		52	138
VY1201SBL01	VY1201SBL01	1,2-Dichloroethane-d4	50	58.3	117		63	155
		Dibromofluoromethane	50	51.0	102		70	134
		Toluene-d8	50	49.5	99		74	123
		4-Bromofluorobenzene	50	42.2	84		52	138
VY1201SBS01	VY1201SBS01	1,2-Dichloroethane-d4	50	47.7	95		63	155
		Dibromofluoromethane	50	50.5	101		70	134
		Toluene-d8	50	50.2	100		74	123
		4-Bromofluorobenzene	50	49.2	98		52	138
VY1201SBSD01	VY1201SBSD01	1,2-Dichloroethane-d4	50	48.0	96		63	155
		Dibromofluoromethane	50	50.5	101		70	134
		Toluene-d8	50	50.0	100		74	123
		4-Bromofluorobenzene	50	48.3	97		52	138