

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3740</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>G Environmental</u>	Datafile :	<u>VY023832.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY1201SBSD01	Dichlorodifluoromethane	20	22.3	ug/Kg	112	0		64	136	20
	Chloromethane	20	18.7	ug/Kg	94	1		73	129	15
	Vinyl chloride	20	19.8	ug/Kg	99	0		73	133	15
	Bromomethane	20	19.4	ug/Kg	97	0		77	137	15
	Chloroethane	20	20.2	ug/Kg	101	1		69	130	15
	Trichlorofluoromethane	20	21.9	ug/Kg	110	0		69	134	27
	1,1,2-Trichlorotrifluoroethane	20	23.1	ug/Kg	116	1		81	123	20
	1,1-Dichloroethene	20	21.4	ug/Kg	107	1		79	121	16
	Acetone	100	130	ug/Kg	130	8		60	174	28
	Carbon disulfide	20	18.0	ug/Kg	90	1		73	125	15
	Methyl tert-butyl Ether	20	21.8	ug/Kg	109	6		77	129	20
	Methyl Acetate	20	20.3	ug/Kg	102	10		51	140	30
	Methylene Chloride	20	19.2	ug/Kg	96	2		54	158	20
	trans-1,2-Dichloroethene	20	21.3	ug/Kg	106	0		80	123	15
	1,1-Dichloroethane	20	22.0	ug/Kg	110	1		82	123	15
	Cyclohexane	20	21.3	ug/Kg	106	1		76	122	15
	2-Butanone	100	120	ug/Kg	120	9		69	131	29
	Carbon Tetrachloride	20	22.3	ug/Kg	112	4		76	129	15
	cis-1,2-Dichloroethene	20	21.5	ug/Kg	108	3		82	123	15
	Bromochloromethane	20	19.5	ug/Kg	98	2		80	127	15
	Chloroform	20	22.1	ug/Kg	111	1		82	125	15
	1,1,1-Trichloroethane	20	22.5	ug/Kg	113	1		80	126	15
	Methylcyclohexane	20	21.8	ug/Kg	109	1		77	123	15
	Benzene	20	21.8	ug/Kg	109	1		84	121	15
	1,2-Dichloroethane	20	22.5	ug/Kg	113	1		81	126	15
	Trichloroethene	20	22.3	ug/Kg	112	1		83	122	15
	1,2-Dichloropropane	20	22.7	ug/Kg	114	1		83	122	15
	Bromodichloromethane	20	22.4	ug/Kg	112	1		82	123	15
	4-Methyl-2-Pentanone	100	110	ug/Kg	110	0		70	135	26
	Toluene	20	22.4	ug/Kg	112	1		83	122	15
	t-1,3-Dichloropropene	20	21.8	ug/Kg	109	2		78	124	15
	cis-1,3-Dichloropropene	20	21.8	ug/Kg	109	1		81	122	15
	1,1,2-Trichloroethane	20	23.0	ug/Kg	115	4		82	125	15
	2-Hexanone	100	120	ug/Kg	120	9		66	138	27
	Dibromochloromethane	20	22.4	ug/Kg	112	2		79	125	15
	1,2-Dibromoethane	20	22.1	ug/Kg	111	5		80	125	16
	Tetrachloroethene	20	23.1	ug/Kg	116	1		83	125	15
	Chlorobenzene	20	22.2	ug/Kg	111	2		84	122	15
	Ethyl Benzene	20	22.6	ug/Kg	113	2		82	124	15
	m/p-Xylenes	40	45.5	ug/Kg	114	2		83	124	15
	o-Xylene	20	22.3	ug/Kg	112	1		83	123	15
	Styrene	20	22.7	ug/Kg	114	0		82	124	15
	Bromoform	20	22.2	ug/Kg	111	1		75	127	16
	Isopropylbenzene	20	22.7	ug/Kg	114	1		82	124	15
	1,1,2,2-Tetrachloroethane	20	22.6	ug/Kg	113	8		77	127	21
	1,3-Dichlorobenzene	20	22.3	ug/Kg	112	4		83	122	15
	1,4-Dichlorobenzene	20	21.9	ug/Kg	110	2		84	121	15
	1,2-Dichlorobenzene	20	22.5	ug/Kg	113	3		83	124	15

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3740</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>G Environmental</u>	Datafile :	<u>VY023832.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VY1201SBSD01	1,2-Dibromo-3-Chloropropane	20	21.8	ug/Kg	109	11		66	134	20
	1,2,4-Trichlorobenzene	20	21.7	ug/Kg	109	5		78	127	20
	1,2,3-Trichlorobenzene	20	22.2	ug/Kg	111	5		70	137	16