

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01  
 Lab Code: ACE SDG NO.: Q3740  
 Lab File ID: VY023829.D Date Analyzed: 12/01/2025  
 Instrument ID: MSVOA\_Y Time Analyzed: 07:55  
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

|                | IS1<br>AREA # | RT #  | IS2<br>AREA # | RT #  | IS3<br>AREA # | RT #  |
|----------------|---------------|-------|---------------|-------|---------------|-------|
| 12 HOUR STD    | 565841        | 7.71  | 855261        | 8.62  | 736930        | 11.42 |
| UPPER LIMIT    | 1131680       | 8.213 | 1710520       | 9.122 | 1473860       | 11.92 |
| LOWER LIMIT    | 282921        | 7.213 | 427631        | 8.122 | 368465        | 10.92 |
| EPA SAMPLE NO. |               |       |               |       |               |       |
| WC1            | 675919        | 7.71  | 1275968       | 8.62  | 1114794       | 11.42 |
| VY1201SBL01    | 719047        | 7.71  | 1221329       | 8.62  | 995220        | 11.42 |
| VY1201SBS01    | 535284        | 7.71  | 821167        | 8.62  | 693280        | 11.42 |
| VY1201SBSD01   | 526036        | 7.71  | 818117        | 8.62  | 695554        | 11.42 |

IS1 = Pentafluorobenzene  
 IS2 = 1,4-Difluorobenzene  
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.